

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF120718\
 Data File : VF060941.D
 Acq On : 7 Dec 2018 21:00
 Operator : VA/AP
 Sample : J6214-27RE
 Misc : 4.25g/5mL/MSVOA-F/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-2(0-2)RE

Manual Integrations
 APPROVED

apatel
 12/10/2018 4:38:16 PM

Quant Time: Dec 08 03:52:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	147392	50.00	ug/l	-0.06
34) 1,4-Difluorobenzene	5.58	114	291642	50.00	ug/l	-0.06
63) Chlorobenzene-d5	9.75	117	184621	50.00	ug/l	-0.06
72) 1,4-Dichlorobenzene-d4	12.54	152	42679	50.00	ug/l	-0.03
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.82	65	111812	64.40	ug/l	-0.06
Spiked Amount 50.000			Recovery	=	128.80%	
35) Dibromofluoromethane	4.09	113	119074	51.47	ug/l	-0.06
Spiked Amount 50.000			Recovery	=	102.94%	
50) Toluene-d8	7.53	98	201490	29.56	ug/l	-0.06
Spiked Amount 50.000			Recovery	=	59.12%	
62) 4-Bromofluorobenzene	11.40	95	59952	19.36	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	38.72%	
Target Compounds						
20) Methylene Chloride	2.24	84	5881m	5.007	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF120718\
Data File : VF060941.D
Acq On : 7 Dec 2018 21:00
Operator : VA/AP
Sample : J6214-27RE
Misc : 4.25g/5mL/MSVOA-F/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-2(0-2)RE

Manual Integrations
APPROVED

apatel
12/10/2018 4:38:16 PM

Quant Time: Dec 08 03:52:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

