


Data Path : Z:\voasrv\HPCHEM1\MSVOA_F\Data\VF120718\
 Data File : VF060932.D
 Acq On : 7 Dec 2018 15:47
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Title : SW846 8260
 Last Update : Tue Nov 27 01:36:26 2018



AutoFind: Scans 1091, 1092, 1093; Background Corrected with Scan 1085

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	27138	PASS
75	95	30	60	48.4	64405	PASS
95	95	100	100	100.0	133178	PASS
96	95	5	9	7.0	9299	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	67.7	90216	PASS
175	174	5	9	7.6	6891	PASS
176	174	95	101	98.8	89141	PASS
177	176	5	9	6.9	6149	PASS