


Data Path : Z:\voasrv\HPCHEM1\MSVOA_F\Data\VF121018\
 Data File : VF060949.D
 Acq On : 10 Dec 2018 12:05
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Title : SW846 8260
 Last Update : Tue Nov 27 01:36:26 2018



AutoFind: Scans 1091, 1092, 1093; Background Corrected with Scan 1085

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	24322	PASS
75	95	30	60	46.3	59461	PASS
95	95	100	100	100.0	128333	PASS
96	95	5	9	6.7	8649	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	68.3	87658	PASS
175	174	5	9	7.5	6617	PASS
176	174	95	101	97.0	84989	PASS
177	176	5	9	6.7	5698	PASS