


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF121118\
Data File : VF060971.D
Acq On : 11 Dec 2018 12:17
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Title : SW846 8260
Last Update : Tue Nov 27 01:36:26 2018



AutoFind: Scans 1092, 1093, 1094; Background Corrected with Scan 1086

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	26141	PASS
75	95	30	60	46.8	64928	PASS
95	95	100	100	100.0	138848	PASS
96	95	5	9	6.4	8853	PASS
173	174	0.00	2	0.4	385	PASS
174	95	50	100	66.6	92442	PASS
175	174	5	9	7.3	6794	PASS
176	174	95	101	99.3	91818	PASS
177	176	5	9	6.3	5739	PASS