

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF121418\
 Data File : VF061016.D
 Acq On : 14 Dec 2018 13:20
 Operator : VA/AP
 Sample : VF1214SBS01
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :

Quant Time: Dec 15 00:04:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	191316	50.00	ug/l	-0.06
34) 1,4-Difluorobenzene	5.58	114	320480	50.00	ug/l	-0.06
63) Chlorobenzene-d5	9.77	117	291053	50.00	ug/l	-0.05
72) 1,4-Dichlorobenzene-d4	12.54	152	150500	50.00	ug/l	-0.04

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.83	65	122799	54.49	ug/l	-0.05
Spiked Amount			Recovery	=	108.98%	
35) Dibromofluoromethane	4.09	113	129390	50.89	ug/l	-0.06
Spiked Amount			Recovery	=	101.78%	
50) Toluene-d8	7.55	98	353758	47.24	ug/l	-0.05
Spiked Amount			Recovery	=	94.48%	
62) 4-Bromofluorobenzene	11.40	95	167938	49.34	ug/l	-0.04
Spiked Amount			Recovery	=	98.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.96	85	56400	19.716	ug/l	99
3) Chloromethane	1.10	50	42064	22.598	ug/l #	81
4) Vinyl Chloride	1.14	62	40399	23.267	ug/l	99
5) Bromomethane	1.32	94	30366	24.802	ug/l	97
6) Chloroethane	1.38	64	19596	26.109	ug/l	99
7) Trichlorofluoromethane	1.49	101	83057	22.295	ug/l #	82
8) Diethyl Ether	1.63	74	13240	24.715	ug/l	87
9) 1,1,2-Trichlorotrifluoroet	1.83	101	45961	24.404	ug/l	88
10) Methyl Iodide	1.90	142	57593	20.118	ug/l	92
11) Tert butyl alcohol	2.57	59	9924	107.417	ug/l #	80
12) 1,1-Dichloroethene	1.79	96	34813	22.981	ug/l	98
13) Acrolein	1.99	56	6375	72.829	ug/l	100
14) Allyl chloride	2.08	41	42804	22.670	ug/l	96
15) Acrylonitrile	2.93	53	24645	118.986	ug/l	93
16) Acetone	2.23	43	52109	125.380	ug/l	94
17) Carbon Disulfide	1.83	76	95689	23.559	ug/l #	93
18) Methyl Acetate	2.34	43	19315	26.940	ug/l	96
19) Methyl tert-butyl Ether	2.42	73	74057	21.706	ug/l	95
20) Methylene Chloride	2.24	84	25343	16.622	ug/l	92
21) trans-1,2-Dichloroethene	2.31	96	27818	19.048	ug/l	87
22) Diisopropyl ether	2.79	45	124153	24.157	ug/l	98
23) Vinyl Acetate	3.18	43	291878	124.240	ug/l	99
24) 1,1-Dichloroethane	2.86	63	73132	23.917	ug/l	99
25) 2-Butanone	4.32	43	91456	127.224	ug/l	95
26) 2,2-Dichloropropane	3.59	77	45481	22.942	ug/l	97
27) cis-1,2-Dichloroethene	3.46	96	56579	23.797	ug/l	89
28) Bromochloromethane	3.71	49	34912	23.790	ug/l	97
29) Tetrahydrofuran	4.06	42	34749	114.039	ug/l	98
30) Chloroform	3.85	83	106406	22.610	ug/l	99
31) Cyclohexane	3.67	56	47401	14.937	ug/l	91
32) 1,1,1-Trichloroethane	4.08	97	63857	19.565	ug/l	98
36) 1,1-Dichloropropene	4.26	75	81954	23.055	ug/l	96
37) Ethyl Acetate	4.09	43	31080	21.516	ug/l #	93
38) Carbon Tetrachloride	3.98	117	65641	20.907	ug/l	96

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF121418\
 Data File : VF061016.D
 Acq On : 14 Dec 2018 13:20
 Operator : VA/AP
 Sample : VF1214SBS01
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :

Quant Time: Dec 15 00:04:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.45	83	78727	21.134	ug/l	90
40) Benzene	4.62	78	184559	23.502	ug/l	98
41) Methacrylonitrile	4.74	41	19375	23.879	ug/l #	87
42) 1,2-Dichloroethane	4.93	62	67254	23.458	ug/l	100
43) Isopropyl Acetate	6.95	43	41927	22.002	ug/l #	98
44) Trichloroethene	5.48	130	56222	21.896	ug/l	91
45) 1,2-Dichloropropane	6.22	63	49596	24.317	ug/l #	89
46) Dibromomethane	6.08	93	29477	21.306	ug/l	98
47) Bromodichloromethane	6.37	83	79636	21.672	ug/l #	95
48) Methyl methacrylate	6.71	41	27219	20.817	ug/l	98
49) 1,4-Dioxane	6.70	88	3644	380.793	ug/l	96
51) 4-Methyl-2-Pentanone	8.25	43	146014	109.614	ug/l	94
52) Toluene	7.62	92	114835	20.818	ug/l	93
53) t-1,3-Dichloropropene	8.27	75	66270	22.171	ug/l	94
54) cis-1,3-Dichloropropene	7.29	75	82809	22.379	ug/l	100
55) 1,1,2-Trichloroethane	8.47	97	34739	21.613	ug/l	90
56) Ethyl methacrylate	8.60	69	37394	20.943	ug/l	97
57) 1,3-Dichloropropane	8.84	76	62972	22.394	ug/l	92
58) 2-Chloroethyl Vinyl ether	7.28	63	41286	226.062	ug/l	91
59) 2-Hexanone	9.48	43	102509	109.690	ug/l	100
60) Dibromochloromethane	8.70	129	52186	21.791	ug/l	97
61) 1,2-Dibromoethane	8.98	107	37843	21.795	ug/l	100
64) Tetrachloroethene	8.14	164	51259	23.058	ug/l	95
65) Chlorobenzene	9.79	112	131815	23.280	ug/l	100
66) 1,1,1,2-Tetrachloroethane	9.92	131	55774	22.913	ug/l	96
67) Ethyl Benzene	9.88	91	254371	24.131	ug/l	99
68) m/p-Xylenes	10.12	106	169459	44.698	ug/l	96
69) o-Xylene	10.70	106	99527	23.807	ug/l	93
70) Styrene	10.77	104	134409	23.193	ug/l	98
71) Bromoform	10.76	173	24736	21.038	ug/l	95
73) Isopropylbenzene	11.12	105	295270	24.826	ug/l	99
74) N-amyl acetate	11.36	43	67710	24.655	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.69	83	47613	24.325	ug/l	93
76) 1,2,3-Trichloropropane	11.86	75	21706	15.256	ug/l	96
77) Bromobenzene	11.48	156	60957	23.378	ug/l	82
78) n-propylbenzene	11.58	91	334502	24.017	ug/l	99
79) 2-Chlorotoluene	11.71	91	196981	24.665	ug/l	98
80) 1,3,5-Trimethylbenzene	11.82	105	236151	24.458	ug/l	98
81) trans-1,4-Dichloro-2-buten	11.86	75	20721	29.725	ug/l	98
82) 4-Chlorotoluene	11.89	91	198777	23.651	ug/l	97
83) tert-Butylbenzene	12.12	119	239930	24.395	ug/l	98
84) 1,2,4-Trimethylbenzene	12.19	105	229031	23.768	ug/l	99
85) sec-Butylbenzene	12.30	105	317795	23.604	ug/l	99
86) p-Isopropyltoluene	12.45	119	264130	23.400	ug/l	97
87) 1,3-Dichlorobenzene	12.47	146	117375	23.665	ug/l	91
88) 1,4-Dichlorobenzene	12.56	146	111083	22.982	ug/l	97
89) n-Butylbenzene	12.83	91	285121	24.980	ug/l	97
90) Hexachloroethane	12.91	117	65900	24.245	ug/l	86
91) 1,2-Dichlorobenzene	12.93	146	109248	23.260	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.63	75	7230	19.105	ug/l	80


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF121418\
Data File : VF061016.D
Acq On : 14 Dec 2018 13:20
Operator : VA/AP
Sample : VF1214SBS01
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :

Quant Time: Dec 15 00:04:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.20	180	79235	24.144	ug/l	95
94) Hexachlorobutadiene	14.19	225	51544	23.560	ug/l	98
95) Naphthalene	14.45	128	132129	18.542	ug/l	98
96) 1,2,3-Trichlorobenzene	14.59	180	70634	23.180	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.85 min Scan# 428

Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion:168 Resp: 191316

Ion Ratio Lower Upper

168 100

99 68.0 52.6 79.0



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.85

60000



40000

20000

0

4.80 5.00

Time-->



#2

Dichlorodifluoromethane

Concen: 19.716 ug/l

RT: 0.96 min Scan# 33

Delta R.T. -0.02 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 85 Resp: 56400

Ion Ratio Lower Upper

85 100

87 33.0 16.7 50.0



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.96

15000


10000


5000

0

0.80 0.90 1.00 1.10 1.20

Time-->





#3

Chloromethane

Concen: 22.598 ug/l

RT: 1.10 min Scan# 47

Delta R.T. -0.01 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion: 50 Resp: 42064

Ion Ratio Lower Upper




50 100

52 23.1 27.0 40.4#



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#4

Vinyl Chloride

Concen: 23.267 ug/l

RT: 1.14 min Scan# 52

Delta R.T. -0.01 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 62 Resp: 40399

Ion Ratio Lower Upper



62 100


64 33.6 26.4 39.6



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0





#5

Bromomethane

Concen: 24.802 ug/l

RT: 1.32 min Scan# 70

Delta R.T. -0.02 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20





Instrument :
MSVOA_F
ClientSampleId :

Tot Ion: 94 Resp: 30366

Ion Ratio Lower Upper

94 100

96 97.8 75.9 113.9



#6

Chloroethane

Concen: 26.109 ug/l

RT: 1.38 min Scan# 76

Delta R.T. -0.01 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion: 64 Resp: 19596

Ion Ratio Lower Upper

64 100

66 34.3 28.1 42.1





#7

Trichlorofluoromethane

Concen: 22.295 ug/l

RT: 1.49 min Scan# 87

Delta R.T. -0.01 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion:101 Resp: 83057

Ion Ratio Lower Upper




101 100

103 54.9 55.4 83.0#



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 24.715 ug/l

RT: 1.63 min Scan# 101

Delta R.T. -0.02 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 74 Resp: 13240

Ion Ratio Lower Upper



74 100


45 125.4 71.0 212.8



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 24.404 ug/l

RT: 1.83 min Scan# 121

Delta R.T. -0.01 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :





Tot Ion:101 Resp: 45961

Ion Ratio Lower Upper

101 100

85 48.0 42.8 64.2

151 55.9 54.4 81.6



#10

Methyl Iodide

Concen: 20.118 ug/l

RT: 1.90 min Scan# 129

Delta R.T. -0.01 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20




Tgt Ion:142 Resp: 57593


Ion Ratio Lower Upper

142 100

127 56.8 51.9 77.9

141 15.5 12.7 19.1









#11
Tert butyl alcohol
Concen: 107.417 ug/l
RT: 2.57 min Scan# 196
Delta R.T. -0.02 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 59 Resp: 9924
Ion Ratio Lower Upper
59 100
57 24.5 12.7 19.1#



#12
1,1-Dichloroethene
Concen: 22.981 ug/l
RT: 1.79 min Scan# 117
Delta R.T. -0.02 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion: 96 Resp: 34813
Ion Ratio Lower Upper
96 100
61 223.5 177.3 265.9
98 64.6 49.6 74.4









#13
Acrolein
Concen: 72.829 ug/l
RT: 1.99 min Scan# 138
Delta R.T. -0.03 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 56 Resp: 6375
Ion Ratio Lower Upper
56 100
55 76.6 61.2 91.8



#14
Allyl chloride
Concen: 22.670 ug/l
RT: 2.08 min Scan# 147
Delta R.T. -0.04 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion: 41 Resp: 42804
Ion Ratio Lower Upper
41 100
39 119.7 91.7 137.5
76 30.8 23.5 35.3





#15

Acrylonitrile

Concen: 118.986 ug/l

RT: 2.93 min Scan# 233

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 53 Resp: 24645

Ion Ratio Lower Upper




53 100

52 91.5 76.5 114.7

51 42.9 40.7 61.1



Abundance Ion 53.00 (52.70 to 53.70): VF0
Ion 52.00 (51.70 to 52.70): VF0
Ion 51.00 (50.70 to 51.70): VF0



#16

Acetone

Concen: 125.380 ug/l

RT: 2.23 min Scan# 162

Delta R.T. -0.03 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 43 Resp: 52109



Ion Ratio Lower Upper


43 100

58 16.9 11.4 17.2



Abundance Ion 43.00 (42.70 to 43.70): VF0
Ion 58.00 (57.70 to 58.70): VF0





#17

Carbon Disulfide

Concen: 23.559 ug/l

RT: 1.83 min Scan# 121

Delta R.T. -0.03 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F





ClientSampleId :

Tot Ion: 76 Resp: 95689

Ion Ratio Lower Upper

76 100

78 9.9 10.2 15.2#



#18

Methyl Acetate

Concen: 26.940 ug/l

RT: 2.34 min Scan# 173

Delta R.T. -0.03 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion: 43 Resp: 19315

Ion Ratio Lower Upper

43 100

74 14.7 13.3 19.9









#19

Methyl tert-butyl Ether
 Concen: 21.706 ug/l
 RT: 2.42 min Scan# 181
 Delta R.T. -0.04 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Instrument :
 MSVOA_F
 ClientSampled :




Tot Ion: 73 Resp: 74057
 Ion Ratio Lower Upper
 73 100
 57 16.8 15.3 22.9




#20

Methylene Chloride
 Concen: 16.622 ug/l
 RT: 2.24 min Scan# 163
 Delta R.T. -0.02 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion: 84 Resp: 25343
 Ion Ratio Lower Upper
 84 100
 49 149.8 111.3 166.9
 51 41.2 38.0 57.0
 86 66.5 49.8 74.8





#21

trans-1,2-Dichloroethene

Concen: 19.048 ug/l

RT: 2.31 min Scan# 170

Delta R.T. -0.03 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 96 Resp: 27818

Ion Ratio Lower Upper

96 100

61 151.6 134.5 201.7




98 53.7 51.6 77.4



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 24.157 ug/l

RT: 2.79 min Scan# 219

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Tgt Ion: 45 Resp: 124153


Ion Ratio Lower Upper

45 100

43 50.4 41.0 61.6

87 19.6 14.6 22.0

59 9.4 7.5 11.3






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 124.240 ug/l

RT: 3.18 min Scan# 258

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion: 43 Resp: 291878

Ion Ratio Lower Upper




43 100

86 6.5 5.4 8.0



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0



#24

1,1-Dichloroethane

Concen: 23.917 ug/l

RT: 2.86 min Scan# 226

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion: 63 Resp: 73132

Ion Ratio Lower Upper

63 100

98 4.9 2.7 8.1



100 2.8 1.5 4.5




Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#25

2-Butanone

Concen: 127.224 ug/l

RT: 4.32 min Scan# 374

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20





Instrument :
MSVOA_F
ClientSampled :

Tot Ion: 43 Resp: 91456

Ion Ratio Lower Upper

43 100

72 21.7 15.4 23.2



#26

2,2-Dichloropropane

Concen: 22.942 ug/l

RT: 3.59 min Scan# 300

Delta R.T. -0.05 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion: 77 Resp: 45481

Ion Ratio Lower Upper

77 100

97 15.9 8.6 25.8






#27




cis-1,2-Dichloroethene
Concen: 23.797 ug/l
RT: 3.46 min Scan# 287
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
96	100		
61	131.7	0.0	290.6
98	60.8	0.0	137.8




Abundance Ion 96.00 (95.70 to 96.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 98.00 (97.70 to 98.70): VF0





#28


Bromochloromethane
Concen: 23.790 ug/l
RT: 3.71 min Scan# 312
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.4
130	82.8	64.2	96.4



Abundance Ion 49.00 (48.70 to 49.70): VF0
Ion 129.00 (128.70 to 129.70): V
Ion 130.00 (129.70 to 130.70): V





#29

Tetrahydrofuran

Concen: 114.039 ug/l

RT: 4.06 min Scan# 348

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 42 Resp: 34749

Ion Ratio Lower Upper




42 100

72 39.7 31.9 47.9

71 37.7 28.7 43.1



Abundance Ion 42.00 (41.70 to 42.70): VF0
Ion 72.00 (71.70 to 72.70): VF0
Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 22.610 ug/l

RT: 3.85 min Scan# 326

Delta R.T. -0.05 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 83 Resp: 106406



Ion Ratio Lower Upper


83 100

85 67.7 53.4 80.2



Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 14.937 ug/l

RT: 3.67 min Scan# 308

Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 56 Resp: 47401

Ion Ratio Lower Upper




56 100

69 26.7 23.0 34.4

84 78.1 70.4 105.6



Abundance Ion 56.00 (55.70 to 56.70): VF0
Ion 69.00 (68.70 to 69.70): VF0
Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 19.565 ug/l

RT: 4.08 min Scan# 350

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion: 97 Resp: 63857

Ion Ratio Lower Upper



97 100


99 65.6 53.5 80.3

61 46.8 38.6 58.0



Abundance Ion 97.00 (96.70 to 97.70): VF0
Ion 99.00 (98.70 to 99.70): VF0
Ion 61.00 (60.70 to 61.70): VF0









#33
 1,2-Dichloroethane-d4
 Concen: 54.486 ug/l
 RT: 4.83 min Scan# 426
 Delta R.T. -0.05 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion: 65 Resp: 122799
 Ion Ratio Lower Upper
 65 100
 67 47.5 0.0 96.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.58 min Scan# 502
 Delta R.T. -0.06 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion: 114 Resp: 320480
 Ion Ratio Lower Upper
 114 100
 63 21.6 0.0 41.6
 88 19.6 0.0 36.6





#35

Dibromofluoromethane

Concen: 50.892 ug/l

RT: 4.09 min Scan# 351

Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampled :


Tot Ion:113 Resp: 129390

Ion Ratio Lower Upper




113 100

111 102.8 83.0 124.6

192 14.8 12.3 18.5



Abundance Ion 113.00 (112.70 to 113.70): V
Ion 111.00 (110.70 to 111.70): V
Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 23.055 ug/l

RT: 4.26 min Scan# 368

Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion: 75 Resp: 81954

Ion Ratio Lower Upper



75 100


110 28.9 15.6 46.8

77 27.2 23.4 35.0



Abundance Ion 75.00 (74.70 to 75.70): VF0
Ion 110.00 (109.70 to 110.70): V
Ion 77.00 (76.70 to 77.70): VF0









#37
Ethyl Acetate
Concen: 21.516 ug/l
RT: 4.09 min Scan# 351
Delta R.T. -0.06 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 43 Resp: 31080
Ion Ratio Lower Upper
43 100
61 79.5 68.6 103.0
70 7.1 9.0 13.4#



#38
Carbon Tetrachloride
Concen: 20.907 ug/l
RT: 3.98 min Scan# 339
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion: 117 Resp: 65641
Ion Ratio Lower Upper
117 100
119 97.7 74.4 111.6
121 29.9 23.5 35.3





#39

Methylvlcyclohexane

Concen: 21.134 ug/l

RT: 5.45 min Scan# 488

Delta R.T. -0.05 min


Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion: 83 Resp: 78727

Ion	Ratio	Lower	Upper
83	100		
55	80.4	72.1	108.1
98	43.3	39.1	58.7






Abundance

Ion 83.00 (82.70 to 83.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#40

Benzene

Concen: 23.502 ug/l

RT: 4.62 min Scan# 404


Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Tgt Ion: 78 Resp: 184559



Ion	Ratio	Lower	Upper
78	100		
77	24.5	18.7	28.1




Abundance

Ion 78.10 (77.80 to 78.80): VF0

Ion 77.00 (76.70 to 77.70): VF0





#41

Methacrylonitrile

Concen: 23.879 ug/l

RT: 4.74 min Scan# 416

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :

Tot Ion: 41 Resp: 19375





Ion Ratio Lower Upper

41 100

39 50.6 53.0 79.6#

67 53.8 43.8 65.6

52 36.7 37.8 56.8#



#42

1,2-Dichloroethane

Concen: 23.458 ug/l

RT: 4.93 min Scan# 436

Delta R.T. -0.05 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion: 62 Resp: 67254

Ion Ratio Lower Upper

62 100

98 8.8 0.0 17.2





#43

Isopropyl Acetate

Concen: 22.002 ug/l

RT: 6.95 min Scan# 641

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 43 Resp: 41927

Ion Ratio Lower Upper

43 100

61 28.0 21.4 32.0




87 0.0 0.2 0.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 21.896 ug/l

RT: 5.48 min Scan# 492

Delta R.T. -0.06 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion:130 Resp: 56222

Ion Ratio Lower Upper



130 100


95 106.8 0.0 195.6



Abundance Ion 129.90 (129.60 to 130.60): V

Ion 94.90 (94.60 to 95.60): VF0









#45
 1,2-Dichloropropane
 Concen: 24.317 ug/l
 RT: 6.22 min Scan# 567
 Delta R.T. -0.05 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20




Instrument :
 MSVOA_F
 ClientSampled :


Tot Ion: 63 Resp: 49596
 Ion Ratio Lower Upper
 63 100
 65 35.0 23.2 34.8#



#46
 Dibromomethane
 Concen: 21.306 ug/l
 RT: 6.08 min Scan# 552
 Delta R.T. -0.05 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion: 93 Resp: 29477
 Ion Ratio Lower Upper
 93 100
 95 79.6 64.7 97.1
 174 93.8 73.0 109.6





#47

Bromodichloromethane

Concen: 21.672 ug/l

RT: 6.37 min Scan# 582

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 83 Resp: 79636

Ion Ratio Lower Upper

83 100

85 67.6 51.1 76.7




127 6.5 6.8 10.2#



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 20.817 ug/l

RT: 6.71 min Scan# 616

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion: 41 Resp: 27219

Ion Ratio Lower Upper

41 100

69 65.2 53.9 80.9



39 60.5 49.0 73.6




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#49

1,4-Dioxane

Concen: 380.793 ug/l

RT: 6.70 min Scan# 615

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 88 Resp: 3644

Ion Ratio Lower Upper




88 100

43 56.0 42.0 63.0

58 80.4 61.8 92.8



Abundance Ion 88.00 (87.70 to 88.70): VF0
Ion 43.00 (42.70 to 43.70): VF0
Ion 58.00 (57.70 to 58.70): VF0



#50

Toluene-d8

Concen: 47.235 ug/l

RT: 7.55 min Scan# 701

Delta R.T. -0.05 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 98 Resp: 353758



Ion Ratio Lower Upper


98 100

100 71.4 59.3 88.9



Abundance Ion 98.00 (97.70 to 98.70): VF0
Ion 100.00 (99.70 to 100.70): VF0





#51

4-Methyl-2-Pentanone

Concen: 109.614 ug/l

RT: 8.25 min Scan# 772

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion: 43 Resp: 146014

Ion Ratio Lower Upper




43 100

58 40.6 29.7 44.5



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 20.818 ug/l

RT: 7.62 min Scan# 708

Delta R.T. -0.05 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 92 Resp: 114835

Ion Ratio Lower Upper



92 100


91 167.1 126.2 189.2



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0









#53
 t-1,3-Dichloropropene
 Concen: 22.171 ug/l
 RT: 8.27 min Scan# 774
 Delta R.T. -0.05 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion: 75 Resp: 66270
 Ion Ratio Lower Upper
 75 100
 77 30.6 27.4 41.2



#54
 cis-1,3-Dichloropropene
 Concen: 22.379 ug/l
 RT: 7.29 min Scan# 675
 Delta R.T. -0.05 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion: 75 Resp: 82809
 Ion Ratio Lower Upper
 75 100
 77 31.3 25.0 37.6
 39 63.9 51.1 76.7





#55

1,1,2-Trichloroethane

Concen: 21.613 ug/l

RT: 8.47 min Scan# 795

Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :

Tot Ion: 97 Resp: 34739



Ion	Ratio	Lower	Upper
97	100		
83	97.6	67.8	101.6
85	60.5	45.0	67.6
99	69.3	50.5	75.7

97 100

83 97.6 67.8 101.6

85 60.5 45.0 67.6

99 69.3 50.5 75.7





Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 20.943 ug/l

RT: 8.60 min Scan# 808

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20



Tgt Ion: 69 Resp: 37394

Ion	Ratio	Lower	Upper
69	100		
41	81.5	62.6	94.0
39	46.1	37.8	56.6

69 100

41 81.5 62.6 94.0


39 46.1 37.8 56.6




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 22.394 ug/l

RT: 8.84 min Scan# 832

Delta R.T. -0.06 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F





ClientSampled :

Tot Ion: 76 Resp: 62972

Ion Ratio Lower Upper

76 100

78 29.2 27.0 40.4



#58

2-Chloroethyl Vinyl ether

Concen: 226.062 ug/l

RT: 7.28 min Scan# 674

Delta R.T. -0.06 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion: 63 Resp: 41286

Ion Ratio Lower Upper

63 100

106 23.8 23.0 34.6









#59
2-Hexanone
Concen: 109.690 ug/l
RT: 9.48 min Scan# 897
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampled :


Tot Ion: 43 Resp: 102509
Ion Ratio Lower Upper
43 100
58 50.9 25.6 76.6



#60
Dibromochloromethane
Concen: 21.791 ug/l
RT: 8.70 min Scan# 818
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion:129 Resp: 52186
Ion Ratio Lower Upper
129 100
127 71.6 37.3 111.9





#61

1,2-Dibromoethane

Concen: 21.795 ug/l

RT: 8.98 min Scan# 846

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion:107 Resp: 37843

Ion Ratio Lower Upper




107 100

109 92.2 73.6 110.4



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 49.339 ug/l

RT: 11.40 min Scan# 1092

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion: 95 Resp: 167938

Ion Ratio Lower Upper

95 100

174 76.2 0.0 138.2



176 70.3 0.0 146.2




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.77 min Scan# 926
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion:117 Resp: 291053
Ion Ratio Lower Upper
117 100
82 56.8 46.6 69.8
119 33.3 26.4 39.6



#64
Tetrachloroethene
Concen: 23.058 ug/l
RT: 8.14 min Scan# 761
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion:164 Resp: 51259
Ion Ratio Lower Upper
164 100
166 119.4 100.3 150.5
129 91.7 77.9 116.9
131 100.5 77.5 116.3









#65
Chlorobenzene
Concen: 23.280 ug/l
RT: 9.79 min Scan# 928
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion:112 Resp: 131815
Ion Ratio Lower Upper
112 100
114 32.3 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 22.913 ug/l
RT: 9.92 min Scan# 942
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion:131 Resp: 55774
Ion Ratio Lower Upper
131 100
133 91.6 46.5 139.3
119 67.5 31.1 93.3









#67
Ethyl Benzene
Concen: 24.131 ug/l
RT: 9.88 min Scan# 938
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20




Instrument :
MSVOA_F
ClientSampled :


Tot Ion: 91 Resp: 254371
Ion Ratio Lower Upper
91 100
106 28.4 22.2 33.2



#68
m/p-Xylenes
Concen: 44.698 ug/l
RT: 10.12 min Scan# 962
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion: 106 Resp: 169459
Ion Ratio Lower Upper
106 100
91 214.8 166.7 250.1









#69
o-Xylene
Concen: 23.807 ug/l
RT: 10.70 min Scan# 1021
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:106 Resp: 99527
Ion Ratio Lower Upper
106 100
91 221.1 104.8 314.6




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VF0






#70
Styrene
Concen: 23.193 ug/l
RT: 10.77 min Scan# 1028
Delta R.T. -0.05 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion:104 Resp: 134409
Ion Ratio Lower Upper
104 100
78 49.2 37.3 55.9
103 45.5 36.8 55.2



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VF0
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 21.038 ug/l

RT: 10.76 min Scan# 1027

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:173 Resp: 24736

Ion Ratio Lower Upper

173 100

175 47.6 25.5 76.5




254 7.6 7.5 11.3



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.54 min Scan# 1207

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion:152 Resp: 150500

Ion Ratio Lower Upper

152 100

115 65.0 30.1 90.5



150 155.4 0.0 327.6




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 24.826 ug/l

RT: 11.12 min Scan# 1063

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion:105 Resp: 295270

Ion Ratio Lower Upper




105 100

120 25.0 12.2 36.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 24.655 ug/l

RT: 11.36 min Scan# 1088

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Tgt Ion: 43 Resp: 67710


Ion Ratio Lower Upper

43 100

70 33.6 29.6 44.4

55 20.1 17.2 25.8

61 25.0 21.8 32.6






Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0





#75

1,1,2,2-Tetrachloroethane

Concen: 24.325 ug/l

RT: 11.69 min Scan# 1121

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 83 Resp: 47613

Ion Ratio Lower Upper

83 100

131 8.8 4.7 14.1




85 71.1 32.6 97.8



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): V

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 15.256 ug/l

RT: 11.86 min Scan# 1138

Delta R.T. 0.03 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 75 Resp: 21706

Ion Ratio Lower Upper



75 100


77 39.3 18.4 55.4



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 23.378 ug/l

RT: 11.48 min Scan# 1100

Delta R.T. -0.05 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:156 Resp: 60957

Ion Ratio Lower Upper

156 100

77 174.8 69.7 209.0




158 95.1 47.8 143.4



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 24.017 ug/l

RT: 11.58 min Scan# 1110

Delta R.T. -0.04 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 91 Resp: 334502

Ion Ratio Lower Upper



91 100


120 20.0 10.3 30.8



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 24.665 ug/l

RT: 11.71 min Scan# 1123

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F





ClientSampled :

Tot Ion: 91 Resp: 196981

Ion Ratio Lower Upper

91 100

126 30.6 15.9 47.6



#80

1,3,5-Trimethylbenzene

Concen: 24.458 ug/l

RT: 11.82 min Scan# 1134

Delta R.T. -0.04 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion:105 Resp: 236151

Ion Ratio Lower Upper

105 100

120 50.3 25.8 77.3





#81

trans-1,4-Dichloro-2-butene

Concen: 29.725 ug/l

RT: 11.86 min Scan# 1138

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampled :





Tot Ion: 75 Resp: 20721

Ion Ratio Lower Upper

75 100

53 82.5 65.9 98.9

89 51.5 38.8 58.2



#82

4-Chlorotoluene

Concen: 23.651 ug/l

RT: 11.89 min Scan# 1141

Delta R.T. -0.04 min

Lab File: VF061016.D




Acq: 14 Dec 2018 13:20


Tgt Ion: 91 Resp: 198777

Ion Ratio Lower Upper

91 100

126 31.1 14.8 44.4









#83
 tert-Butylbenzene
 Concen: 24.395 ug/l
 RT: 12.12 min Scan# 1165
 Delta R.T. -0.04 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion:119 Resp: 239930
 Ion Ratio Lower Upper
 119 100
 91 59.5 30.1 90.3
 134 23.4 12.6 37.8



#84
 1,2,4-Trimethylbenzene
 Concen: 23.768 ug/l
 RT: 12.19 min Scan# 1172
 Delta R.T. -0.04 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion:105 Resp: 229031
 Ion Ratio Lower Upper
 105 100
 120 47.0 23.8 71.5









#85
 sec-Butylbenzene
 Concen: 23.604 ug/l
 RT: 12.30 min Scan# 1183
 Delta R.T. -0.03 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20




Instrument :
 MSVOA_F
 ClientSampled :


Tot Ion:105 Resp: 317795
 Ion Ratio Lower Upper
 105 100
 134 19.5 9.5 28.5



#86
 p-Isopropyltoluene
 Concen: 23.400 ug/l
 RT: 12.45 min Scan# 1198
 Delta R.T. -0.04 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion:119 Resp: 264130
 Ion Ratio Lower Upper
 119 100
 134 24.4 12.9 38.6
 91 25.7 11.8 35.3









#87
 1,3-Dichlorobenzene
 Concen: 23.665 ug/l
 RT: 12.47 min Scan# 1200
 Delta R.T. -0.04 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
146	100		
111	38.8	20.0	59.9
148	55.0	32.9	98.6



#88
 1,4-Dichlorobenzene
 Concen: 22.982 ug/l
 RT: 12.56 min Scan# 1209
 Delta R.T. -0.04 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion	Ratio	Lower	Upper
146	100		
111	40.3	18.9	56.7
148	66.8	32.3	96.9





#89

n-Butylbenzene

Concen: 24.980 ug/l

RT: 12.83 min Scan# 1237

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampled :


Tot Ion: 91 Resp: 285121

Ion Ratio Lower Upper

91 100

92 63.3 31.6 94.7




134 22.9 13.8 41.4



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 92.00 (91.70 to 92.70): VF0

Ion 134.00 (133.70 to 134.70): V



#90

Hexachloroethane

Concen: 24.245 ug/l

RT: 12.91 min Scan# 1245

Delta R.T. -0.03 min

Lab File: VF061016.D


Acq: 14 Dec 2018 13:20

Tgt Ion: 117 Resp: 65900

Ion Ratio Lower Upper



117 100


201 63.1 26.7 80.1



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 201.00 (200.70 to 201.70): V





#91

1,2-Dichlorobenzene

Concen: 23.260 ug/l

RT: 12.93 min Scan# 1247

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:146 Resp: 109248

Ion Ratio Lower Upper

146 100

111 41.3 21.2 63.6




148 63.1 32.5 97.5



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.105 ug/l

RT: 13.63 min Scan# 1318

Delta R.T. -0.04 min

Lab File: VF061016.D

Acq: 14 Dec 2018 13:20


Tgt Ion: 75 Resp: 7230

Ion Ratio Lower Upper

75 100

155 96.8 45.6 136.9



157 93.6 63.8 191.5




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V






#93
 1,2,4-Trichlorobenzene
 Concen: 24.144 ug/l
 RT: 14.20 min Scan# 1375
 Delta R.T. -0.03 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Instrument :
 MSVOA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
180	100		
182	93.9	48.5	145.5
145	29.6	18.1	54.3






Abundance

Ion 180.00 (179.70 to 180.70): V


Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V



#94
 Hexachlorobutadiene
 Concen: 23.560 ug/l
 RT: 14.19 min Scan# 1374
 Delta R.T. -0.03 min
 Lab File: VF061016.D
 Acq: 14 Dec 2018 13:20

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.9	31.4	94.3
227	62.2	32.5	97.5






Abundance

Ion 225.00 (224.70 to 225.70): V

Ion 223.00 (222.70 to 223.70): V

Ion 227.00 (226.70 to 227.70): V









#95
Naphthalene
Concen: 18.542 ug/l
RT: 14.45 min Scan# 1401
Delta R.T. -0.03 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Instrument :
MSVOA_F
ClientSampled :

Tot Ion	128	Resp	132	129
Ion Ratio	100			
Lower	9.9			
Upper	14.9			



#96
1,2,3-Trichlorobenzene
Concen: 23.180 ug/l
RT: 14.59 min Scan# 1415
Delta R.T. -0.04 min
Lab File: VF061016.D
Acq: 14 Dec 2018 13:20

Tgt Ion	180	Resp	70	63
Ion Ratio	100			
Lower	47.9			
Upper	143.8			

