

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF121718\
 Data File : VF061042.D
 Acq On : 17 Dec 2018 16:00
 Operator : VA/AP
 Sample : J6319-12
 Misc : 4.90g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 066_SW-W-2

Manual Integrations
 APPROVED

apatel
 12/18/2018 2:32:25 PM

Quant Time: Dec 18 03:35:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	214505	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.62	114	363913	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.79	117	245595	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.56	152	108151	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	128139	50.71	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	101.42%	
35) Dibromofluoromethane	4.13	113	139814	48.43	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	96.86%	
50) Toluene-d8	7.57	98	303965	35.74	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	71.48%	
62) 4-Bromofluorobenzene	11.43	95	208002	53.82	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	107.64%	
Target Compounds						
16) Acetone	2.25	43	13883	29.793	ug/l	95
20) Methylene Chloride	2.26	84	4970m	2.907	ug/l	
84) 1,2,4-Trimethylbenzene	12.22	105	9313	1.345	ug/l	86
85) sec-Butylbenzene	12.32	105	39024	4.033	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF121718\
Data File : VF061042.D
Acq On : 17 Dec 2018 16:00
Operator : VA/AP
Sample : J6319-12
Misc : 4.90g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
066_SW-W-2

Manual Integrations
APPROVED

apatel
12/18/2018 2:32:25 PM

Quant Time: Dec 18 03:35:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

