

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF121718\
 Data File : VF061045.D
 Acq On : 17 Dec 2018 17:27
 Operator : VA/AP
 Sample : J6319-12RE
 Misc : 4.51g/5mL/MSVOA-F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 066_SW-W-2RE

Manual Integrations
 APPROVED

apatel
 12/18/2018 2:32:36 PM

Quant Time: Dec 18 03:40:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	250224	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.61	114	430647	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.80	117	340592	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.57	152	167678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	108023	36.65	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	73.30%	
35) Dibromofluoromethane	4.13	113	123843	36.25	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	72.50%	
50) Toluene-d8	7.57	98	293950	29.21	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	58.42%	
62) 4-Bromofluorobenzene	11.42	95	131708	28.80	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	57.60%	
Target Compounds						
16) Acetone	2.24	43	12341m	22.703	ug/l	Qvalue
20) Methylene Chloride	2.26	84	5366m	2.691	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF121718\
Data File : VF061045.D
Acq On : 17 Dec 2018 17:27
Operator : VA/AP
Sample : J6319-12RE
Misc : 4.51g/5mL/MSVOA-F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
066_SW-W-2RE

Manual Integrations
APPROVED

apatel
12/18/2018 2:32:36 PM

Quant Time: Dec 18 03:40:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

