

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF121818\
 Data File : VF061064.D
 Acq On : 18 Dec 2018 15:55
 Operator : VA/AP
 Sample : VF1218SBS01
 Misc : 5.00µg/5mL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01

Quant Time: Dec 19 00:51:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	167810	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	5.60	114	286860	50.00	ug/l	-0.04
63) Chlorobenzene-d5	9.78	117	266065	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.55	152	138556	50.00	ug/l	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.85	65	103370	52.29	ug/l	-0.03
Spiked Amount			Recovery	=	104.58%	
35) Dibromofluoromethane	4.11	113	113842	50.02	ug/l	-0.04
Spiked Amount			Recovery	=	100.04%	
50) Toluene-d8	7.55	98	295845	44.13	ug/l	-0.04
Spiked Amount			Recovery	=	88.26%	
62) 4-Bromofluorobenzene	11.41	95	147817	48.52	ug/l	-0.03
Spiked Amount			Recovery	=	97.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.96	85	56115	22.365	ug/l	93
3) Chloromethane	1.11	50	41177	25.220	ug/l #	87
4) Vinyl Chloride	1.15	62	39768	26.112	ug/l	99
5) Bromomethane	1.33	94	28932	26.941	ug/l	99
6) Chloroethane	1.40	64	15953	24.233	ug/l	90
7) Trichlorofluoromethane	1.50	101	76334	23.361	ug/l	91
8) Diethyl Ether	1.63	74	10987	23.382	ug/l	97
9) 1,1,2-Trichlorotrifluoroet	1.85	101	39907	24.157	ug/l	97
10) Methyl Iodide	1.92	142	53691	21.382	ug/l	97
11) Tert butyl alcohol	2.58	59	10009	123.513	ug/l #	89
12) 1,1-Dichloroethene	1.80	96	29904	22.506	ug/l #	75
13) Acrolein	2.01	56	5232	68.144	ug/l	90
14) Allyl chloride	2.10	41	33661	20.325	ug/l	93
15) Acrylonitrile	2.95	53	24615	135.488	ug/l	93
16) Acetone	2.25	43	49203	134.971	ug/l	97
17) Carbon Disulfide	1.84	76	62243	17.471	ug/l	94
18) Methyl Acetate	2.35	43	21522	34.685	ug/l	97
19) Methyl tert-butyl Ether	2.43	73	75571	25.252	ug/l	100
20) Methylene Chloride	2.27	84	23925	17.890	ug/l	95
21) trans-1,2-Dichloroethene	2.31	96	22688	17.711	ug/l	93
22) Diisopropyl ether	2.81	45	114902	25.488	ug/l	97
23) Vinyl Acetate	3.19	43	287123	139.336	ug/l	98
24) 1,1-Dichloroethane	2.88	63	63233	23.576	ug/l	99
25) 2-Butanone	4.35	43	86544	137.255	ug/l	98
26) 2,2-Dichloropropane	3.60	77	42408	24.389	ug/l	98
27) cis-1,2-Dichloroethene	3.48	96	54966	26.356	ug/l	96
28) Bromochloromethane	3.72	49	34510	26.810	ug/l	91
29) Tetrahydrofuran	4.08	42	35668	133.451	ug/l	97
30) Chloroform	3.85	83	106116	25.707	ug/l	96
31) Cyclohexane	3.70	56	39639	14.241	ug/l	97
32) 1,1,1-Trichloroethane	4.09	97	65275	22.801	ug/l	93
36) 1,1-Dichloropropene	4.28	75	77608	24.392	ug/l	97
37) Ethyl Acetate	4.12	43	34376	26.587	ug/l #	72
38) Carbon Tetrachloride	3.99	117	58636	20.865	ug/l	90

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF121818\
 Data File : VF061064.D
 Acq On : 18 Dec 2018 15:55
 Operator : VA/AP
 Sample : VF1218SBS01
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01

Quant Time: Dec 19 00:51:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.45	83	64099	19.223	ug/l	96
40) Benzene	4.63	78	170562	24.265	ug/l	97
41) Methacrylonitrile	4.75	41	17522	24.170	ug/l #	88
42) 1,2-Dichloroethane	4.95	62	64588	25.169	ug/l	94
43) Isopropyl Acetate	6.96	43	40177	23.757	ug/l #	100
44) Trichloroethene	5.50	130	55353	24.084	ug/l	94
45) 1,2-Dichloropropane	6.24	63	44008	24.106	ug/l	94
46) Dibromomethane	6.08	93	28101	22.691	ug/l	95
47) Bromodichloromethane	6.39	83	81376	24.740	ug/l	99
48) Methyl methacrylate	6.72	41	25921	22.148	ug/l #	80
49) 1,4-Dioxane	6.71	88	3522	411.179	ug/l #	83
51) 4-Methyl-2-Pentanone	8.26	43	144168	120.912	ug/l	98
52) Toluene	7.62	92	107863	21.846	ug/l	97
53) t-1,3-Dichloropropene	8.28	75	63295	23.658	ug/l	94
54) cis-1,3-Dichloropropene	7.30	75	80113	24.188	ug/l	97
55) 1,1,2-Trichloroethane	8.49	97	33727	23.443	ug/l	98
56) Ethyl methacrylate	8.62	69	35316	22.097	ug/l	96
57) 1,3-Dichloropropane	8.85	76	61170	24.302	ug/l	97
58) 2-Chloroethyl Vinyl ether	7.30	63	39561	242.004	ug/l	98
59) 2-Hexanone	9.50	43	98307	117.522	ug/l	98
60) Dibromochloromethane	8.72	129	47301	22.066	ug/l	96
61) 1,2-Dibromoethane	8.99	107	36838	23.702	ug/l	96
64) Tetrachloroethene	8.15	164	48667	23.948	ug/l	96
65) Chlorobenzene	9.80	112	128415	24.810	ug/l	97
66) 1,1,1,2-Tetrachloroethane	9.93	131	52332	23.518	ug/l	92
67) Ethyl Benzene	9.90	91	242041	25.118	ug/l	98
68) m/p-Xylenes	10.14	106	164958	47.597	ug/l	99
69) o-Xylene	10.71	106	87186	22.813	ug/l	86
70) Styrene	10.79	104	127799	24.124	ug/l	97
71) Bromoform	10.78	173	24274	22.584	ug/l	96
73) Isopropylbenzene	11.12	105	272975	24.930	ug/l	98
74) N-amyl acetate	11.37	43	62567	24.746	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.70	83	47144	26.161	ug/l	93
76) 1,2,3-Trichloropropane	11.79	75	36424	27.808	ug/l	92
77) Bromobenzene	11.50	156	57467	23.939	ug/l	92
78) n-propylbenzene	11.60	91	322940	25.185	ug/l	98
79) 2-Chlorotoluene	11.72	91	186529	25.370	ug/l	98
80) 1,3,5-Trimethylbenzene	11.82	105	227357	25.578	ug/l	95
81) trans-1,4-Dichloro-2-buten	11.87	75	19633	30.592	ug/l	93
82) 4-Chlorotoluene	11.90	91	192150	24.834	ug/l	96
83) tert-Butylbenzene	12.13	119	231713	25.591	ug/l	99
84) 1,2,4-Trimethylbenzene	12.21	105	227313	25.623	ug/l	97
85) sec-Butylbenzene	12.31	105	308512	24.890	ug/l	99
86) p-Isopropyltoluene	12.46	119	254201	24.461	ug/l	99
87) 1,3-Dichlorobenzene	12.47	146	111572	24.434	ug/l	99
88) 1,4-Dichlorobenzene	12.56	146	103802	23.327	ug/l	95
89) n-Butylbenzene	12.84	91	272964	25.977	ug/l	95
90) Hexachloroethane	12.92	117	61746	24.675	ug/l	91
91) 1,2-Dichlorobenzene	12.94	146	105456	24.388	ug/l	96
92) 1,2-Dibromo-3-Chloropropan	13.64	75	8857	25.422	ug/l	62

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF121818\
Data File : VF061064.D
Acq On : 18 Dec 2018 15:55
Operator : VA/AP
Sample : VF1218SBS01
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01

Quant Time: Dec 19 00:51:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.20	180	75379	24.949	ug/l	97
94) Hexachlorobutadiene	14.19	225	48518	24.089	ug/l	97
95) Naphthalene	14.46	128	132721	20.573	ug/l	100
96) 1,2,3-Trichlorobenzene	14.61	180	66720	23.783	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF121818\
 Data File : VF061064.D
 Acq On : 18 Dec 2018 15:55
 Operator : VA/AP
 Sample : VF1218SBS01
 Misc : 5.00µL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01

Quant Time: Dec 19 00:51:43 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.87 min Scan# 430

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion:168 Resp: 167810

Ion Ratio Lower Upper

168 100

99 77.7 52.6 79.0



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.87

60000

50000

40000

30000

20000



10000

0

Time-->

4.80

5.00



#2

Dichlorodifluoromethane

Concen: 22.365 ug/l

RT: 0.96 min Scan# 34

Delta R.T. -0.01 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 85 Resp: 56115

Ion Ratio Lower Upper

85 100

87 29.3 16.7 50.0



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.96

15000

10000

5000


0

Time-->

0.90

1.00

1.10





#3

Chloromethane

Concen: 25.220 ug/l

RT: 1.11 min Scan# 49

Delta R.T. 0.01 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion: 50 Resp: 41177

Ion Ratio Lower Upper

50 100

52 26.4 27.0 40.4#



#4

Vinyl Chloride

Concen: 26.112 ug/l

RT: 1.15 min Scan# 53

Delta R.T. -0.00 min

Lab File: VF061064.D




Acq: 18 Dec 2018 15:55


Tgt Ion: 62 Resp: 39768

Ion Ratio Lower Upper

62 100

64 32.5 26.4 39.6





#5

Bromomethane

Concen: 26.941 ug/l

RT: 1.33 min Scan# 71

Delta R.T. -0.01 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion: 94 Resp: 28932

Ion Ratio Lower Upper

94 100

96 93.7 75.9 113.9



#6

Chloroethane

Concen: 24.233 ug/l

RT: 1.40 min Scan# 78

Delta R.T. 0.01 min

Lab File: VF061064.D




Acq: 18 Dec 2018 15:55

Tgt Ion: 64 Resp: 15953

Ion Ratio Lower Upper

64 100

66 41.2 28.1 42.1





#7

Trichlorofluoromethane

Concen: 23.361 ug/l

RT: 1.50 min Scan# 88

Delta R.T. -0.00 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion:101 Resp: 76334

Ion Ratio Lower Upper




101 100

103 62.1 55.4 83.0



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 23.382 ug/l

RT: 1.63 min Scan# 102

Delta R.T. -0.01 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 74 Resp: 10987

Ion Ratio Lower Upper



74 100


45 138.4 71.0 212.8



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 24.157 ug/l

RT: 1.85 min Scan# 124

Delta R.T. 0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:101 Resp: 39907

Ion Ratio Lower Upper

101 100

85 54.2 42.8 64.2




151 63.9 54.4 81.6



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 21.382 ug/l

RT: 1.92 min Scan# 131

Delta R.T. 0.01 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion:142 Resp: 53691

Ion Ratio Lower Upper

142 100

127 67.5 51.9 77.9



141 16.1 12.7 19.1




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11
Tert butyl alcohol
Concen: 123.513 ug/l
RT: 2.58 min Scan# 198
Delta R.T. -0.00 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion: 59 Resp: 10009
Ion Ratio Lower Upper
59 100
57 11.4 12.7 19.1#



#12
1,1-Dichloroethene
Concen: 22.506 ug/l
RT: 1.80 min Scan# 119
Delta R.T. -0.00 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion: 96 Resp: 29904
Ion Ratio Lower Upper
96 100
61 268.0 177.3 265.9#
98 70.8 49.6 74.4





#13

Acrolein

Concen: 68.144 ug/l

RT: 2.01 min Scan# 140

Delta R.T. -0.01 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion: 56 Resp: 5232

Ion Ratio Lower Upper

56 100

55 85.2 61.2 91.8



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

2.01

2000

1500



1000

500

0

1.90 1.95 2.00 2.05 2.10

Time-->



#14

Allyl chloride

Concen: 20.325 ug/l

RT: 2.10 min Scan# 149

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 41 Resp: 33661

Ion Ratio Lower Upper

41 100

39 123.4 91.7 137.5

76 31.6 23.5 35.3



Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

Ion 76.00 (75.70 to 76.70): VF0

25000

20000

15000


10000


5000

0

2.00 2.05 2.10 2.15

Time-->





#15

Acrylonitrile

Concen: 135.488 ug/l

RT: 2.95 min Scan# 235

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion: 53 Resp: 24615

Ion Ratio Lower Upper

53 100

52 90.0 76.5 114.7




51 44.8 40.7 61.1



Abundance Ion 53.00 (52.70 to 53.70): VF0

Ion 52.00 (51.70 to 52.70): VF0

Ion 51.00 (50.70 to 51.70): VF0



#16

Acetone

Concen: 134.971 ug/l

RT: 2.25 min Scan# 164

Delta R.T. -0.01 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 43 Resp: 49203

Ion Ratio Lower Upper



43 100


58 15.5 11.4 17.2



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0









#17
 Carbon Disulfide
 Concen: 17.471 ug/l
 RT: 1.84 min Scan# 123
 Delta R.T. -0.01 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55




Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01


Tot Ion: 76 Resp: 62243
 Ion Ratio Lower Upper
 76 100
 78 10.4 10.2 15.2



#18
 Methyl Acetate
 Concen: 34.685 ug/l
 RT: 2.35 min Scan# 175
 Delta R.T. -0.01 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion: 43 Resp: 21522
 Ion Ratio Lower Upper
 43 100
 74 17.8 13.3 19.9









#19

Methyl tert-butyl Ether
 Concen: 25.252 ug/l
 RT: 2.43 min Scan# 183
 Delta R.T. -0.02 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01




Tot Ion: 73 Resp: 75571
 Ion Ratio Lower Upper
 73 100
 57 19.0 15.3 22.9




#20

Methylene Chloride
 Concen: 17.890 ug/l
 RT: 2.27 min Scan# 166
 Delta R.T. 0.01 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion: 84 Resp: 23925
 Ion Ratio Lower Upper
 84 100
 49 147.0 111.3 166.9
 51 43.2 38.0 57.0
 86 61.5 49.8 74.8





#21

trans-1,2-Dichloroethene

Concen: 17.711 ug/l

RT: 2.31 min Scan# 171

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01





Tot Ion: 96 Resp: 22688

Ion Ratio Lower Upper

96 100

61 158.1 134.5 201.7

98 60.9 51.6 77.4



#22

Diisopropyl ether

Concen: 25.488 ug/l

RT: 2.81 min Scan# 221

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Tgt Ion: 45 Resp: 114902




Ion Ratio Lower Upper


45 100

43 53.6 41.0 61.6

87 19.8 14.6 22.0

59 9.0 7.5 11.3





#23

Vinyl Acetate

Concen: 139.336 ug/l

RT: 3.19 min Scan# 260

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion: 43 Resp: 287123

Ion Ratio Lower Upper




43 100

86 7.2 5.4 8.0



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0



#24

1,1-Dichloroethane

Concen: 23.576 ug/l

RT: 2.88 min Scan# 228

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 63 Resp: 63233

Ion Ratio Lower Upper

63 100

98 4.9 2.7 8.1



100 2.7 1.5 4.5




Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#25

2-Butanone

Concen: 137.255 ug/l

RT: 4.35 min Scan# 377

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion: 43 Resp: 86544

Ion Ratio Lower Upper

43 100

72 20.1 15.4 23.2



#26

2,2-Dichloropropane

Concen: 24.389 ug/l

RT: 3.60 min Scan# 301

Delta R.T. -0.04 min

Lab File: VF061064.D




Acq: 18 Dec 2018 15:55


Tgt Ion: 77 Resp: 42408

Ion Ratio Lower Upper

77 100

97 18.2 8.6 25.8






#27




cis-1,2-Dichloroethene
Concen: 26.356 ug/l
RT: 3.48 min Scan# 289
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01

Tot Ion	Ratio	Lower	Upper
96	100		
61	141.5	0.0	290.6
98	64.4	0.0	137.8




Abundance Ion 96.00 (95.70 to 96.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 98.00 (97.70 to 98.70): VF0





#28

Bromochloromethane
Concen: 26.810 ug/l
RT: 3.72 min Scan# 313
Delta R.T. -0.04 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.4
130	72.4	64.2	96.4



Abundance Ion 49.00 (48.70 to 49.70): VF0
Ion 129.00 (128.70 to 129.70): V
Ion 130.00 (129.70 to 130.70): V





#29

Tetrahydrofuran

Concen: 133.451 ug/l

RT: 4.08 min Scan# 350

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion: 42 Resp: 35668

Ion Ratio Lower Upper




42 100

72 37.4 31.9 47.9

71 35.4 28.7 43.1



Abundance Ion 42.00 (41.70 to 42.70): VF0
Ion 72.00 (71.70 to 72.70): VF0
Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 25.707 ug/l

RT: 3.85 min Scan# 327

Delta R.T. -0.04 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 83 Resp: 106116



Ion Ratio Lower Upper


83 100

85 70.2 53.4 80.2



Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 14.241 ug/l

RT: 3.70 min Scan# 311

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion: 56 Resp: 39639

Ion Ratio Lower Upper

56 100

69 32.7 23.0 34.4




84 86.2 70.4 105.6



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 22.801 ug/l

RT: 4.09 min Scan# 351

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 97 Resp: 65275

Ion Ratio Lower Upper

97 100

99 57.6 53.5 80.3



61 48.1 38.6 58.0




Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

Ion 61.00 (60.70 to 61.70): VF0





#33

1,2-Dichloroethane-d4

Concen: 52.290 ug/l

RT: 4.85 min Scan# 428

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion: 65 Resp: 103370

Ion Ratio Lower Upper

65 100

67 46.2 0.0 96.4



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0

4.85

40000



30000

20000

10000

0

Time--> 4.70 4.80 4.90



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.60 min Scan# 504

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 114 Resp: 286860

Ion Ratio Lower Upper

114 100

63 21.2 0.0 41.6

88 20.1 0.0 36.6



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0


150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80





#35

Dibromofluoromethane

Concen: 50.024 ug/l

RT: 4.11 min Scan# 353

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:113 Resp: 113842

Ion Ratio Lower Upper

113 100

111 98.2 83.0 124.6




192 12.5 12.3 18.5



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 24.392 ug/l

RT: 4.28 min Scan# 370

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 75 Resp: 77608

Ion Ratio Lower Upper

75 100

110 30.2 15.6 46.8



77 31.1 23.4 35.0




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VF0









#37
Ethyl Acetate
Concen: 26.587 ug/l
RT: 4.12 min Scan# 354
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion: 43 Resp: 34376
Ion Ratio Lower Upper
43 100
61 56.6 68.6 103.0#
70 11.7 9.0 13.4



#38
Carbon Tetrachloride
Concen: 20.865 ug/l
RT: 3.99 min Scan# 341
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion: 117 Resp: 58636
Ion Ratio Lower Upper
117 100
119 104.4 74.4 111.6
121 31.8 23.5 35.3





#39

Methvlcvclohexane

Concen: 19.223 ug/l

RT: 5.45 min Scan# 489

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01





Tot Ion: 83 Resp: 64099

Ion Ratio Lower Upper

83 100

55 89.5 72.1 108.1

98 41.0 39.1 58.7



#40

Benzene

Concen: 24.265 ug/l

RT: 4.63 min Scan# 406

Delta R.T. -0.04 min

Lab File: VF061064.D




Acq: 18 Dec 2018 15:55


Tgt Ion: 78 Resp: 170562

Ion Ratio Lower Upper

78 100

77 24.9 18.7 28.1









#41
Methacrylonitrile
Concen: 24.170 ug/l
RT: 4.75 min Scan# 418
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01

Tot Ion:	41	Resp:	17522
Ion	Ratio	Lower	Upper
41	100		
39	56.2	53.0	79.6
67	40.7	43.8	65.6#
52	45.5	37.8	56.8




Abundance Ion 41.00 (40.70 to 41.70): VF0
Ion 39.00 (38.70 to 39.70): VF0
Ion 67.00 (66.70 to 67.70): VF0
Ion 52.00 (51.70 to 52.70): VF0






#42
1,2-Dichloroethane
Concen: 25.169 ug/l
RT: 4.95 min Scan# 438
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion:	62	Resp:	64588
Ion	Ratio	Lower	Upper
62	100		
98	6.4	0.0	17.2



Abundance Ion 62.00 (61.70 to 62.70): VF0
Ion 98.00 (97.70 to 98.70): VF0






#43




Isopropyl Acetate
 Concen: 23.757 ug/l
 RT: 6.96 min Scan# 642
 Delta R.T. -0.03 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01

Tot Ion: 43 Resp: 40177
 Ion Ratio Lower Upper
 43 100
 61 26.9 21.4 32.0
 87 0.9 0.2 0.4#




Abundance Ion 43.00 (42.70 to 43.70): VF0
 Ion 61.00 (60.70 to 61.70): VF0
 Ion 87.00 (86.70 to 87.70): VF0





#44


Trichloroethene
 Concen: 24.084 ug/l
 RT: 5.50 min Scan# 494
 Delta R.T. -0.04 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion:130 Resp: 55353
 Ion Ratio Lower Upper
 130 100
 95 103.8 0.0 195.6



Abundance Ion 129.90 (129.60 to 130.60): V
 Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 24.106 ug/l

RT: 6.24 min Scan# 569

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion: 63 Resp: 44008

Ion Ratio Lower Upper

63 100

65 25.8 23.2 34.8



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 65.00 (64.70 to 65.70): VF0

20000

15000

10000

5000



0

6.24

6.20

6.30

Time-->



#46

Dibromomethane

Concen: 22.691 ug/l

RT: 6.08 min Scan# 553

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 93 Resp: 28101

Ion Ratio Lower Upper

93 100

95 80.5 64.7 97.1

174 82.4 73.0 109.6



Abundance Ion 93.00 (92.70 to 93.70): VF0

Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

15000

10000

5000

0


6.08


6.00

6.10

6.20

Time-->





#47

Bromodichloromethane

Concen: 24.740 ug/l

RT: 6.39 min Scan# 584

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion: 83 Resp: 81376

Ion Ratio Lower Upper

83 100

85 63.7 51.1 76.7




127 7.1 6.8 10.2



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 22.148 ug/l

RT: 6.72 min Scan# 618

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 41 Resp: 25921

Ion Ratio Lower Upper

41 100

69 73.7 53.9 80.9



39 86.9 49.0 73.6#




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#49

1,4-Dioxane

Concen: 411.179 ug/l

RT: 6.71 min Scan# 617

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion: 88 Resp: 3522

Ion Ratio Lower Upper

88 100

43 78.7 42.0 63.0#




58 74.3 61.8 92.8



Abundance Ion 88.00 (87.70 to 88.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#50

Toluene-d8

Concen: 44.132 ug/l

RT: 7.55 min Scan# 702

Delta R.T. -0.04 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 98 Resp: 295845

Ion Ratio Lower Upper



98 100


100 68.2 59.3 88.9



Abundance Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#51

4-Methyl-2-Pentanone

Concen: 120.912 ug/l

RT: 8.26 min Scan# 774

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion: 43 Resp: 144168

Ion Ratio Lower Upper




43 100

58 35.9 29.7 44.5



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 21.846 ug/l

RT: 7.62 min Scan# 709

Delta R.T. -0.04 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 92 Resp: 107863

Ion Ratio Lower Upper



92 100


91 153.6 126.2 189.2



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0





#53

t-1,3-Dichloropropene

Concen: 23.658 ug/l

RT: 8.28 min Scan# 776

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion: 75 Resp: 63295

Ion Ratio Lower Upper

75 100

77 30.8 27.4 41.2



#54

cis-1,3-Dichloropropene

Concen: 24.188 ug/l

RT: 7.30 min Scan# 676

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55




Tgt Ion: 75 Resp: 80113


Ion Ratio Lower Upper

75 100

77 30.4 25.0 37.6

39 61.2 51.1 76.7





#55

1,1,2-Trichloroethane

Concen: 23.443 ug/l

RT: 8.49 min Scan# 797

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01

Tot Ion: 97 Resp: 33727



Ion Ratio Lower Upper

97 100

83 85.8 67.8 101.6

85 54.8 45.0 67.6

99 64.2 50.5 75.7





Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 22.097 ug/l

RT: 8.62 min Scan# 810

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55



Tgt Ion: 69 Resp: 35316

Ion Ratio Lower Upper

69 100

41 76.4 62.6 94.0


39 41.8 37.8 56.6




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 24.302 ug/l

RT: 8.85 min Scan# 834

Delta R.T. -0.04 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion: 76 Resp: 61170

Ion Ratio Lower Upper

76 100

78 35.3 27.0 40.4



#58

2-Chloroethyl Vinyl ether

Concen: 242.004 ug/l

RT: 7.30 min Scan# 676

Delta R.T. -0.04 min

Lab File: VF061064.D




Acq: 18 Dec 2018 15:55


Tgt Ion: 63 Resp: 39561

Ion Ratio Lower Upper

63 100

106 27.8 23.0 34.6









#59
2-Hexanone
Concen: 117.522 ug/l
RT: 9.50 min Scan# 899
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion: 43 Resp: 98307
Ion Ratio Lower Upper
43 100
58 52.3 25.6 76.6



#60
Dibromochloromethane
Concen: 22.066 ug/l
RT: 8.72 min Scan# 820
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion: 129 Resp: 47301
Ion Ratio Lower Upper
129 100
127 71.5 37.3 111.9





#61

1,2-Dibromoethane

Concen: 23.702 ug/l

RT: 8.99 min Scan# 848

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :


VF1218SBS01

Tot Ion:107 Resp: 36838

Ion Ratio Lower Upper




107 100

109 87.8 73.6 110.4



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 48.517 ug/l

RT: 11.41 min Scan# 1093

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion: 95 Resp: 147817

Ion Ratio Lower Upper

95 100

174 66.3 0.0 138.2



176 62.8 0.0 146.2




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.78 min Scan# 928

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:117 Resp: 266065

Ion Ratio Lower Upper

117 100

82 57.8 46.6 69.8




119 30.5 26.4 39.6



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 82.00 (81.70 to 82.70): VF0

Ion 119.00 (118.70 to 119.70): V



#64

Tetrachloroethene

Concen: 23.948 ug/l

RT: 8.15 min Scan# 763

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Tgt Ion:164 Resp: 48667


Ion Ratio Lower Upper

164 100

166 127.3 100.3 150.5

129 102.0 77.9 116.9

131 104.5 77.5 116.3






Abundance Ion 163.85 (163.55 to 164.55): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.90 (128.60 to 129.60): V

Ion 130.90 (130.60 to 131.60): V









#65
Chlorobenzene
Concen: 24.810 ug/l
RT: 9.80 min Scan# 930
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion:112 Resp: 128415
Ion Ratio Lower Upper
112 100
114 34.3 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 23.518 ug/l
RT: 9.93 min Scan# 943
Delta R.T. -0.04 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion:131 Resp: 52332
Ion Ratio Lower Upper
131 100
133 100.2 46.5 139.3
119 68.4 31.1 93.3





#67

Ethyl Benzene

Concen: 25.118 ug/l

RT: 9.90 min Scan# 940

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion: 91 Resp: 242041

Ion Ratio Lower Upper

91 100

106 28.8 22.2 33.2



#68

m/p-Xylenes

Concen: 47.597 ug/l

RT: 10.14 min Scan# 964

Delta R.T. -0.03 min

Lab File: VF061064.D




Acq: 18 Dec 2018 15:55


Tgt Ion: 106 Resp: 164958

Ion Ratio Lower Upper

106 100

91 209.4 166.7 250.1









#69
o-Xylene
Concen: 22.813 ug/l
RT: 10.71 min Scan# 1022
Delta R.T. -0.04 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion:106 Resp: 87186
Ion Ratio Lower Upper
106 100
91 232.3 104.8 314.6



#70
Styrene
Concen: 24.124 ug/l
RT: 10.79 min Scan# 1030
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion:104 Resp: 127799
Ion Ratio Lower Upper
104 100
78 48.7 37.3 55.9
103 47.6 36.8 55.2





#71

Bromoform

Concen: 22.584 ug/l

RT: 10.78 min Scan# 1029

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:173 Resp: 24274

Ion Ratio Lower Upper




173 100

175 48.6 25.5 76.5

254 7.5 7.5 11.3



Abundance Ion 173.00 (172.70 to 173.70): V
Ion 175.00 (174.70 to 175.70): V
Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.55 min Scan# 1209

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion:152 Resp: 138556

Ion Ratio Lower Upper



152 100


115 57.7 30.1 90.5

150 155.8 0.0 327.6



Abundance Ion 152.00 (151.70 to 152.70): V
Ion 115.00 (114.70 to 115.70): V
Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 24.930 ug/l

RT: 11.12 min Scan# 1064

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :





VF1218SBS01

Tot Ion:105 Resp: 272975

Ion Ratio Lower Upper

105 100

120 25.3 12.2 36.4



#74

N-ethyl acetate

Concen: 24.746 ug/l

RT: 11.37 min Scan# 1089

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Tgt Ion: 43 Resp: 62567




Ion Ratio Lower Upper


43 100

70 38.1 29.6 44.4

55 20.7 17.2 25.8

61 29.2 21.8 32.6





#75

1,1,2,2-Tetrachloroethane

Concen: 26.161 ug/l

RT: 11.70 min Scan# 1122

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion: 83 Resp: 47144

Ion Ratio Lower Upper

83 100

131 6.9 4.7 14.1




85 59.9 32.6 97.8



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): V

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 27.808 ug/l

RT: 11.79 min Scan# 1132

Delta R.T. -0.03 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 75 Resp: 36424

Ion Ratio Lower Upper



75 100


77 32.4 18.4 55.4



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 23.939 ug/l

RT: 11.50 min Scan# 1102

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:156 Resp: 57467

Ion Ratio Lower Upper

156 100

77 151.6 69.7 209.0




158 91.7 47.8 143.4



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 25.185 ug/l

RT: 11.60 min Scan# 1112

Delta R.T. -0.02 min

Lab File: VF061064.D


Acq: 18 Dec 2018 15:55

Tgt Ion: 91 Resp: 322940

Ion Ratio Lower Upper



91 100

120 21.5 10.3 30.8



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 25.370 ug/l
 RT: 11.72 min Scan# 1124
 Delta R.T. -0.03 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55




Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01


Tot Ion: 91 Resp: 186529
 Ion Ratio Lower Upper
 91 100
 126 30.8 15.9 47.6



#80
 1,3,5-Trimethylbenzene
 Concen: 25.578 ug/l
 RT: 11.82 min Scan# 1135
 Delta R.T. -0.03 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion:105 Resp: 227357
 Ion Ratio Lower Upper
 105 100
 120 48.2 25.8 77.3









#81
trans-1,4-Dichloro-2-butene
Concen: 30.592 ug/l
RT: 11.87 min Scan# 1140
Delta R.T. -0.02 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion: 75 Resp: 19633
Ion Ratio Lower Upper
75 100
53 77.1 65.9 98.9
89 54.1 38.8 58.2



#82
4-Chlorotoluene
Concen: 24.834 ug/l
RT: 11.90 min Scan# 1143
Delta R.T. -0.02 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion: 91 Resp: 192150
Ion Ratio Lower Upper
91 100
126 31.8 14.8 44.4









#83
tert-Butylbenzene
Concen: 25.591 ug/l
RT: 12.13 min Scan# 1166
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01

Tot Ion:119 Resp: 231713
Ion Ratio Lower Upper
119 100
91 60.3 30.1 90.3
134 23.4 12.6 37.8




Abundance Ion 119.00 (118.70 to 119.70): V
Ion 91.00 (90.70 to 91.70): VF0
Ion 134.00 (133.70 to 134.70): V






#84
1,2,4-Trimethylbenzene
Concen: 25.623 ug/l
RT: 12.21 min Scan# 1174
Delta R.T. -0.02 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion:105 Resp: 227313
Ion Ratio Lower Upper
105 100
120 45.7 23.8 71.5



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V









#85
 sec-Butylbenzene
 Concen: 24.890 ug/l
 RT: 12.31 min Scan# 1184
 Delta R.T. -0.02 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55




Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01


Tot Ion:105 Resp: 308512
 Ion Ratio Lower Upper
 105 100
 134 18.6 9.5 28.5



#86
 p-Isopropyltoluene
 Concen: 24.461 ug/l
 RT: 12.46 min Scan# 1200
 Delta R.T. -0.02 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion:119 Resp: 254201
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.6
 91 22.9 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 24.434 ug/l

RT: 12.47 min Scan# 1201

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:146 Resp: 111572

Ion Ratio Lower Upper

146 100

111 40.6 20.0 59.9

148 66.7 32.9 98.6






Abundance

Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#88

1,4-Dichlorobenzene

Concen: 23.327 ug/l

RT: 12.56 min Scan# 1210

Delta R.T. -0.03 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion:146 Resp: 103802

Ion Ratio Lower Upper

146 100

111 44.1 18.9 56.7

148 65.8 32.3 96.9






Abundance

Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V









#89
n-Butylbenzene
Concen: 25.977 ug/l
RT: 12.84 min Scan# 1238
Delta R.T. -0.03 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55




Instrument :
MSVOA_F
ClientSampleId :
VF1218SBS01


Tot Ion: 91 Resp: 272964
Ion Ratio Lower Upper
91 100
92 61.3 31.6 94.7
134 21.9 13.8 41.4



#90
Hexachloroethane
Concen: 24.675 ug/l
RT: 12.92 min Scan# 1246
Delta R.T. -0.02 min
Lab File: VF061064.D
Acq: 18 Dec 2018 15:55

Tgt Ion: 117 Resp: 61746
Ion Ratio Lower Upper
117 100
201 59.8 26.7 80.1









#91
 1,2-Dichlorobenzene
 Concen: 24.388 ug/l
 RT: 12.94 min Scan# 1248
 Delta R.T. -0.03 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55




Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01


Tot Ion:146 Resp: 105456
 Ion Ratio Lower Upper
 146 100
 111 44.8 21.2 63.6
 148 61.8 32.5 97.5



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 25.422 ug/l
 RT: 13.64 min Scan# 1319
 Delta R.T. -0.03 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion: 75 Resp: 8857
 Ion Ratio Lower Upper
 75 100
 155 63.7 45.6 136.9
 157 77.2 63.8 191.5









#93
 1,2,4-Trichlorobenzene
 Concen: 24.949 ug/l
 RT: 14.20 min Scan# 1376
 Delta R.T. -0.02 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55




Instrument :
 MSVOA_F
 ClientSampleId :
 VF1218SBS01

Tot Ion:180 Resp: 75379
 Ion Ratio Lower Upper
 180 100
 182 94.3 48.5 145.5
 145 37.5 18.1 54.3



#94
 Hexachlorobutadiene
 Concen: 24.089 ug/l
 RT: 14.19 min Scan# 1375
 Delta R.T. -0.02 min
 Lab File: VF061064.D
 Acq: 18 Dec 2018 15:55

Tgt Ion:225 Resp: 48518
 Ion Ratio Lower Upper
 225 100
 223 66.8 31.4 94.3
 227 64.8 32.5 97.5





#95

Naphthalene

Concen: 20.573 ug/l

RT: 14.46 min Scan# 1402

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55

Instrument :

MSVOA_F

ClientSampleId :

VF1218SBS01


Tot Ion:128 Resp: 132721

Ion Ratio Lower Upper

128 100

127 12.5 9.9 14.9




129 11.4 9.1 13.7



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 23.783 ug/l

RT: 14.61 min Scan# 1417

Delta R.T. -0.02 min

Lab File: VF061064.D

Acq: 18 Dec 2018 15:55


Tgt Ion:180 Resp: 66720

Ion Ratio Lower Upper

180 100

182 99.2 47.9 143.8

145 34.5 17.2 51.5



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

