


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF121818\
Data File : VF061055.D
Acq On : 18 Dec 2018 11:10
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Title : SW846 8260
Last Update : Tue Nov 27 01:36:26 2018



AutoFind: Scans 1092, 1093, 1094; Background Corrected with Scan 1086

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	22826	PASS
75	95	30	60	45.9	54490	PASS
95	95	100	100	100.0	118690	PASS
96	95	5	9	6.3	7425	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	67.5	80117	PASS
175	174	5	9	7.5	6000	PASS
176	174	95	101	95.5	76493	PASS
177	176	5	9	6.6	5053	PASS