

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_F\DATA\VF121918\  
 Data File : VF061074.D  
 Acq On : 18 Dec 2018 21:18  
 Operator : VA/AP  
 Sample : VSTDIC150  
 Misc : 5.00G/5mL/MSVOA-F/SOIL  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

Quant Time: Dec 19 04:21:08 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_F\METHODS\82F121918S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 19 04:15:13 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 4.87  | 168  | 203163   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.60  | 114  | 352816   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 9.77  | 117  | 317601   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.55 | 152  | 156111   | 50.00 | ug/l  | 0.00     |

## System Monitoring Compounds

|                           |       |     |          |        |         |      |
|---------------------------|-------|-----|----------|--------|---------|------|
| 33) 1,2-Dichloroethane-d4 | 4.84  | 65  | 381608   | 159.76 | ug/l    | 0.00 |
| Spiked Amount             |       |     |          |        |         |      |
|                           |       |     | Recovery | =      | 319.52% |      |
| 35) Dibromofluoromethane  | 4.11  | 113 | 389689   | 141.84 | ug/l    | 0.00 |
| Spiked Amount             |       |     |          |        |         |      |
|                           |       |     | Recovery | =      | 283.68% |      |
| 50) Toluene-d8            | 7.55  | 98  | 1058348  | 130.48 | ug/l    | 0.00 |
| Spiked Amount             |       |     |          |        |         |      |
|                           |       |     | Recovery | =      | 260.96% |      |
| 62) 4-Bromofluorobenzene  | 11.41 | 95  | 493004   | 134.19 | ug/l    | 0.00 |
| Spiked Amount             |       |     |          |        |         |      |
|                           |       |     | Recovery | =      | 268.38% |      |

## Target Compounds

|                               |      |     |         |         |        | Qvalue |
|-------------------------------|------|-----|---------|---------|--------|--------|
| 2) Dichlorodifluoromethane    | 0.95 | 85  | 606779  | 183.986 | ug/l   | 97     |
| 3) Chloromethane              | 1.10 | 50  | 371491  | 175.759 | ug/l   | 100    |
| 4) Vinyl Chloride             | 1.14 | 62  | 338808  | 173.194 | ug/l   | 89     |
| 5) Bromomethane               | 1.33 | 94  | 223198  | 164.100 | ug/l   | 99     |
| 6) Chloroethane               | 1.39 | 64  | 146232  | 173.420 | ug/l   | 89     |
| 7) Trichlorofluoromethane     | 1.49 | 101 | 641366  | 158.136 | ug/l   | 100    |
| 8) Diethyl Ether              | 1.63 | 74  | 86208   | 148.498 | ug/l   | 99     |
| 9) 1,1,2-Trichlorotrifluoroet | 1.83 | 101 | 306484  | 149.661 | ug/l   | 95     |
| 10) Methyl Iodide             | 1.90 | 142 | 451486  | 143.907 | ug/l   | 94     |
| 11) Tert butyl alcohol        | 2.58 | 59  | 78074   | 788.308 | ug/l   | 99     |
| 12) 1,1-Dichloroethene        | 1.79 | 96  | 247118  | 150.159 | ug/l   | 95     |
| 13) Acrolein                  | 2.00 | 56  | 49695   | 577.912 | ug/l   | 86     |
| 14) Allyl chloride            | 2.09 | 41  | 247085  | 128.742 | ug/l   | 95     |
| 15) Acrylonitrile             | 2.95 | 53  | 179248  | 790.765 | ug/l   | 95     |
| 16) Acetone                   | 2.24 | 43  | 325589  | 710.661 | ug/l   | 96     |
| 17) Carbon Disulfide          | 1.83 | 76  | 755354  | 167.373 | ug/l   | 98     |
| 18) Methyl Acetate            | 2.34 | 43  | 137331  | 161.872 | ug/l   | 95     |
| 19) Methyl tert-butyl Ether   | 2.43 | 73  | 555531  | 151.171 | ug/l   | 99     |
| 20) Methylene Chloride        | 2.25 | 84  | 166258  | 105.945 | ug/l   | 96     |
| 21) trans-1,2-Dichloroethene  | 2.30 | 96  | 171343  | 107.967 | ug/l   | 91     |
| 22) Diisopropyl ether         | 2.80 | 45  | 887537  | 159.661 | ug/l   | 95     |
| 23) Vinyl Acetate             | 3.18 | 43  | 1892615 | 741.821 | ug/l   | 100    |
| 24) 1,1-Dichloroethane        | 2.87 | 63  | 489414  | 148.195 | ug/l   | 98     |
| 25) 2-Butanone                | 4.34 | 43  | 535516  | 690.990 | ug/l   | 95     |
| 26) 2,2-Dichloropropane       | 3.60 | 77  | 282056  | 135.029 | ug/l   | 99     |
| 27) cis-1,2-Dichloroethene    | 3.47 | 96  | 374625  | 145.641 | ug/l   | 98     |
| 28) Bromochloromethane        | 3.72 | 49  | 244957  | 151.096 | ug/l   | 98     |
| 29) Tetrahydrofuran           | 4.07 | 42  | 226707  | 693.632 | ug/l   | 98     |
| 30) Chloroform                | 3.85 | 83  | 725353  | 142.253 | ug/l   | 100    |
| 31) Cyclohexane               | 3.69 | 56  | 291883  | 84.267  | ug/l   | 96     |
| 32) 1,1,1-Trichloroethane     | 4.09 | 97  | 456419  | 135.226 | ug/l   | 96     |
| 36) 1,1-Dichloropropene       | 4.28 | 75  | 557724  | 141.168 | ug/l   | 98     |
| 37) Ethyl Acetate             | 4.11 | 43  | 223470  | 140.024 | ug/l # | 93     |
| 38) Carbon Tetrachloride      | 3.98 | 117 | 444145  | 132.771 | ug/l   | 95     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_F\DATA\VF121918\  
 Data File : VF061074.D  
 Acq On : 18 Dec 2018 21:18  
 Operator : VA/AP  
 Sample : VSTDIC150  
 Misc : 5.00G/5mL/MSVOA-F/SOIL  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

Quant Time: Dec 19 04:21:08 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_F\METHODS\82F121918S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 19 04:15:13 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc     | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|----------|--------|----------|
| 39) Methylcyclohexane          | 5.45  | 83   | 450386   | 109.665  | ug/l   | 97       |
| 40) Benzene                    | 4.63  | 78   | 1217666  | 140.110  | ug/l   | 96       |
| 41) Methacrylonitrile          | 4.75  | 41   | 122410   | 129.565  | ug/l   | 96       |
| 42) 1,2-Dichloroethane         | 4.94  | 62   | 474533   | 149.764  | ug/l   | 96       |
| 43) Isopropyl Acetate          | 6.96  | 43   | 311971   | 145.735  | ug/l # | 97       |
| 44) Trichloroethene            | 5.50  | 130  | 365309   | 128.878  | ug/l   | 99       |
| 45) 1,2-Dichloropropane        | 6.23  | 63   | 315522   | 140.346  | ug/l   | 97       |
| 46) Dibromomethane             | 6.08  | 93   | 221623   | 145.541  | ug/l   | 97       |
| 47) Bromodichloromethane       | 6.38  | 83   | 577969   | 142.556  | ug/l   | 99       |
| 48) Methyl methacrylate        | 6.71  | 41   | 210327   | 147.004  | ug/l   | 96       |
| 49) 1,4-Dioxane                | 6.70  | 88   | 28549    | 2759.675 | ug/l # | 90       |
| 51) 4-Methyl-2-Pentanone       | 8.25  | 43   | 858999   | 590.867  | ug/l   | 100      |
| 52) Toluene                    | 7.62  | 92   | 792388   | 130.220  | ug/l   | 93       |
| 53) t-1,3-Dichloropropene      | 8.28  | 75   | 426886   | 130.413  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene    | 7.30  | 75   | 537283   | 133.009  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane      | 8.49  | 97   | 245320   | 140.100  | ug/l   | 95       |
| 56) Ethyl methacrylate         | 8.62  | 69   | 285461   | 147.837  | ug/l   | 99       |
| 57) 1,3-Dichloropropane        | 8.85  | 76   | 451509   | 146.908  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether  | 7.30  | 63   | 250402   | 1095.304 | ug/l   | 94       |
| 59) 2-Hexanone                 | 9.50  | 43   | 643946   | 626.802  | ug/l   | 99       |
| 60) Dibromochloromethane       | 8.72  | 129  | 368517   | 140.909  | ug/l   | 98       |
| 61) 1,2-Dibromoethane          | 8.98  | 107  | 277344   | 145.708  | ug/l   | 96       |
| 64) Tetrachloroethene          | 8.15  | 164  | 351987   | 144.583  | ug/l   | 97       |
| 65) Chlorobenzene              | 9.80  | 112  | 887536   | 144.653  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane  | 9.93  | 131  | 357701   | 136.100  | ug/l   | 97       |
| 67) Ethyl Benzene              | 9.90  | 91   | 1518380  | 132.571  | ug/l   | 94       |
| 68) m/p-Xylenes                | 10.14 | 106  | 1116210  | 270.031  | ug/l   | 85       |
| 69) o-Xylene                   | 10.71 | 106  | 613306   | 134.801  | ug/l   | 86       |
| 70) Styrene                    | 10.79 | 104  | 856552   | 136.780  | ug/l   | 98       |
| 71) Bromoform                  | 10.77 | 173  | 184872   | 146.343  | ug/l # | 98       |
| 73) Isopropylbenzene           | 11.12 | 105  | 1640337  | 131.221  | ug/l   | 96       |
| 74) N-amyl acetate             | 11.37 | 43   | 429084   | 150.195  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane  | 11.70 | 83   | 292770   | 142.810  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane     | 11.79 | 75   | 231538   | 154.223  | ug/l   | 97       |
| 77) Bromobenzene               | 11.50 | 156  | 375549   | 138.885  | ug/l   | 84       |
| 78) n-propylbenzene            | 11.60 | 91   | 1855201  | 126.658  | ug/l   | 91       |
| 79) 2-Chlorotoluene            | 11.73 | 91   | 1126500  | 133.964  | ug/l   | 95       |
| 80) 1,3,5-Trimethylbenzene     | 11.82 | 105  | 1342515  | 132.391  | ug/l   | 96       |
| 81) trans-1,4-Dichloro-2-buten | 11.87 | 75   | 148954   | 206.222  | ug/l   | 87       |
| 82) 4-Chlorotoluene            | 11.90 | 91   | 1229255  | 139.424  | ug/l   | 94       |
| 83) tert-Butylbenzene          | 12.13 | 119  | 1339325  | 129.118  | ug/l   | 94       |
| 84) 1,2,4-Trimethylbenzene     | 12.21 | 105  | 1377354  | 137.403  | ug/l   | 92       |
| 85) sec-Butylbenzene           | 12.31 | 105  | 1875750  | 133.405  | ug/l   | 92       |
| 86) p-Isopropyltoluene         | 12.46 | 119  | 1455311  | 123.056  | ug/l   | 96       |
| 87) 1,3-Dichlorobenzene        | 12.47 | 146  | 654402   | 127.249  | ug/l   | 93       |
| 88) 1,4-Dichlorobenzene        | 12.56 | 146  | 703605   | 140.310  | ug/l   | 94       |
| 89) n-Butylbenzene             | 12.84 | 91   | 1444499  | 121.362  | ug/l   | 96       |
| 90) Hexachloroethane           | 12.92 | 117  | 414879   | 146.252  | ug/l   | 74       |
| 91) 1,2-Dichlorobenzene        | 12.94 | 146  | 659125   | 134.741  | ug/l   | 97       |
| 92) 1,2-Dibromo-3-Chloropropan | 13.64 | 75   | 66445    | 167.625  | ug/l   | 94       |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_F\DATA\VF121918\  
Data File : VF061074.D  
Acq On : 18 Dec 2018 21:18  
Operator : VA/AP  
Sample : VSTDIC150  
Misc : 5.00G/5mL/MSVOA-F/SOIL  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDICC150

Quant Time: Dec 19 04:21:08 2018  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_F\METHODS\82F121918S.M  
Quant Title : SW846 8260  
QLast Update : Wed Dec 19 04:15:13 2018  
Response via : Initial Calibration


| Internal Standards         | R.T.  | QIon | Response | Conc    | Units | Dev(Min) |
|----------------------------|-------|------|----------|---------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 14.20 | 180  | 461532   | 134.917 | ug/l  | 94       |
| 94) Hexachlorobutadiene    | 14.19 | 225  | 320318   | 140.646 | ug/l  | 98       |
| 95) Naphthalene            | 14.46 | 128  | 1011917  | 134.755 | ug/l  | 98       |
| 96) 1,2,3-Trichlorobenzene | 14.60 | 180  | 463374   | 143.958 | ug/l  | 96       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_F\DATA\VF121918\  
Data File : VF061074.D  
Acq On : 18 Dec 2018 21:18  
Operator : VA/AP  
Sample : VSTDIC150  
Misc : 5.00G/5mL/MSVOA-F/SOIL  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150

Quant Time: Dec 19 04:21:08 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_F\METHODS\82F121918S.M  
Quant Title : SW846 8260  
QLast Update : Wed Dec 19 04:15:13 2018  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.87 min Scan# 430

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion:168 Resp: 203163

Ion Ratio Lower Upper

168 100

99 69.1 54.4 81.6



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.87



60000

40000

20000

0

Time--> 4.70 4.80 4.90 5.00



#2

Dichlorodifluoromethane

Concen: 183.986 ug/l

RT: 0.95 min Scan# 33

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 85 Resp: 606779

Ion Ratio Lower Upper

85 100

87 33.2 15.9 47.6



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.95


150000


100000

50000

0

Time--> 0.80 0.90 1.00 1.10 1.20





#3

Chloromethane

Concen: 175.759 ug/l

RT: 1.10 min Scan# 48

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC150

Tgt Ion: 50 Resp: 371491

Ion Ratio Lower Upper

50 100

52 33.0 26.3 39.5



#4

Vinyl Chloride

Concen: 173.194 ug/l

RT: 1.14 min Scan# 52

Delta R.T. 0.01 min

Lab File: VF061074.D




Acq: 18 Dec 2018 21:18

Tgt Ion: 62 Resp: 338808

Ion Ratio Lower Upper

62 100

64 31.5 30.4 45.6





#5

Bromomethane

Concen: 164.100 ug/l

RT: 1.33 min Scan# 71

Delta R.T. -0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 94 Resp: 223198

Ion Ratio Lower Upper

94 100

96 94.2 74.4 111.6



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0

1.33

40000

30000



20000

10000

0

Time-->

1.20 1.40 1.60



#6

Chloroethane

Concen: 173.420 ug/l

RT: 1.39 min Scan# 77

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 64 Resp: 146232

Ion Ratio Lower Upper

64 100

66 32.0 30.9 46.3



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0

1.39

50000

40000

30000


20000

10000

0

Time-->

1.20 1.30 1.40 1.50 1.60






#7




Trichlorofluoromethane  
Concen: 158.136 ug/l  
RT: 1.49 min Scan# 87  
Delta R.T. -0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150

Tgt Ion:101 Resp: 641366  
Ion Ratio Lower Upper  
101 100  
103 63.7 50.7 76.1




Abundance Ion 101.00 (100.70 to 101.70): V  
Ion 103.00 (102.70 to 103.70): V





#8

Diethyl Ether  
Concen: 148.498 ug/l  
RT: 1.63 min Scan# 102  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion: 74 Resp: 86208  
Ion Ratio Lower Upper  
74 100  
45 134.2 67.8 203.4



Abundance Ion 74.00 (73.70 to 74.70): VF0  
Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 149.661 ug/l

RT: 1.83 min Scan# 122

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion:101 Resp: 306484

Ion Ratio Lower Upper

101 100

85 55.3 44.5 66.7




151 70.0 49.9 74.9



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 143.907 ug/l

RT: 1.90 min Scan# 129

Delta R.T. -0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion:142 Resp: 451486

Ion Ratio Lower Upper

142 100

127 62.0 54.7 82.1



141 15.2 12.2 18.2




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11  
Tert butyl alcohol  
Concen: 788.308 ug/l  
RT: 2.58 min Scan# 198  
Delta R.T. 0.02 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion: 59 Resp: 78074  
Ion Ratio Lower Upper  
59 100  
57 14.8 11.5 17.3



#12  
1,1-Dichloroethene  
Concen: 150.159 ug/l  
RT: 1.79 min Scan# 118  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion: 96 Resp: 247118  
Ion Ratio Lower Upper  
96 100  
61 220.8 182.0 273.0  
98 63.4 46.5 69.7









#13  
Acrolein  
Concen: 577.912 ug/l  
RT: 2.00 min Scan# 139  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion: 56 Resp: 49695  
Ion Ratio Lower Upper  
56 100  
55 74.8 70.2 105.2



#14  
Allyl chloride  
Concen: 128.742 ug/l  
RT: 2.09 min Scan# 148  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion: 41 Resp: 247085  
Ion Ratio Lower Upper  
41 100  
39 113.4 96.4 144.6  
76 30.8 25.4 38.0





#15

Acrylonitrile

Concen: 790.765 ug/l

RT: 2.95 min Scan# 235

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 53 Resp: 179248

Ion Ratio Lower Upper

53 100

52 95.0 71.3 106.9




51 47.9 36.8 55.2



Abundance Ion 53.00 (52.70 to 53.70): VF0

Ion 52.00 (51.70 to 52.70): VF0

Ion 51.00 (50.70 to 51.70): VF0



#16

Acetone

Concen: 710.661 ug/l

RT: 2.24 min Scan# 163

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 43 Resp: 325589

Ion Ratio Lower Upper



43 100

58 15.6 11.1 16.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0









#17  
Carbon Disulfide  
Concen: 167.373 ug/l  
RT: 1.83 min Scan# 122  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion: 76 Resp: 755354  
Ion Ratio Lower Upper  
76 100  
78 11.9 10.2 15.4



#18  
Methyl Acetate  
Concen: 161.872 ug/l  
RT: 2.34 min Scan# 174  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion: 43 Resp: 137331  
Ion Ratio Lower Upper  
43 100  
74 15.8 11.0 16.6









#19

Methyl tert-butyl Ether  
 Concen: 151.171 ug/l  
 RT: 2.43 min Scan# 183  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150




Tgt Ion: 73 Resp: 55531  
 Ion Ratio Lower Upper  
 73 100  
 57 18.7 14.5 21.7




#20

Methylene Chloride  
 Concen: 105.945 ug/l  
 RT: 2.25 min Scan# 164  
 Delta R.T. -0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion: 84 Resp: 166258  
 Ion Ratio Lower Upper  
 84 100  
 49 159.3 129.4 194.2  
 51 47.3 44.0 66.0  
 86 62.2 46.3 69.5





#21

trans-1,2-Dichloroethene

Concen: 107.967 ug/l

RT: 2.30 min Scan# 170

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 96 Resp: 171343

Ion Ratio Lower Upper

96 100

61 168.7 146.0 219.0




98 59.4 51.6 77.4



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 159.661 ug/l

RT: 2.80 min Scan# 220

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Tgt Ion: 45 Resp: 887537


Ion Ratio Lower Upper

45 100

43 48.2 42.2 63.2

87 17.9 14.6 22.0

59 9.3 7.0 10.4






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 741.821 ug/l

RT: 3.18 min Scan# 259

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 43 Resp: 1892615

Ion Ratio Lower Upper

43 100

86 7.9 6.3 9.5



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0

800000

600000



400000

200000

0

3.10 3.20 3.30 3.40

Time-->



#24

1,1-Dichloroethane

Concen: 148.195 ug/l

RT: 2.87 min Scan# 227

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 63 Resp: 489414

Ion Ratio Lower Upper

63 100

98 5.0 2.1 6.5

100 2.7 1.7 5.1



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0

250000

200000

150000


100000


50000

0

2.80 2.90 3.00

Time-->





#25

2-Butanone

Concen: 690.990 ug/l

RT: 4.34 min Scan# 376

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 43 Resp: 53516

Ion Ratio Lower Upper




43 100

72 17.1 15.6 23.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 72.00 (71.70 to 72.70): VF0



#26

2,2-Dichloropropane

Concen: 135.029 ug/l

RT: 3.60 min Scan# 301

Delta R.T. 0.01 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 77 Resp: 282056

Ion Ratio Lower Upper



77 100


97 16.6 8.2 24.4



Abundance Ion 77.00 (76.70 to 77.70): VF0

Ion 97.00 (96.70 to 97.70): VF0






#27

cis-1,2-Dichloroethene  
 Concen: 145.641 ug/l  
 RT: 3.47 min Scan# 288  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 96      | 100   |       |       |
| 61      | 145.0 | 0.0   | 287.8 |
| 98      | 64.5  | 0.0   | 134.2 |






Abundance

Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0


Ion 98.00 (97.70 to 98.70): VF0



#28

Bromochloromethane  
 Concen: 151.096 ug/l  
 RT: 3.72 min Scan# 313  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 49      | 100   |       |       |
| 129     | 2.6   | 0.0   | 3.6   |
| 130     | 80.8  | 65.8  | 98.6  |






Abundance

Ion 49.00 (48.70 to 49.70): VF0

Ion 129.00 (128.70 to 129.70): V

Ion 130.00 (129.70 to 130.70): V





#29

Tetrahydrofuran

Concen: 693.632 ug/l

RT: 4.07 min Scan# 349

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 42 Resp: 226707

Ion Ratio Lower Upper

42 100

72 39.0 31.8 47.8




71 38.8 29.7 44.5



Abundance Ion 42.00 (41.70 to 42.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 142.253 ug/l

RT: 3.85 min Scan# 327

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 83 Resp: 725353

Ion Ratio Lower Upper



83 100


85 68.1 54.5 81.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 84.267 ug/l

RT: 3.69 min Scan# 310

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 56 Resp: 291883

Ion Ratio Lower Upper

56 100

69 30.5 26.0 39.0




84 85.8 71.4 107.2



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 135.226 ug/l

RT: 4.09 min Scan# 351

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 97 Resp: 456419

Ion Ratio Lower Upper

97 100

99 67.2 51.6 77.4



61 48.9 36.2 54.2




Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

Ion 61.00 (60.70 to 61.70): VF0





#33

1,2-Dichloroethane-d4

Concen: 159.760 ug/l

RT: 4.84 min Scan# 427

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 65 Resp: 381608

Ion Ratio Lower Upper




65 100

67 46.8 0.0 94.6



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.60 min Scan# 504

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 114 Resp: 352816

Ion Ratio Lower Upper

114 100

63 20.4 0.0 41.6



88 18.9 0.0 38.0



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0









#35  
 Dibromofluoromethane  
 Concen: 141.842 ug/l  
 RT: 4.11 min Scan# 353  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 113     | 100   |       |       |
| 111     | 99.8  | 85.2  | 127.8 |
| 192     | 16.4  | 12.8  | 19.2  |



#36  
 1,1-Dichloropropene  
 Concen: 141.168 ug/l  
 RT: 4.28 min Scan# 370  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 100   |       |       |
| 110     | 31.3  | 16.1  | 48.2  |
| 77      | 32.0  | 24.5  | 36.7  |





#37

Ethyl Acetate

Concen: 140.024 ug/l

RT: 4.11 min Scan# 353

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 43 Resp: 223470

Ion Ratio Lower Upper

43 100

61 76.6 56.6 84.8




70 9.6 6.3 9.5#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 70.00 (69.70 to 70.70): VF0



#38

Carbon Tetrachloride

Concen: 132.771 ug/l

RT: 3.98 min Scan# 340

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 117 Resp: 444145

Ion Ratio Lower Upper

117 100

119 96.8 82.1 123.1



121 30.2 25.5 38.3




Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V





#39

Methylcyclohexane

Concen: 109.665 ug/l

RT: 5.45 min Scan# 489

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 83 Resp: 450386

Ion Ratio Lower Upper

83 100

55 84.9 70.6 105.8




98 47.5 38.5 57.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#40

Benzene

Concen: 140.110 ug/l

RT: 4.63 min Scan# 406

Delta R.T. 0.01 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 78 Resp: 1217666

Ion Ratio Lower Upper



78 100


77 23.7 20.7 31.1



Abundance Ion 78.10 (77.80 to 78.80): VF0

Ion 77.00 (76.70 to 77.70): VF0









#41  
Methacrylonitrile  
Concen: 129.565 ug/l  
RT: 4.75 min Scan# 418  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 41    | Resp: | 122410 |
| Ion      | Ratio | Lower | Upper  |
| 41       | 100   |       |        |
| 39       | 60.0  | 45.7  | 68.5   |
| 67       | 50.3  | 38.9  | 58.3   |
| 52       | 40.1  | 35.2  | 52.8   |



#42  
1,2-Dichloroethane  
Concen: 149.764 ug/l  
RT: 4.94 min Scan# 437  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 62    | Resp: | 474533 |
| Ion      | Ratio | Lower | Upper  |
| 62       | 100   |       |        |
| 98       | 6.8   | 0.0   | 16.6   |





#43

Isopropyl Acetate

Concen: 145.735 ug/l

RT: 6.96 min Scan# 642

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 43 Resp: 311971

Ion Ratio Lower Upper




43 100

61 28.1 23.8 35.8

87 0.4 0.0 0.0#



Abundance Ion 43.00 (42.70 to 43.70): VF0  
Ion 61.00 (60.70 to 61.70): VF0  
Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 128.878 ug/l

RT: 5.50 min Scan# 494

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion:130 Resp: 365309



Ion Ratio Lower Upper

130 100

95 101.3 0.0 205.2



Abundance Ion 129.90 (129.60 to 130.60): V  
Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 140.346 ug/l

RT: 6.23 min Scan# 568

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 63 Resp: 315522

Ion Ratio Lower Upper

63 100

65 32.1 24.4 36.6



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 65.00 (64.70 to 65.70): VF0

6.23

120000

100000

80000



60000

40000

20000

0

Time-->



#46

Dibromomethane

Concen: 145.541 ug/l

RT: 6.08 min Scan# 553

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 93 Resp: 221623

Ion Ratio Lower Upper

93 100

95 83.5 66.9 100.3

174 87.6 74.1 111.1



Abundance Ion 93.00 (92.70 to 93.70): VF0

Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

6.08

100000

80000


60000


40000

20000

0

Time-->





#47

Bromodichloromethane

Concen: 142.556 ug/l

RT: 6.38 min Scan# 583

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 83 Resp: 577969

Ion Ratio Lower Upper




83 100

85 64.6 52.2 78.4

127 7.8 6.1 9.1



Abundance Ion 83.00 (82.70 to 83.70): VF0  
Ion 85.00 (84.70 to 85.70): VF0  
Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 147.004 ug/l

RT: 6.71 min Scan# 617

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 41 Resp: 210327

Ion Ratio Lower Upper



41 100

69 69.3 52.7 79.1

39 63.2 47.8 71.6



Abundance Ion 41.00 (40.70 to 41.70): VF0  
Ion 69.00 (68.70 to 69.70): VF0  
Ion 39.00 (38.70 to 39.70): VF0





#49

1,4-Dioxane

Concen: 2759.675 ug/l

RT: 6.70 min Scan# 616

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150





Tgt Ion: 88 Resp: 28549

Ion Ratio Lower Upper

88 100

43 50.2 33.3 49.9#

58 73.0 53.2 79.8



#50

Toluene-d8

Concen: 130.482 ug/l

RT: 7.55 min Scan# 702

Delta R.T. -0.00 min

Lab File: VF061074.D




Acq: 18 Dec 2018 21:18

Tgt Ion: 98 Resp: 1058348

Ion Ratio Lower Upper

98 100

100 74.6 55.3 82.9





#51

4-Methyl-2-Pentanone

Concen: 590.867 ug/l

RT: 8.25 min Scan# 773

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 43 Resp: 858999

Ion Ratio Lower Upper




43 100

58 36.1 29.1 43.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 130.220 ug/l

RT: 7.62 min Scan# 709

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 92 Resp: 792388

Ion Ratio Lower Upper



92 100


91 152.3 128.8 193.2



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0









#53  
 t-1,3-Dichloropropene  
 Concen: 130.413 ug/l  
 RT: 8.28 min Scan# 776  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

Tgt Ion: 75 Resp: 426886  
 Ion Ratio Lower Upper  
 75 100  
 77 33.7 25.8 38.8



#54  
 cis-1,3-Dichloropropene  
 Concen: 133.009 ug/l  
 RT: 7.30 min Scan# 676  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion: 75 Resp: 537283  
 Ion Ratio Lower Upper  
 75 100  
 77 31.9 25.0 37.4  
 39 64.4 50.1 75.1





#55

1,1,2-Trichloroethane

Concen: 140.100 ug/l

RT: 8.49 min Scan# 797

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150

Tgt Ion: 97 Resp: 245320


Ion Ratio Lower Upper

97 100

83 84.8 75.0 112.4

85 57.1 47.5 71.3

99 64.4 51.9 77.9






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 147.837 ug/l

RT: 8.62 min Scan# 810

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 69 Resp: 285461

Ion Ratio Lower Upper

69 100

41 74.3 59.7 89.5



39 44.8 35.0 52.6




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 146.908 ug/l

RT: 8.85 min Scan# 834

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 451509

Ion Ratio Lower Upper

76 100

78 32.8 26.5 39.7



#58

2-Chloroethyl Vinyl ether

Concen: 1095.304 ug/l

RT: 7.30 min Scan# 676

Delta R.T. 0.01 min

Lab File: VF061074.D




Acq: 18 Dec 2018 21:18


Tgt Ion: 63 Resp: 250402

Ion Ratio Lower Upper

63 100

106 27.7 19.8 29.8








#59  
2-Hexanone  
Concen: 626.802 ug/l  
RT: 9.50 min Scan# 899  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion: 43 Resp: 643946  
Ion Ratio Lower Upper  
43 100  
58 50.9 25.1 75.2



#60  
Dibromochloromethane  
Concen: 140.909 ug/l  
RT: 8.72 min Scan# 820  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion: 129 Resp: 368517  
Ion Ratio Lower Upper  
129 100  
127 75.2 36.8 110.3





#61

1,2-Dibromoethane

Concen: 145.708 ug/l

RT: 8.98 min Scan# 847

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion: 107 Resp: 277344

Ion Ratio Lower Upper




107 100

109 96.0 73.7 110.5



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 134.188 ug/l

RT: 11.41 min Scan# 1093

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 95 Resp: 493004

Ion Ratio Lower Upper

95 100

174 66.7 0.0 143.8



176 65.2 0.0 147.6




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 9.77 min Scan# 927  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150

Tgt Ion:117 Resp: 317601  
Ion Ratio Lower Upper  
117 100  
82 59.9 45.0 67.4  
119 31.9 24.8 37.2




Abundance Ion 117.00 (116.70 to 117.70): V  
Ion 82.00 (81.70 to 82.70): VF0  
Ion 119.00 (118.70 to 119.70): V





#64  
Tetrachloroethene  
Concen: 144.583 ug/l  
RT: 8.15 min Scan# 763  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion:164 Resp: 351987  
Ion Ratio Lower Upper  
164 100  
166 131.1 104.0 156.0  
129 97.9 81.1 121.7  
131 96.1 81.2 121.8



Abundance Ion 163.85 (163.55 to 164.55): V  
Ion 165.80 (165.50 to 166.50): V  
Ion 128.90 (128.60 to 129.60): V  
Ion 130.90 (130.60 to 131.60): V









#65  
Chlorobenzene  
Concen: 144.653 ug/l  
RT: 9.80 min Scan# 930  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion:112 Resp: 887536  
Ion Ratio Lower Upper  
112 100  
114 35.5 27.8 41.6



#66  
1,1,1,2-Tetrachloroethane  
Concen: 136.100 ug/l  
RT: 9.93 min Scan# 943  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion:131 Resp: 357701  
Ion Ratio Lower Upper  
131 100  
133 98.5 49.5 148.7  
119 64.7 34.5 103.5









#67  
 Ethyl Benzene  
 Concen: 132.571 ug/l  
 RT: 9.90 min Scan# 940  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150


Tgt Ion: 91 Resp: 1518380  
 Ion Ratio Lower Upper  
 91 100  
 106 31.1 22.5 33.7



#68  
 m/p-Xylenes  
 Concen: 270.031 ug/l  
 RT: 10.14 min Scan# 964  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion: 106 Resp: 1116210  
 Ion Ratio Lower Upper  
 106 100  
 91 186.3 168.2 252.2









#69  
o-Xylene  
Concen: 134.801 ug/l  
RT: 10.71 min Scan# 1022  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion:106 Resp: 613306  
Ion Ratio Lower Upper  
106 100  
91 208.9 115.9 347.7



#70  
Styrene  
Concen: 136.780 ug/l  
RT: 10.79 min Scan# 1030  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion:104 Resp: 856552  
Ion Ratio Lower Upper  
104 100  
78 49.2 38.4 57.6  
103 44.7 37.0 55.6





#71

Bromoform

Concen: 146.343 ug/l

RT: 10.77 min Scan# 1028

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion:173 Resp: 184872

Ion Ratio Lower Upper

173 100

175 48.1 24.8 74.4




254 7.1 0.0 0.0#



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.55 min Scan# 1209

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion:152 Resp: 156111

Ion Ratio Lower Upper

152 100

115 56.2 33.3 99.9



150 179.9 0.0 328.2




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 131.221 ug/l

RT: 11.12 min Scan# 1064

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC150

Tgt Ion:105 Resp: 1640337

Ion Ratio Lower Upper




105 100

120 28.0 13.0 39.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 150.195 ug/l

RT: 11.37 min Scan# 1089

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Tgt Ion: 43 Resp: 429084


Ion Ratio Lower Upper

43 100

70 36.2 29.2 43.8

55 20.1 16.5 24.7

61 26.7 20.7 31.1





Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0





#75

1,1,2,2-Tetrachloroethane

Concen: 142.810 ug/l

RT: 11.70 min Scan# 1122

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion: 83 Resp: 292770

Ion Ratio Lower Upper

83 100

131 9.7 4.8 14.3




85 66.7 33.9 101.6



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): V

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 154.223 ug/l

RT: 11.79 min Scan# 1132

Delta R.T. -0.00 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 75 Resp: 231538

Ion Ratio Lower Upper



75 100


77 33.7 17.6 52.9



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 138.885 ug/l

RT: 11.50 min Scan# 1102

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion:156 Resp: 375549

Ion Ratio Lower Upper

156 100

77 154.4 93.8 281.4




158 100.4 48.9 146.8



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 126.658 ug/l

RT: 11.60 min Scan# 1112

Delta R.T. 0.01 min

Lab File: VF061074.D


Acq: 18 Dec 2018 21:18

Tgt Ion: 91 Resp: 1855201

Ion Ratio Lower Upper



91 100


120 24.8 10.4 31.2



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V









#79  
 2-Chlorotoluene  
 Concen: 133.964 ug/l  
 RT: 11.73 min Scan# 1125  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

Tgt Ion: 91 Resp: 1126500  
 Ion Ratio Lower Upper  
 91 100  
 126 35.3 16.4 49.1



#80  
 1,3,5-Trimethylbenzene  
 Concen: 132.391 ug/l  
 RT: 11.82 min Scan# 1135  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion: 105 Resp: 1342515  
 Ion Ratio Lower Upper  
 105 100  
 120 53.4 25.4 76.4





#81

trans-1,4-Dichloro-2-butene

Concen: 206.222 ug/l

RT: 11.87 min Scan# 1140

Delta R.T. 0.01 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150





Tgt Ion: 75 Resp: 148954

Ion Ratio Lower Upper

75 100

53 75.3 74.1 111.1

89 51.3 42.6 64.0



#82

4-Chlorotoluene

Concen: 139.424 ug/l

RT: 11.90 min Scan# 1143

Delta R.T. 0.01 min

Lab File: VF061074.D




Acq: 18 Dec 2018 21:18


Tgt Ion: 91 Resp: 1229255

Ion Ratio Lower Upper

91 100

126 35.2 16.0 47.9









#83  
 tert-Butylbenzene  
 Concen: 129.118 ug/l  
 RT: 12.13 min Scan# 1166  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

Tgt Ion:119 Resp: 1339325  
 Ion Ratio Lower Upper  
 119 100  
 91 66.6 30.5 91.5  
 134 26.6 12.4 37.4




Abundance Ion 119.00 (118.70 to 119.70): V  
 Ion 91.00 (90.70 to 91.70): VF0  
 Ion 134.00 (133.70 to 134.70): V






#84  
 1,2,4-Trimethylbenzene  
 Concen: 137.403 ug/l  
 RT: 12.21 min Scan# 1174  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion:105 Resp: 1377354  
 Ion Ratio Lower Upper  
 105 100  
 120 52.8 23.6 71.0



Abundance Ion 105.00 (104.70 to 105.70): V  
 Ion 120.00 (119.70 to 120.70): V









#85  
 sec-Butylbenzene  
 Concen: 133.405 ug/l  
 RT: 12.31 min Scan# 1184  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150


Tgt Ion:105 Resp: 1875750  
 Ion Ratio Lower Upper  
 105 100  
 134 22.6 9.4 28.2



#86  
 p-Isopropyltoluene  
 Concen: 123.056 ug/l  
 RT: 12.46 min Scan# 1199  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion:119 Resp: 1455311  
 Ion Ratio Lower Upper  
 119 100  
 134 26.8 12.8 38.4  
 91 27.1 12.2 36.6









#87  
 1,3-Dichlorobenzene  
 Concen: 127.249 ug/l  
 RT: 12.47 min Scan# 1201  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150


Tgt Ion:146 Resp: 654402  
 Ion Ratio Lower Upper  
 146 100  
 111 42.6 19.3 57.9  
 148 70.5 32.4 97.2



#88  
 1,4-Dichlorobenzene  
 Concen: 140.310 ug/l  
 RT: 12.56 min Scan# 1210  
 Delta R.T. -0.00 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Tgt Ion:146 Resp: 703605  
 Ion Ratio Lower Upper  
 146 100  
 111 43.9 19.4 58.4  
 148 68.3 32.4 97.0









#89  
n-Butylbenzene  
Concen: 121.362 ug/l  
RT: 12.84 min Scan# 1238  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150


Tgt Ion: 91 Resp: 1444499  
Ion Ratio Lower Upper  
91 100  
92 69.4 32.8 98.4  
134 27.2 13.2 39.5



#90  
Hexachloroethane  
Concen: 146.252 ug/l  
RT: 12.92 min Scan# 1246  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion: 117 Resp: 414879  
Ion Ratio Lower Upper  
117 100  
201 67.8 24.9 74.8





#91

1,2-Dichlorobenzene

Concen: 134.741 ug/l

RT: 12.94 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC150


Tgt Ion:146 Resp: 659125

Ion Ratio Lower Upper

146 100

111 46.7 23.2 69.5




148 68.4 32.4 97.0



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 167.625 ug/l

RT: 13.64 min Scan# 1319

Delta R.T. -0.00 min

Lab File: VF061074.D

Acq: 18 Dec 2018 21:18


Tgt Ion: 75 Resp: 66445

Ion Ratio Lower Upper

75 100

155 75.9 41.5 124.5



157 101.5 53.2 159.6



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V









#93  
 1,2,4-Trichlorobenzene  
 Concen: 134.917 ug/l  
 RT: 14.20 min Scan# 1376  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC150

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 180     | 461532 |       |       |
| 182     | 99.9   | 46.9  | 140.6 |
| 145     | 37.3   | 20.5  | 61.6  |




Abundance Ion 180.00 (179.70 to 180.70): V  
 Ion 182.00 (181.70 to 182.70): V  
 Ion 145.00 (144.70 to 145.70): V






#94  
 Hexachlorobutadiene  
 Concen: 140.646 ug/l  
 RT: 14.19 min Scan# 1375  
 Delta R.T. 0.01 min  
 Lab File: VF061074.D  
 Acq: 18 Dec 2018 21:18

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 225     | 320318 |       |       |
| 223     | 62.7   | 32.8  | 98.3  |
| 227     | 67.4   | 34.0  | 101.8 |



Abundance Ion 225.00 (224.70 to 225.70): V  
 Ion 223.00 (222.70 to 223.70): V  
 Ion 227.00 (226.70 to 227.70): V









#95  
Naphthalene  
Concen: 134.755 ug/l  
RT: 14.46 min Scan# 1402  
Delta R.T. 0.01 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC150

Tgt Ion:128 Resp: 1011917  
Ion Ratio Lower Upper  
128 100  
127 12.3 9.4 14.2  
129 11.8 8.8 13.2



#96  
1,2,3-Trichlorobenzene  
Concen: 143.958 ug/l  
RT: 14.60 min Scan# 1416  
Delta R.T. -0.00 min  
Lab File: VF061074.D  
Acq: 18 Dec 2018 21:18

Tgt Ion:180 Resp: 463374  
Ion Ratio Lower Upper  
180 100  
182 100.0 47.9 143.8  
145 38.4 18.6 55.8

