

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF122118\
 Data File : VF061118.D
 Acq On : 21 Dec 2018 11:10
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 21 14:18:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.90	168	210443	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	5.62	114	356539	50.00	ug/l	0.03
63) Chlorobenzene-d5	9.81	117	327354	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	12.58	152	156227	50.00	ug/l	0.03

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	4.87	65	122675	45.32	ug/l	0.03
Spiked Amount	50.000		Recovery	=	90.64%	
35) Dibromofluoromethane	4.14	113	137569	48.17	ug/l	0.03
Spiked Amount	50.000		Recovery	=	96.34%	
50) Toluene-d8	7.59	98	416585	51.47	ug/l	0.03
Spiked Amount	50.000		Recovery	=	102.94%	
62) 4-Bromofluorobenzene	11.43	95	175529	47.04	ug/l	0.02
Spiked Amount	50.000		Recovery	=	94.08%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	0.98	85	193433	41.349	ug/l	99
3) Chloromethane	1.12	50	122197	45.602	ug/l	100
4) Vinyl Chloride	1.15	62	113213	45.544	ug/l	99
5) Bromomethane	1.37	94	80116	46.467	ug/l	99
6) Chloroethane	1.41	64	57101	49.853	ug/l #	86
7) Trichlorofluoromethane	1.50	101	139805	29.456	ug/l	99
8) Diethyl Ether	1.65	74	33148	51.001	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	1.86	101	112226	45.603	ug/l	94
10) Methyl Iodide	1.93	142	152455	39.301	ug/l	91
11) Tert butyl alcohol	2.60	59	29137	266.762	ug/l	96
12) 1,1-Dichloroethene	1.82	96	87280	47.050	ug/l	93
13) Acrolein	2.01	56	24963	316.287	ug/l	87
14) Allyl chloride	2.11	41	63996	33.150	ug/l	96
15) Acrylonitrile	2.97	53	66138	261.860	ug/l	93
16) Acetone	2.26	43	144345	266.767	ug/l #	90
17) Carbon Disulfide	1.85	76	189441	33.792	ug/l	99
18) Methyl Acetate	2.37	43	59442	53.098	ug/l	96
19) Methyl tert-butyl Ether	2.45	73	185637	45.084	ug/l	97
20) Methylene Chloride	2.29	84	65676	36.797	ug/l	92
21) trans-1,2-Dichloroethene	2.33	96	47910	24.147	ug/l	93
22) Diisopropyl ether	2.82	45	299582	46.588	ug/l	96
23) Vinyl Acetate	3.22	43	776164	261.641	ug/l	99
24) 1,1-Dichloroethane	2.89	63	162920	42.761	ug/l	98
25) 2-Butanone	4.37	43	246662	281.082	ug/l	99
26) 2,2-Dichloropropane	3.63	77	99287	45.409	ug/l	95
27) cis-1,2-Dichloroethene	3.50	96	148211	48.794	ug/l	93
28) Bromochloromethane	3.75	49	97532	50.451	ug/l	92
29) Tetrahydrofuran	4.11	42	91069	254.973	ug/l	96
30) Chloroform	3.89	83	260264	45.558	ug/l	94
31) Cyclohexane	3.72	56	96963	24.526	ug/l	98
32) 1,1,1-Trichloroethane	4.12	97	136595	39.891	ug/l	92
36) 1,1-Dichloropropene	4.31	75	205650	47.785	ug/l	99
37) Ethyl Acetate	4.15	43	82442	49.099	ug/l #	77
38) Carbon Tetrachloride	4.02	117	141945	43.310	ug/l	95

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF122118\
 Data File : VF061118.D
 Acq On : 21 Dec 2018 11:10
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 21 14:18:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.48	83	171790	44.494	ug/l	97
40) Benzene	4.67	78	457431	49.425	ug/l	98
41) Methacrylonitrile	4.78	41	43648	48.023	ug/l #	90
42) 1,2-Dichloroethane	4.97	62	161015	45.797	ug/l	98
43) Isopropyl Acetate	6.99	43	116905	56.258	ug/l #	100
44) Trichloroethene	5.53	130	143595	49.212	ug/l	99
45) 1,2-Dichloropropane	6.26	63	119072	47.288	ug/l	96
46) Dibromomethane	6.12	93	77611	47.523	ug/l	97
47) Bromodichloromethane	6.41	83	201208	47.418	ug/l	98
48) Methyl methacrylate	6.75	41	76222	52.542	ug/l	92
49) 1,4-Dioxane	6.75	88	11393	1082.353	ug/l #	90
51) 4-Methyl-2-Pentanone	8.29	43	356768	250.336	ug/l	98
52) Toluene	7.65	92	282863	46.171	ug/l	98
53) t-1,3-Dichloropropene	8.31	75	165245	48.273	ug/l	97
54) cis-1,3-Dichloropropene	7.33	75	206619	48.093	ug/l	99
55) 1,1,2-Trichloroethane	8.52	97	90871	49.478	ug/l	95
56) Ethyl methacrylate	8.65	69	103160	51.909	ug/l	97
57) 1,3-Dichloropropane	8.89	76	160793	48.407	ug/l	99
58) 2-Chloroethyl Vinyl ether	7.33	63	109449	286.754	ug/l	98
59) 2-Hexanone	9.52	43	303357	294.338	ug/l	97
60) Dibromochloromethane	8.75	129	127552	48.449	ug/l	89
61) 1,2-Dibromoethane	9.02	107	95435	48.655	ug/l	99
64) Tetrachloroethene	8.19	164	126088	50.071	ug/l	93
65) Chlorobenzene	9.83	112	324439	47.975	ug/l	97
66) 1,1,1,2-Tetrachloroethane	9.96	131	130164	47.099	ug/l	97
67) Ethyl Benzene	9.93	91	590782	47.952	ug/l	97
68) m/p-Xylenes	10.16	106	423103	97.583	ug/l	97
69) o-Xylene	10.74	106	217223	45.173	ug/l	91
70) Styrene	10.81	104	313455	53.872	ug/l	99
71) Bromoform	10.80	173	61847	48.371	ug/l	99
73) Isopropylbenzene	11.15	105	659097	48.957	ug/l	99
74) N-amyl acetate	11.39	43	163029	54.543	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.72	83	108180	48.416	ug/l	99
76) 1,2,3-Trichloropropane	11.82	75	79771	47.973	ug/l	98
77) Bromobenzene	11.52	156	137614	47.484	ug/l	90
78) n-propylbenzene	11.62	91	746268	47.686	ug/l	95
79) 2-Chlorotoluene	11.74	91	432420	46.821	ug/l	95
80) 1,3,5-Trimethylbenzene	11.85	105	534272	48.195	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.89	75	48642	68.075	ug/l	90
82) 4-Chlorotoluene	11.93	91	437045	45.745	ug/l	95
83) tert-Butylbenzene	12.15	119	521971	45.757	ug/l	99
84) 1,2,4-Trimethylbenzene	12.23	105	516228	46.697	ug/l	98
85) sec-Butylbenzene	12.33	105	707560	47.529	ug/l	98
86) p-Isopropyltoluene	12.48	119	607397	49.155	ug/l	99
87) 1,3-Dichlorobenzene	12.50	146	259230	55.390	ug/l	97
88) 1,4-Dichlorobenzene	12.59	146	249941	46.967	ug/l	98
89) n-Butylbenzene	12.86	91	604694	47.992	ug/l	98
90) Hexachloroethane	12.94	117	144274	47.258	ug/l	75
91) 1,2-Dichlorobenzene	12.96	146	238598	46.052	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	13.66	75	21578	52.627	ug/l	82


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF122118\
Data File : VF061118.D
Acq On : 21 Dec 2018 11:10
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Quant Time: Dec 21 14:18:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 00:56:59 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.22	180	176076	47.920	ug/l	95
94) Hexachlorobutadiene	14.21	225	117014	50.001	ug/l	94
95) Naphthalene	14.48	128	335342	49.980	ug/l	99
96) 1,2,3-Trichlorobenzene	14.63	180	158399	46.632	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.90 min Scan# 433

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:168 Resp: 210443

Ion Ratio Lower Upper

168 100

99 66.4 54.4 81.6



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.90



60000

40000

20000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 41.349 ug/l

RT: 0.98 min Scan# 35

Delta R.T. 0.02 min

Lab File: VF061118.D


Acq: 21 Dec 2018 11:10

Tgt Ion: 85 Resp: 193433

Ion Ratio Lower Upper

85 100

87 32.3 15.9 47.6



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.98

50000

40000


30000


20000

10000

0

Time-->









#3
Chloromethane
Concen: 45.602 ug/l
RT: 1.12 min Scan# 50
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 50 Resp: 122197
Ion Ratio Lower Upper
50 100
52 33.2 26.3 39.5



#4
Vinyl Chloride
Concen: 45.544 ug/l
RT: 1.15 min Scan# 53
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 62 Resp: 113213
Ion Ratio Lower Upper
62 100
64 37.4 30.4 45.6





#5

Bromomethane

Concen: 46.467 ug/l

RT: 1.37 min Scan# 75

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 94 Resp: 80116

Ion Ratio Lower Upper

94 100

96 93.6 74.4 111.6



#6

Chloroethane

Concen: 49.853 ug/l

RT: 1.41 min Scan# 79

Delta R.T. 0.02 min

Lab File: VF061118.D




Acq: 21 Dec 2018 11:10


Tgt Ion: 64 Resp: 57101

Ion Ratio Lower Upper

64 100

66 29.9 30.9 46.3#









#7

Trichlorofluoromethane
Concen: 29.456 ug/l
RT: 1.50 min Scan# 88
Delta R.T. 0.00 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050




Tot Ion:101 Resp: 139805
Ion Ratio Lower Upper
101 100
103 64.5 50.7 76.1




#8

Diethyl Ether
Concen: 51.001 ug/l
RT: 1.65 min Scan# 103
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 74 Resp: 33148
Ion Ratio Lower Upper
74 100
45 132.7 67.8 203.4





#9

1,1,2-Trichlorotrifluoroethane

Concen: 45.603 ug/l

RT: 1.86 min Scan# 125

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:101 Resp: 112226

Ion Ratio Lower Upper

101 100

85 48.6 44.5 66.7




151 64.0 49.9 74.9



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 39.301 ug/l

RT: 1.93 min Scan# 132

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Tgt Ion:142 Resp: 152455

Ion Ratio Lower Upper

142 100

127 60.0 54.7 82.1



141 13.7 12.2 18.2




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11
Tert butyl alcohol
Concen: 266.762 ug/l
RT: 2.60 min Scan# 200
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 59 Resp: 29137
Ion Ratio Lower Upper
59 100
57 12.7 11.5 17.3



#12
1,1-Dichloroethene
Concen: 47.050 ug/l
RT: 1.82 min Scan# 120
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 96 Resp: 87280
Ion Ratio Lower Upper
96 100
61 216.4 182.0 273.0
98 64.1 46.5 69.7









#13
Acrolein
Concen: 316.287 ug/l
RT: 2.01 min Scan# 140
Delta R.T. 0.01 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 56 Resp: 24963
Ion Ratio Lower Upper
56 100
55 75.4 70.2 105.2



#14
Allyl chloride
Concen: 33.150 ug/l
RT: 2.11 min Scan# 150
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 41 Resp: 63996
Ion Ratio Lower Upper
41 100
39 117.5 96.4 144.6
76 37.9 25.4 38.0





#15

Acrylonitrile

Concen: 261.860 ug/l

RT: 2.97 min Scan# 237

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 53 Resp: 66138

Ion Ratio Lower Upper

53 100

52 81.4 71.3 106.9

51 42.3 36.8 55.2



#16

Acetone

Concen: 266.767 ug/l

RT: 2.26 min Scan# 165

Delta R.T. 0.02 min

Lab File: VF061118.D




Acq: 21 Dec 2018 11:10


Tgt Ion: 43 Resp: 144345

Ion Ratio Lower Upper

43 100

58 18.1 11.1 16.7#





#17

Carbon Disulfide

Concen: 33.792 ug/l

RT: 1.85 min Scan# 124

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 76 Resp: 189441

Ion Ratio Lower Upper

76 100

78 12.4 10.2 15.4



#18

Methyl Acetate

Concen: 53.098 ug/l

RT: 2.37 min Scan# 176

Delta R.T. 0.02 min

Lab File: VF061118.D




Acq: 21 Dec 2018 11:10


Tgt Ion: 43 Resp: 59442

Ion Ratio Lower Upper

43 100

74 15.5 11.0 16.6









#19
Methyl tert-butyl Ether
Concen: 45.084 ug/l
RT: 2.45 min Scan# 184
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 73 Resp: 185637
Ion Ratio Lower Upper
73 100
57 16.8 14.5 21.7



#20
Methylene Chloride
Concen: 36.797 ug/l
RT: 2.29 min Scan# 168
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 84 Resp: 65676
Ion Ratio Lower Upper
84 100
49 151.9 129.4 194.2
51 47.7 44.0 66.0
86 63.9 46.3 69.5





#21

trans-1,2-Dichloroethene

Concen: 24.147 ug/l

RT: 2.33 min Scan# 172

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 96 Resp: 47910

Ion Ratio Lower Upper

96 100

61 169.6 146.0 219.0




98 64.0 51.6 77.4



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 46.588 ug/l

RT: 2.82 min Scan# 222

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Tgt Ion: 45 Resp: 299582


Ion Ratio Lower Upper

45 100

43 48.8 42.2 63.2

87 16.8 14.6 22.0

59 8.7 7.0 10.4






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 261.641 ug/l

RT: 3.22 min Scan# 262

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 43 Resp: 776164

Ion Ratio Lower Upper

43 100

86 8.1 6.3 9.5



#24

1,1-Dichloroethane

Concen: 42.761 ug/l

RT: 2.89 min Scan# 229

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10




Tgt Ion: 63 Resp: 162920


Ion Ratio Lower Upper

63 100

98 3.8 2.1 6.5

100 2.5 1.7 5.1





#25

2-Butanone

Concen: 281.082 ug/l

RT: 4.37 min Scan# 379

Delta R.T. 0.04 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 246662

Ion Ratio Lower Upper




43 100

72 19.9 15.6 23.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 72.00 (71.70 to 72.70): VF0



#26

2,2-Dichloropropane

Concen: 45.409 ug/l

RT: 3.63 min Scan# 304

Delta R.T. 0.04 min

Lab File: VF061118.D


Acq: 21 Dec 2018 11:10

Tgt Ion: 77 Resp: 99287

Ion Ratio Lower Upper



77 100


97 18.6 8.2 24.4



Abundance Ion 77.00 (76.70 to 77.70): VF0

Ion 97.00 (96.70 to 97.70): VF0









#27

cis-1,2-Dichloroethene
Concen: 48.794 ug/l
RT: 3.50 min Scan# 291
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050




Tot Ion	Ratio	Lower	Upper
96	100		
61	134.7	0.0	287.8
98	63.2	0.0	134.2




#28

Bromochloromethane
Concen: 50.451 ug/l
RT: 3.75 min Scan# 316
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	3.6
130	74.6	65.8	98.6






#29
Tetrahydrofuran
Concen: 254.973 ug/l
RT: 4.11 min Scan# 353
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion	42	Resp	91069
Ion Ratio	Lower	Upper	
42	100		
72	36.3	31.8	47.8
71	38.3	29.7	44.5






Abundance

Ion 42.00 (41.70 to 42.70): VF0


Ion 72.00 (71.70 to 72.70): VF0

Ion 71.00 (70.70 to 71.70): VF0



#30
Chloroform
Concen: 45.558 ug/l
RT: 3.89 min Scan# 330
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10



Tgt Ion	83	Resp	260264
Ion Ratio	Lower	Upper	
83	100		
85	63.2	54.5	81.7




Abundance

Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0









#31
Cyclohexane
Concen: 24.526 ug/l
RT: 3.72 min Scan# 313
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 56 Resp: 96963
Ion Ratio Lower Upper
56 100
69 29.8 26.0 39.0
84 88.7 71.4 107.2



#32
1,1,1-Trichloroethane
Concen: 39.891 ug/l
RT: 4.12 min Scan# 354
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 97 Resp: 136595
Ion Ratio Lower Upper
97 100
99 69.9 51.6 77.4
61 50.8 36.2 54.2





#33

1,2-Dichloroethane-d4

Concen: 45.323 ug/l

RT: 4.87 min Scan# 430

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 65 Resp: 122675

Ion Ratio Lower Upper

65 100

67 49.7 0.0 94.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.62 min Scan# 506

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10




Tgt Ion: 114 Resp: 356539


Ion Ratio Lower Upper

114 100

63 21.8 0.0 41.6

88 19.8 0.0 38.0









#35
 Dibromofluoromethane
 Concen: 48.167 ug/l
 RT: 4.14 min Scan# 356
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion:113 Resp: 137569
 Ion Ratio Lower Upper
 113 100
 111 103.1 85.2 127.8
 192 12.0 12.8 19.2#



#36
 1,1-Dichloropropene
 Concen: 47.785 ug/l
 RT: 4.31 min Scan# 373
 Delta R.T. 0.04 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion: 75 Resp: 205650
 Ion Ratio Lower Upper
 75 100
 110 32.5 16.1 48.2
 77 31.2 24.5 36.7





#37

Ethyl Acetate

Concen: 49.099 ug/l

RT: 4.15 min Scan# 357

Delta R.T. 0.04 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 43 Resp: 82442

Ion Ratio Lower Upper

43 100

61 49.8 56.6 84.8#




70 8.9 6.3 9.5



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 70.00 (69.70 to 70.70): VF0



#38

Carbon Tetrachloride

Concen: 43.310 ug/l

RT: 4.02 min Scan# 344

Delta R.T. 0.04 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Tgt Ion: 117 Resp: 141945

Ion Ratio Lower Upper

117 100

119 96.0 82.1 123.1



121 30.9 25.5 38.3




Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V









#39
Methylvlcyclohexane
Concen: 44.494 ug/l
RT: 5.48 min Scan# 492
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion: 83 Resp: 171790
Ion Ratio Lower Upper
83 100
55 85.2 70.6 105.8
98 47.1 38.5 57.7




Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 55.00 (54.70 to 55.70): VF0
Ion 98.00 (97.70 to 98.70): VF0






#40
Benzene
Concen: 49.425 ug/l
RT: 4.67 min Scan# 409
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 78 Resp: 457431
Ion Ratio Lower Upper
78 100
77 25.1 20.7 31.1



Abundance Ion 78.10 (77.80 to 78.80): VF0
Ion 77.00 (76.70 to 77.70): VF0






#41
Methacrylonitrile
Concen: 48.023 ug/l
RT: 4.78 min Scan# 421
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:	41	Resp:	43648
Ion	Ratio	Lower	Upper
41	100		
39	54.1	45.7	68.5
67	59.3	38.9	58.3#
52	36.0	35.2	52.8






Abundance

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0


Ion 67.00 (66.70 to 67.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#42
1,2-Dichloroethane
Concen: 45.797 ug/l
RT: 4.97 min Scan# 440
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10



Tgt Ion:	62	Resp:	161015
Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	16.6




Abundance

Ion 62.00 (61.70 to 62.70): VF0

Ion 98.00 (97.70 to 98.70): VF0





#43

Isopropyl Acetate

Concen: 56.258 ug/l

RT: 6.99 min Scan# 645

Delta R.T. 0.04 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 43 Resp: 116905

Ion Ratio Lower Upper




43 100

61 30.0 23.8 35.8

87 0.4 0.0 0.0#



Abundance Ion 43.00 (42.70 to 43.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 49.212 ug/l

RT: 5.53 min Scan# 497

Delta R.T. 0.03 min

Lab File: VF061118.D


Acq: 21 Dec 2018 11:10

Tgt Ion:130 Resp: 143595



Ion Ratio Lower Upper


130 100

95 101.1 0.0 205.2



Abundance Ion 129.90 (129.60 to 130.60): V
Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 47.288 ug/l

RT: 6.26 min Scan# 571

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 63 Resp: 119072

Ion Ratio Lower Upper

63 100

65 32.4 24.4 36.6



#46

Dibromomethane

Concen: 47.523 ug/l

RT: 6.12 min Scan# 556

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10




Tgt Ion: 93 Resp: 77611


Ion Ratio Lower Upper

93 100

95 79.8 66.9 100.3

174 90.9 74.1 111.1





#47

Bromodichloromethane

Concen: 47.418 ug/l

RT: 6.41 min Scan# 586

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 201208

Ion Ratio Lower Upper

83 100

85 63.8 52.2 78.4




127 7.4 6.1 9.1



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 52.542 ug/l

RT: 6.75 min Scan# 620

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Tgt Ion: 41 Resp: 76222

Ion Ratio Lower Upper

41 100

69 71.8 52.7 79.1



39 65.7 47.8 71.6




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#49
 1,4-Dioxane
 Concen: 1082.353 ug/l
 RT: 6.75 min Scan# 620
 Delta R.T. 0.04 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion: 88 Resp: 11393
 Ion Ratio Lower Upper
 88 100
 43 52.4 33.3 49.9#
 58 71.0 53.2 79.8



#50
 Toluene-d8
 Concen: 51.470 ug/l
 RT: 7.59 min Scan# 705
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion: 98 Resp: 416585
 Ion Ratio Lower Upper
 98 100
 100 71.7 55.3 82.9





#51

4-Methyl-2-Pentanone

Concen: 250.336 ug/l

RT: 8.29 min Scan# 776

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 356768

Ion Ratio Lower Upper




43 100

58 37.7 29.1 43.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 46.171 ug/l

RT: 7.65 min Scan# 712

Delta R.T. 0.03 min

Lab File: VF061118.D


Acq: 21 Dec 2018 11:10

Tgt Ion: 92 Resp: 282863

Ion Ratio Lower Upper



92 100


91 163.7 128.8 193.2



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0





#53

t-1,3-Dichloropropene

Concen: 48.273 ug/l

RT: 8.31 min Scan# 778

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 75 Resp: 165245

Ion Ratio Lower Upper

75 100

77 30.9 25.8 38.8



#54

cis-1,3-Dichloropropene

Concen: 48.093 ug/l

RT: 7.33 min Scan# 679

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10




Tgt Ion: 75 Resp: 206619


Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.4

39 62.0 50.1 75.1









#55
 1,1,2-Trichloroethane
 Concen: 49.478 ug/l
 RT: 8.52 min Scan# 800
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Tot Ion: 97 Resp: 90871
 Ion Ratio Lower Upper
 97 100
 83 87.1 75.0 112.4
 85 57.2 47.5 71.3
 99 62.0 51.9 77.9




Abundance Ion 97.00 (96.70 to 97.70): VF0
 Ion 82.95 (82.65 to 83.65): VF0
 Ion 84.95 (84.65 to 85.65): VF0
 Ion 98.95 (98.65 to 99.65): VF0






#56
 Ethyl methacrylate
 Concen: 51.909 ug/l
 RT: 8.65 min Scan# 813
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion: 69 Resp: 103160
 Ion Ratio Lower Upper
 69 100
 41 77.5 59.7 89.5
 39 42.3 35.0 52.6



Abundance Ion 69.00 (68.70 to 69.70): VF0
 Ion 41.00 (40.70 to 41.70): VF0
 Ion 39.00 (38.70 to 39.70): VF0









#57
 1,3-Dichloropropane
 Concen: 48.407 ug/l
 RT: 8.89 min Scan# 837
 Delta R.T. 0.04 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion: 76 Resp: 160793
 Ion Ratio Lower Upper
 76 100
 78 33.8 26.5 39.7



#58
 2-Chloroethyl Vinyl ether
 Concen: 286.754 ug/l
 RT: 7.33 min Scan# 679
 Delta R.T. 0.04 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion: 63 Resp: 109449
 Ion Ratio Lower Upper
 63 100
 106 26.0 19.8 29.8









#59
2-Hexanone
Concen: 294.338 ug/l
RT: 9.52 min Scan# 901
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 43 Resp: 303357
Ion Ratio Lower Upper
43 100
58 52.4 25.1 75.2



#60
Dibromochloromethane
Concen: 48.449 ug/l
RT: 8.75 min Scan# 823
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 129 Resp: 127552
Ion Ratio Lower Upper
129 100
127 82.9 36.8 110.3





#61

1,2-Dibromoethane

Concen: 48.655 ug/l

RT: 9.02 min Scan# 850

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:107 Resp: 95435

Ion Ratio Lower Upper




107 100

109 93.3 73.7 110.5



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 47.042 ug/l

RT: 11.43 min Scan# 1095

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Tgt Ion: 95 Resp: 175529

Ion Ratio Lower Upper

95 100

174 63.3 0.0 143.8



176 62.0 0.0 147.6




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.81 min Scan# 930
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:117 Resp: 327354
Ion Ratio Lower Upper
117 100
82 60.1 45.0 67.4
119 30.6 24.8 37.2



#64
Tetrachloroethene
Concen: 50.071 ug/l
RT: 8.19 min Scan# 766
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion:164 Resp: 126088
Ion Ratio Lower Upper
164 100
166 125.7 104.0 156.0
129 93.1 81.1 121.7
131 92.7 81.2 121.8









#65
Chlorobenzene
Concen: 47.975 ug/l
RT: 9.83 min Scan# 933
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:112 Resp: 324439
Ion Ratio Lower Upper
112 100
114 33.0 27.8 41.6



#66
1,1,1,2-Tetrachloroethane
Concen: 47.099 ug/l
RT: 9.96 min Scan# 946
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion:131 Resp: 130164
Ion Ratio Lower Upper
131 100
133 97.6 49.5 148.7
119 65.4 34.5 103.5









#67
Ethyl Benzene
Concen: 47.952 ug/l
RT: 9.93 min Scan# 943
Delta R.T. 0.04 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 590782
Ion Ratio Lower Upper
91 100
106 29.5 22.5 33.7



#68
m/p-Xylenes
Concen: 97.583 ug/l
RT: 10.16 min Scan# 966
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 106 Resp: 423103
Ion Ratio Lower Upper
106 100
91 205.9 168.2 252.2









#69
o-Xylene
Concen: 45.173 ug/l
RT: 10.74 min Scan# 1025
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:106 Resp: 217223
Ion Ratio Lower Upper
106 100
91 217.0 115.9 347.7



#70
Styrene
Concen: 53.872 ug/l
RT: 10.81 min Scan# 1032
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion:104 Resp: 313455
Ion Ratio Lower Upper
104 100
78 48.7 38.4 57.6
103 45.9 37.0 55.6





#71

Bromoform

Concen: 48.371 ug/l

RT: 10.80 min Scan# 1031

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:173 Resp: 61847

Ion Ratio Lower Upper

173 100

175 50.5 24.8 74.4




254 0.0 0.0 0.0



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.58 min Scan# 1211

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Tgt Ion:152 Resp: 156227

Ion Ratio Lower Upper

152 100

115 54.2 33.3 99.9



150 164.9 0.0 328.2




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 48.957 ug/l

RT: 11.15 min Scan# 1066

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion:105 Resp: 659097

Ion Ratio Lower Upper

105 100

120 25.3 13.0 39.0



#74

N-ethyl acetate

Concen: 54.543 ug/l

RT: 11.39 min Scan# 1091

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Tgt Ion: 43 Resp: 163029




Ion Ratio Lower Upper


43 100

70 35.0 29.2 43.8

55 19.9 16.5 24.7

61 24.4 20.7 31.1





#75

1,1,2,2-Tetrachloroethane

Concen: 48.416 ug/l

RT: 11.72 min Scan# 1124

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 108180

Ion	Ratio	Lower	Upper
83	100		
131	10.2	4.8	14.3
85	66.9	33.9	101.6

83 100

131 10.2 4.8 14.3




85 66.9 33.9 101.6



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): V

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 47.973 ug/l

RT: 11.82 min Scan# 1134

Delta R.T. 0.02 min

Lab File: VF061118.D


Acq: 21 Dec 2018 11:10

Tgt Ion: 75 Resp: 79771

Ion	Ratio	Lower	Upper
75	100		
77	34.1	17.6	52.9



75 100


77 34.1 17.6 52.9



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 47.484 ug/l

RT: 11.52 min Scan# 1104

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:156 Resp: 137614

Ion Ratio Lower Upper

156 100

77 166.3 93.8 281.4




158 95.7 48.9 146.8



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 47.686 ug/l

RT: 11.62 min Scan# 1114

Delta R.T. 0.03 min

Lab File: VF061118.D


Acq: 21 Dec 2018 11:10

Tgt Ion: 91 Resp: 746268

Ion Ratio Lower Upper



91 100


120 23.2 10.4 31.2



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 46.821 ug/l
 RT: 11.74 min Scan# 1126
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion: 91 Resp: 432420
 Ion Ratio Lower Upper
 91 100
 126 29.8 16.4 49.1



#80
 1,3,5-Trimethylbenzene
 Concen: 48.195 ug/l
 RT: 11.85 min Scan# 1137
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion: 105 Resp: 534272
 Ion Ratio Lower Upper
 105 100
 120 51.5 25.4 76.4





#81

trans-1,4-Dichloro-2-butene

Concen: 68.075 ug/l

RT: 11.89 min Scan# 1141

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 75 Resp: 48642

Ion Ratio Lower Upper

75 100

53 80.7 74.1 111.1

89 48.7 42.6 64.0



#82

4-Chlorotoluene

Concen: 45.745 ug/l

RT: 11.93 min Scan# 1145

Delta R.T. 0.03 min

Lab File: VF061118.D




Acq: 21 Dec 2018 11:10


Tgt Ion: 91 Resp: 437045

Ion Ratio Lower Upper

91 100

126 34.7 16.0 47.9









#83
 tert-Butylbenzene
 Concen: 45.757 ug/l
 RT: 12.15 min Scan# 1168
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion:119 Resp: 521971
 Ion Ratio Lower Upper
 119 100
 91 60.6 30.5 91.5
 134 24.1 12.4 37.4



#84
 1,2,4-Trimethylbenzene
 Concen: 46.697 ug/l
 RT: 12.23 min Scan# 1176
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion:105 Resp: 516228
 Ion Ratio Lower Upper
 105 100
 120 48.9 23.6 71.0





#85

sec-Butylbenzene

Concen: 47.529 ug/l

RT: 12.33 min Scan# 1186

Delta R.T. 0.03 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:105 Resp: 707560

Ion Ratio Lower Upper




105 100

134 19.6 9.4 28.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 49.155 ug/l

RT: 12.48 min Scan# 1201

Delta R.T. 0.02 min

Lab File: VF061118.D

Acq: 21 Dec 2018 11:10


Tgt Ion:119 Resp: 607397

Ion Ratio Lower Upper

119 100

134 24.9 12.8 38.4



91 24.3 12.2 36.6




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 55.390 ug/l
 RT: 12.50 min Scan# 1203
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion	Ratio	Lower	Upper
146	100		
111	40.9	19.3	57.9
148	66.7	32.4	97.2



#88
 1,4-Dichlorobenzene
 Concen: 46.967 ug/l
 RT: 12.59 min Scan# 1212
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion	Ratio	Lower	Upper
146	100		
111	39.4	19.4	58.4
148	66.8	32.4	97.0









#89
n-Butylbenzene
Concen: 47.992 ug/l
RT: 12.86 min Scan# 1240
Delta R.T. 0.02 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 604694
Ion Ratio Lower Upper
91 100
92 64.0 32.8 98.4
134 24.6 13.2 39.5



#90
Hexachloroethane
Concen: 47.258 ug/l
RT: 12.94 min Scan# 1248
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion: 117 Resp: 144274
Ion Ratio Lower Upper
117 100
201 67.0 24.9 74.8









#91
 1,2-Dichlorobenzene
 Concen: 46.052 ug/l
 RT: 12.96 min Scan# 1250
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion	Ratio	Lower	Upper
146	100		
111	43.5	23.2	69.5
148	62.7	32.4	97.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 52.627 ug/l
 RT: 13.66 min Scan# 1321
 Delta R.T. 0.02 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion	Ratio	Lower	Upper
75	100		
155	68.5	41.5	124.5
157	86.3	53.2	159.6









#93
 1,2,4-Trichlorobenzene
 Concen: 47.920 ug/l
 RT: 14.22 min Scan# 1378
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion:180 Resp: 176076
 Ion Ratio Lower Upper
 180 100
 182 94.6 46.9 140.6
 145 32.6 20.5 61.6



#94
 Hexachlorobutadiene
 Concen: 50.001 ug/l
 RT: 14.21 min Scan# 1377
 Delta R.T. 0.03 min
 Lab File: VF061118.D
 Acq: 21 Dec 2018 11:10

Tgt Ion:225 Resp: 117014
 Ion Ratio Lower Upper
 225 100
 223 59.6 32.8 98.3
 227 63.5 34.0 101.8









#95
Naphthalene
Concen: 49.980 ug/l
RT: 14.48 min Scan# 1404
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:128 Resp: 335342
Ion Ratio Lower Upper
128 100
127 12.4 9.4 14.2
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 46.632 ug/l
RT: 14.63 min Scan# 1419
Delta R.T. 0.03 min
Lab File: VF061118.D
Acq: 21 Dec 2018 11:10

Tgt Ion:180 Resp: 158399
Ion Ratio Lower Upper
180 100
182 99.0 47.9 143.8
145 32.7 18.6 55.8

