

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF122118\
 Data File : VF061129.D
 Acq On : 21 Dec 2018 16:44
 Operator : VA/AP
 Sample : J6475-01RE
 Misc : 5.03g/5mL/MSVOA-F/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SLUDGERE

Quant Time: Dec 21 17:13:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-4.87
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-5.59
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-9.78
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-12.55

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		4.70	65	63	0.00	ug/l	-0.15
Spiked Amount	50.000			Recovery	=	0.00%	
35) Dibromofluoromethane		0.00	113	0	0.00	ug/l	
Spiked Amount	50.000			Recovery	=	0.00%	
50) Toluene-d8		0.00	98	0	0.00	ug/l	
Spiked Amount	50.000			Recovery	=	0.00%	
62) 4-Bromofluorobenzene		11.56	95	72	0.00	ug/l	0.15
Spiked Amount	50.000			Recovery	=	0.00%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF122118\
Data File : VF061129.D
Acq On : 21 Dec 2018 16:44
Operator : VA/AP
Sample : J6475-01RE
Misc : 5.03g/5mL/MSVOA-F/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SLUDGERE

Quant Time: Dec 21 17:13:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 00:56:59 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 0.000 ug/l

Expected RT: 4.87 min

Lab File: VF061129.D

Acq: 21 Dec 2018 16:44

Tot Ion: 168

Sig Exp Ratio

168 100



99 68.0

Instrument :

MSVOA_F

ClientSampleId :

SLUDGERE



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

RT: 4.70 min Scan# 412

Delta R.T. -0.15 min

Lab File: VF061129.D




Acq: 21 Dec 2018 16:44


Tgt Ion: 65 Resp: 63

Ion Ratio Lower Upper

65 100

67 0.0 0.0 94.6







#34
 1,4-Difluorobenzene
 Concen: 0.000 ug/l
 Expected RT: 5.59 min
 Lab File: VF061129.D
 Acq: 21 Dec 2018 16:44


Instrument :
 MSVOA_F
 ClientSampleId :
 SLUDGERE


Tot Ion: 114
 Sig Exp Ratio
 114 100
 63 20.8
 88 19.0



#35
 Dibromofluoromethane
 Concen: 0.000 ug/l
 Expected RT: 4.11 min
 Lab File: VF061129.D
 Acq: 21 Dec 2018 16:44

Tgt Ion: 113
 Sig Exp Ratio
 113 100
 111 106.5
 192 16.0







#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 7.56 min

Lab File: VF061129.D
Acq: 21 Dec 2018 16:44




Instrument :
MSVOA_F
ClientSampleId :
SLUDGERE


Tot Ion: 98
Sig Exp Ratio
98 100
100 69.1



#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
RT: 11.56 min Scan# 1108
Delta R.T. 0.15 min
Lab File: VF061129.D
Acq: 21 Dec 2018 16:44

Tgt Ion: 95 Resp: 72
Ion Ratio Lower Upper
95 100
174 0.0 0.0 143.8
176 0.0 0.0 147.6







#63
 Chlorobenzene-d5
 Concen: 0.000 ug/l
 Expected RT: 9.78 min

Lab File: VF061129.D
 Acq: 21 Dec 2018 16:44

Instrument :
 MSVOA_F
 ClientSampleId :
 SLUDGERE

Tot Ion: 117
 Sig Exp Ratio
 117 100
 82 56.2
 119 31.0



#72
 1,4-Dichlorobenzene-d4
 Concen: 0.000 ug/l
 Expected RT: 12.55 min

Lab File: VF061129.D
 Acq: 21 Dec 2018 16:44

Tgt Ion: 152
 Sig Exp Ratio
 152 100
 115 66.6
 150 164.1

