

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF122218\
 Data File : VF061140.D
 Acq On : 22 Dec 2018 12:49
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 24 02:54:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.90	168	178973	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	5.63	114	323360	50.00	ug/l	0.03
63) Chlorobenzene-d5	9.80	117	297922	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.57	152	136222	50.00	ug/l	0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.87	65	125926	54.71	ug/l	0.02
Spiked Amount			Recovery	=	109.42%	
35) Dibromofluoromethane	4.14	113	136612	52.74	ug/l	0.02
Spiked Amount			Recovery	=	105.48%	
50) Toluene-d8	7.58	98	399863	54.47	ug/l	0.02
Spiked Amount			Recovery	=	108.94%	
62) 4-Bromofluorobenzene	11.43	95	167240	49.42	ug/l	0.01
Spiked Amount			Recovery	=	98.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.96	85	203619	51.180	ug/l	97
3) Chloromethane	1.11	50	136091	59.717	ug/l	94
4) Vinyl Chloride	1.14	62	123569	58.451	ug/l #	86
5) Bromomethane	1.35	94	80977	55.225	ug/l	91
6) Chloroethane	1.40	64	55782	57.264	ug/l	93
7) Trichlorofluoromethane	1.49	101	153235	37.962	ug/l	99
8) Diethyl Ether	1.63	74	33627	60.835	ug/l	95
9) 1,1,2-Trichlorotrifluoroet	1.85	101	112727	53.861	ug/l	97
10) Methyl Iodide	1.93	142	147263	44.638	ug/l	93
11) Tert butyl alcohol	2.59	59	26952	290.146	ug/l	98
12) 1,1-Dichloroethene	1.81	96	89183	56.529	ug/l	91
13) Acrolein	2.01	56	9820	146.299	ug/l #	81
14) Allyl chloride	2.11	41	102092	62.183	ug/l	86
15) Acrylonitrile	2.96	53	64376	299.701	ug/l	94
16) Acetone	2.25	43	84327	183.250	ug/l	98
17) Carbon Disulfide	1.85	76	195284	40.960	ug/l	98
18) Methyl Acetate	2.35	43	58542	61.490	ug/l	95
19) Methyl tert-butyl Ether	2.44	73	201316	57.489	ug/l	100
20) Methylene Chloride	2.19	84	28452	16.323	ug/l	87
21) trans-1,2-Dichloroethene	2.32	96	72633	43.045	ug/l	96
22) Diisopropyl ether	2.82	45	317201	58.002	ug/l	97
23) Vinyl Acetate	3.21	43	720065	285.411	ug/l	98
24) 1,1-Dichloroethane	2.89	63	179933	55.530	ug/l	98
25) 2-Butanone	4.35	43	197905	265.176	ug/l	98
26) 2,2-Dichloropropane	3.62	77	77902	41.894	ug/l	98
27) cis-1,2-Dichloroethene	3.50	96	150917	58.422	ug/l	87
28) Bromochloromethane	3.74	49	93908	57.118	ug/l	95
29) Tetrahydrofuran	4.10	42	98538	324.396	ug/l	97
30) Chloroform	3.88	83	266781	54.910	ug/l	96
31) Cyclohexane	3.71	56	106691	31.731	ug/l	97
32) 1,1,1-Trichloroethane	4.12	97	149914	51.479	ug/l	94
36) 1,1-Dichloropropene	4.31	75	198923	50.964	ug/l	98
37) Ethyl Acetate	4.14	43	86588	58.875	ug/l #	88
38) Carbon Tetrachloride	4.02	117	150580	50.659	ug/l	93

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF122218\
 Data File : VF061140.D
 Acq On : 22 Dec 2018 12:49
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 24 02:54:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.48	83	210744	60.184	ug/l	94
40) Benzene	4.66	78	448278	53.406	ug/l	93
41) Methacrylonitrile	4.77	41	45368	55.037	ug/l	93
42) 1,2-Dichloroethane	4.97	62	165128	51.786	ug/l	99
43) Isopropyl Acetate	6.99	43	110229	58.488	ug/l #	99
44) Trichloroethene	5.53	130	138455	52.319	ug/l	100
45) 1,2-Dichloropropane	6.26	63	124650	54.583	ug/l	99
46) Dibromomethane	6.11	93	80232	54.168	ug/l	92
47) Bromodichloromethane	6.41	83	198510	51.582	ug/l	96
48) Methyl methacrylate	6.74	41	75920	57.704	ug/l	97
49) 1,4-Dioxane	6.73	88	11670	1222.425	ug/l #	76
51) 4-Methyl-2-Pentanone	8.28	43	347488	268.842	ug/l	97
52) Toluene	7.65	92	292623	52.665	ug/l	94
53) t-1,3-Dichloropropene	8.30	75	144974	46.697	ug/l	100
54) cis-1,3-Dichloropropene	7.32	75	189218	48.561	ug/l	98
55) 1,1,2-Trichloroethane	8.52	97	87958	52.806	ug/l	95
56) Ethyl methacrylate	8.65	69	105861	58.733	ug/l	95
57) 1,3-Dichloropropane	8.87	76	165560	54.956	ug/l	98
58) 2-Chloroethyl Vinyl ether	7.31	63	95003	274.445	ug/l	96
59) 2-Hexanone	9.51	43	241029	257.859	ug/l	99
60) Dibromochloromethane	8.74	129	128483	53.810	ug/l	99
61) 1,2-Dibromoethane	9.01	107	98423	55.327	ug/l	97
64) Tetrachloroethene	8.18	164	142136	62.020	ug/l	93
65) Chlorobenzene	9.82	112	334489	54.347	ug/l	99
66) 1,1,1,2-Tetrachloroethane	9.96	131	133937	53.252	ug/l	95
67) Ethyl Benzene	9.93	91	587775	52.421	ug/l	98
68) m/p-Xylenes	10.16	106	398770	101.057	ug/l	97
69) o-Xylene	10.74	106	233489	53.352	ug/l	93
70) Styrene	10.82	104	298696	56.662	ug/l	99
71) Bromoform	10.80	173	63023	54.161	ug/l #	100
73) Isopropylbenzene	11.14	105	667117	56.829	ug/l	98
74) N-amyl acetate	11.39	43	155437	59.640	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.71	83	118087	60.611	ug/l	99
76) 1,2,3-Trichloropropane	11.81	75	80454	55.489	ug/l	100
77) Bromobenzene	11.52	156	136944	54.192	ug/l	86
78) n-propylbenzene	11.61	91	757525	57.963	ug/l	99
79) 2-Chlorotoluene	11.73	91	450269	55.914	ug/l	95
80) 1,3,5-Trimethylbenzene	11.84	105	515667	53.348	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.89	75	44527	71.468	ug/l	93
82) 4-Chlorotoluene	11.92	91	432757	51.948	ug/l	98
83) tert-Butylbenzene	12.15	119	534602	53.747	ug/l	98
84) 1,2,4-Trimethylbenzene	12.22	105	516003	53.532	ug/l	97
85) sec-Butylbenzene	12.33	105	716682	55.212	ug/l	98
86) p-Isopropyltoluene	12.47	119	553162	51.340	ug/l	98
87) 1,3-Dichlorobenzene	12.49	146	237778	58.581	ug/l	99
88) 1,4-Dichlorobenzene	12.58	146	250228	53.926	ug/l	99
89) n-Butylbenzene	12.86	91	598490	56.830	ug/l	97
90) Hexachloroethane	12.94	117	154253	57.947	ug/l	82
91) 1,2-Dichlorobenzene	12.96	146	245375	54.315	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.66	75	22777	63.709	ug/l	91


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF122218\
Data File : VF061140.D
Acq On : 22 Dec 2018 12:49
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Quant Time: Dec 24 02:54:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 00:56:59 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.22	180	168070	52.459	ug/l	95
94) Hexachlorobutadiene	14.21	225	109324	53.575	ug/l	99
95) Naphthalene	14.48	128	348315	59.537	ug/l	99
96) 1,2,3-Trichlorobenzene	14.61	180	160503	54.190	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.90 min Scan# 433

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:168 Resp: 178973

Ion Ratio Lower Upper




168 100

99 65.9 54.4 81.6



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0



#2

Dichlorodifluoromethane

Concen: 51.180 ug/l

RT: 0.96 min Scan# 34

Delta R.T. 0.00 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 85 Resp: 203619

Ion Ratio Lower Upper



85 100

87 33.2 15.9 47.6



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0





#3

Chloromethane

Concen: 59.717 ug/l

RT: 1.11 min Scan# 49

Delta R.T. 0.00 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 50 Resp: 136091

Ion Ratio Lower Upper




50 100

52 36.5 26.3 39.5



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#4

Vinyl Chloride

Concen: 58.451 ug/l

RT: 1.14 min Scan# 52

Delta R.T. 0.00 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 62 Resp: 123569

Ion Ratio Lower Upper



62 100

64 29.8 30.4 45.6#



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0





#5

Bromomethane

Concen: 55.225 ug/l

RT: 1.35 min Scan# 73

Delta R.T. 0.00 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 94 Resp: 80977

Ion Ratio Lower Upper

94 100

96 84.1 74.4 111.6



#6

Chloroethane

Concen: 57.264 ug/l

RT: 1.40 min Scan# 78

Delta R.T. 0.00 min

Lab File: VF061140.D




Acq: 22 Dec 2018 12:49

Tgt Ion: 64 Resp: 55782

Ion Ratio Lower Upper

64 100

66 34.5 30.9 46.3





#7

Trichlorofluoromethane

Concen: 37.962 ug/l

RT: 1.49 min Scan# 88

Delta R.T. -0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:101 Resp: 153235

Ion Ratio Lower Upper




101 100

103 64.2 50.7 76.1



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 60.835 ug/l

RT: 1.63 min Scan# 102

Delta R.T. 0.00 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 74 Resp: 33627

Ion Ratio Lower Upper



74 100

45 141.6 67.8 203.4



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 53.861 ug/l

RT: 1.85 min Scan# 124

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:101 Resp: 112727

Ion Ratio Lower Upper

101 100

85 52.0 44.5 66.7




151 61.6 49.9 74.9



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 44.638 ug/l

RT: 1.93 min Scan# 132

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion:142 Resp: 147263

Ion Ratio Lower Upper

142 100

127 61.7 54.7 82.1



141 16.6 12.2 18.2




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
 Concen: 290.146 ug/l
 RT: 2.59 min Scan# 199
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050




Tot Ion: 59 Resp: 26952
 Ion Ratio Lower Upper
 59 100
 57 15.1 11.5 17.3




#12

1,1-Dichloroethene
 Concen: 56.529 ug/l
 RT: 1.81 min Scan# 120
 Delta R.T. 0.01 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion: 96 Resp: 89183
 Ion Ratio Lower Upper
 96 100
 61 213.0 182.0 273.0
 98 66.7 46.5 69.7





#13

Acrolein

Concen: 146.299 ug/l

RT: 2.01 min Scan# 140

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 56 Resp: 9820

Ion Ratio Lower Upper




56 100

55 70.1 70.2 105.2#



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 55.00 (54.70 to 55.70): VF0



#14

Allyl chloride

Concen: 62.183 ug/l

RT: 2.11 min Scan# 150

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion: 41 Resp: 102092

Ion Ratio Lower Upper

41 100

39 101.7 96.4 144.6



76 33.9 25.4 38.0




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

Ion 76.00 (75.70 to 76.70): VF0





#15

Acrylonitrile

Concen: 299.701 ug/l

RT: 2.96 min Scan# 237

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 53 Resp: 64376

Ion Ratio Lower Upper

53 100

52 97.3 71.3 106.9

51 45.4 36.8 55.2



#16

Acetone

Concen: 183.250 ug/l

RT: 2.25 min Scan# 165

Delta R.T. 0.01 min

Lab File: VF061140.D




Acq: 22 Dec 2018 12:49

Tgt Ion: 43 Resp: 84327

Ion Ratio Lower Upper

43 100

58 14.5 11.1 16.7









#17
Carbon Disulfide
Concen: 40.960 ug/l
RT: 1.85 min Scan# 124
Delta R.T. 0.01 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 76 Resp: 195284
Ion Ratio Lower Upper
76 100
78 13.5 10.2 15.4



#18
Methyl Acetate
Concen: 61.490 ug/l
RT: 2.35 min Scan# 175
Delta R.T. 0.00 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 43 Resp: 58542
Ion Ratio Lower Upper
43 100
74 16.0 11.0 16.6









#19
Methvl tert-butvl Ether
Concen: 57.489 ug/l
RT: 2.44 min Scan# 184
Delta R.T. 0.01 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 73 Resp: 201316
Ion Ratio Lower Upper
73 100
57 18.1 14.5 21.7



#20
Methylene Chloride
Concen: 16.323 ug/l
RT: 2.19 min Scan# 159
Delta R.T. -0.06 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 84 Resp: 28452
Ion Ratio Lower Upper
84 100
49 140.4 129.4 194.2
51 45.1 44.0 66.0
86 59.1 46.3 69.5





#21

trans-1,2-Dichloroethene

Concen: 43.045 ug/l

RT: 2.32 min Scan# 172

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 96 Resp: 72633

Ion Ratio Lower Upper

96 100

61 176.1 146.0 219.0




98 66.9 51.6 77.4



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 58.002 ug/l

RT: 2.82 min Scan# 222

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Tgt Ion: 45 Resp: 317201


Ion Ratio Lower Upper

45 100

43 50.7 42.2 63.2

87 19.2 14.6 22.0

59 9.5 7.0 10.4






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 285.411 ug/l

RT: 3.21 min Scan# 262

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 720065

Ion Ratio Lower Upper

43 100

86 7.3 6.3 9.5



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0

3.21

250000

200000

150000



100000

50000

0

Time-->

3.10 3.20 3.30 3.40



#24

1,1-Dichloroethane

Concen: 55.530 ug/l

RT: 2.89 min Scan# 229

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion: 63 Resp: 179933

Ion Ratio Lower Upper

63 100

98 3.6 2.1 6.5

100 2.9 1.7 5.1



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0

2.89

80000

60000


40000


20000

0

Time-->

2.80 2.90 3.00





#25

2-Butanone

Concen: 265.176 ug/l

RT: 4.35 min Scan# 378

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 197905

Ion Ratio Lower Upper

43 100

72 20.6 15.6 23.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

4.35

50000

40000

30000



20000

10000

0

Time-->

4.20 4.30 4.40 4.50



#26

2,2-Dichloropropane

Concen: 41.894 ug/l

RT: 3.62 min Scan# 303

Delta R.T. 0.02 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 77 Resp: 77902

Ion Ratio Lower Upper

77 100

97 17.3 8.2 24.4



Abundance Ion 77.00 (76.70 to 77.70): VF0

Ion 97.00 (96.70 to 97.70): VF0

3.62

25000

20000

15000


10000

5000

0

Time-->

3.50 3.60 3.70 3.80






#27




cis-1,2-Dichloroethene
Concen: 58.422 ug/l
RT: 3.50 min Scan# 291
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion	96	Resp	150917
Ion	Ratio	Lower	Upper
96	100		
61	125.0	0.0	287.8
98	61.1	0.0	134.2




Abundance Ion 96.00 (95.70 to 96.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 98.00 (97.70 to 98.70): VF0





#28


Bromochloromethane
Concen: 57.118 ug/l
RT: 3.74 min Scan# 316
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion	49	Resp	93908
Ion	Ratio	Lower	Upper
49	100		
129	2.8	0.0	3.6
130	78.0	65.8	98.6



Abundance Ion 49.00 (48.70 to 49.70): VF0
Ion 129.00 (128.70 to 129.70): V
Ion 130.00 (129.70 to 130.70): V





#29

Tetrahydrofuran

Concen: 324.396 ug/l

RT: 4.10 min Scan# 352

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 42 Resp: 98538

Ion Ratio Lower Upper




42 100

72 36.1 31.8 47.8

71 37.4 29.7 44.5



Abundance Ion 42.00 (41.70 to 42.70): VF0
Ion 72.00 (71.70 to 72.70): VF0
Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 54.910 ug/l

RT: 3.88 min Scan# 330

Delta R.T. 0.02 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 83 Resp: 266781



Ion Ratio Lower Upper


83 100

85 65.2 54.5 81.7



Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 31.731 ug/l

RT: 3.71 min Scan# 313

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 56 Resp: 106691

Ion Ratio Lower Upper

56 100

69 34.7 26.0 39.0




84 87.3 71.4 107.2



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 51.479 ug/l

RT: 4.12 min Scan# 354

Delta R.T. 0.03 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion: 97 Resp: 149914

Ion Ratio Lower Upper

97 100

99 59.5 51.6 77.4



61 48.5 36.2 54.2




Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

Ion 61.00 (60.70 to 61.70): VF0









#33
 1,2-Dichloroethane-d4
 Concen: 54.705 ug/l
 RT: 4.87 min Scan# 430
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Tot Ion: 65 Resp: 125926
 Ion Ratio Lower Upper
 65 100
 67 51.7 0.0 94.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.63 min Scan# 507
 Delta R.T. 0.03 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion: 114 Resp: 323360
 Ion Ratio Lower Upper
 114 100
 63 18.3 0.0 41.6
 88 18.7 0.0 38.0






#35
 Dibromofluoromethane
 Concen: 52.740 ug/l
 RT: 4.14 min Scan# 356
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
113	100		
111	97.1	85.2	127.8
192	14.0	12.8	19.2






Abundance

Ion 113.00 (112.70 to 113.70): V


Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#36
 1,1-Dichloropropene
 Concen: 50.964 ug/l
 RT: 4.31 min Scan# 373
 Delta R.T. 0.03 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion	Ratio	Lower	Upper
75	100		
110	31.0	16.1	48.2
77	31.9	24.5	36.7






Abundance

Ion 75.00 (74.70 to 75.70): VF0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VF0





#37

Ethyl Acetate

Concen: 58.875 ug/l

RT: 4.14 min Scan# 356

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 43 Resp: 86588

Ion Ratio Lower Upper

43 100

61 81.5 56.6 84.8




70 10.0 6.3 9.5#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 70.00 (69.70 to 70.70): VF0



#38

Carbon Tetrachloride

Concen: 50.659 ug/l

RT: 4.02 min Scan# 344

Delta R.T. 0.03 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion: 117 Resp: 150580

Ion Ratio Lower Upper

117 100

119 94.5 82.1 123.1



121 30.8 25.5 38.3




Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V









#39
Methvlcvclohexane
Concen: 60.184 ug/l
RT: 5.48 min Scan# 492
Delta R.T. 0.03 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion: 83 Resp: 210744
Ion Ratio Lower Upper
83 100
55 82.4 70.6 105.8
98 51.1 38.5 57.7




Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 55.00 (54.70 to 55.70): VF0
Ion 98.00 (97.70 to 98.70): VF0






#40
Benzene
Concen: 53.406 ug/l
RT: 4.66 min Scan# 409
Delta R.T. 0.03 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 78 Resp: 448278
Ion Ratio Lower Upper
78 100
77 22.5 20.7 31.1



Abundance Ion 78.10 (77.80 to 78.80): VF0
Ion 77.00 (76.70 to 77.70): VF0









#41
Methacrylonitrile
Concen: 55.037 ug/l
RT: 4.77 min Scan# 420
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:	41	Resp:	45368
Ion	Ratio	Lower	Upper
41	100		
39	60.5	45.7	68.5
67	55.6	38.9	58.3
52	49.2	35.2	52.8



#42
1,2-Dichloroethane
Concen: 51.786 ug/l
RT: 4.97 min Scan# 440
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion:	62	Resp:	165128
Ion	Ratio	Lower	Upper
62	100		
98	7.9	0.0	16.6









#43

Isopropyl Acetate
 Concen: 58.488 ug/l
 RT: 6.99 min Scan# 645
 Delta R.T. 0.03 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050




Tot Ion: 43 Resp: 110229
 Ion Ratio Lower Upper
 43 100
 61 29.1 23.8 35.8
 87 0.5 0.0 0.0#



#44

Trichloroethene
 Concen: 52.319 ug/l
 RT: 5.53 min Scan# 497
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion: 130 Resp: 138455
 Ion Ratio Lower Upper
 130 100
 95 102.5 0.0 205.2









#45
 1,2-Dichloropropane
 Concen: 54.583 ug/l
 RT: 6.26 min Scan# 571
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion: 63 Resp: 124650
 Ion Ratio Lower Upper
 63 100
 65 29.9 24.4 36.6



#46
 Dibromomethane
 Concen: 54.168 ug/l
 RT: 6.11 min Scan# 556
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion: 93 Resp: 80232
 Ion Ratio Lower Upper
 93 100
 95 75.9 66.9 100.3
 174 86.1 74.1 111.1





#47

Bromodichloromethane

Concen: 51.582 ug/l

RT: 6.41 min Scan# 586

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 198510

Ion Ratio Lower Upper

83 100

85 69.0 52.2 78.4




127 8.6 6.1 9.1



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 57.704 ug/l

RT: 6.74 min Scan# 620

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion: 41 Resp: 75920

Ion Ratio Lower Upper

41 100

69 68.7 52.7 79.1



39 57.9 47.8 71.6



Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#49
1,4-Dioxane
Concen: 1222.425 ug/l
RT: 6.73 min Scan# 619
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion: 88 Resp: 11670
Ion Ratio Lower Upper
88 100
43 55.7 33.3 49.9#
58 86.6 53.2 79.8#



#50
Toluene-d8
Concen: 54.473 ug/l
RT: 7.58 min Scan# 705
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 98 Resp: 399863
Ion Ratio Lower Upper
98 100
100 73.1 55.3 82.9





#51

4-Methyl-2-Pentanone

Concen: 268.842 ug/l

RT: 8.28 min Scan# 776

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 347488

Ion Ratio Lower Upper

43 100

58 38.2 29.1 43.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0

8.28



150000

100000

50000

0

Time-->



#52

Toluene

Concen: 52.665 ug/l

RT: 7.65 min Scan# 712

Delta R.T. 0.02 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 92 Resp: 292623

Ion Ratio Lower Upper

92 100

91 168.8 128.8 193.2



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0

7.65

200000


150000


100000

50000

0

Time-->





#53

t-1,3-Dichloropropene

Concen: 46.697 ug/l

RT: 8.30 min Scan# 778

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 75 Resp: 144974

Ion Ratio Lower Upper

75 100

77 32.3 25.8 38.8



#54

cis-1,3-Dichloropropene

Concen: 48.561 ug/l

RT: 7.32 min Scan# 679

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49




Tgt Ion: 75 Resp: 189218

Ion Ratio Lower Upper

75 100

77 32.2 25.0 37.4

39 63.8 50.1 75.1





#55

1,1,2-Trichloroethane

Concen: 52.806 ug/l

RT: 8.52 min Scan# 800

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050

Tot Ion: 97 Resp: 87958





Ion Ratio Lower Upper

97 100

83 89.4 75.0 112.4

85 57.7 47.5 71.3

99 59.3 51.9 77.9



#56

Ethyl methacrylate

Concen: 58.733 ug/l

RT: 8.65 min Scan# 813

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49




Tgt Ion: 69 Resp: 105861


Ion Ratio Lower Upper

69 100

41 78.9 59.7 89.5

39 40.6 35.0 52.6





#57

1,3-Dichloropropane

Concen: 54.956 ug/l

RT: 8.87 min Scan# 836

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 76 Resp: 165560

Ion Ratio Lower Upper

76 100

78 31.8 26.5 39.7



#58

2-Chloroethyl Vinyl ether

Concen: 274.445 ug/l

RT: 7.31 min Scan# 678

Delta R.T. 0.02 min

Lab File: VF061140.D




Acq: 22 Dec 2018 12:49


Tgt Ion: 63 Resp: 95003

Ion Ratio Lower Upper

63 100

106 26.9 19.8 29.8








#59
2-Hexanone
Concen: 257.859 ug/l
RT: 9.51 min Scan# 901
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 43 Resp: 241029
Ion Ratio Lower Upper
43 100
58 51.0 25.1 75.2



#60
Dibromochloromethane
Concen: 53.810 ug/l
RT: 8.74 min Scan# 823
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 129 Resp: 128483
Ion Ratio Lower Upper
129 100
127 74.0 36.8 110.3





#61

1,2-Dibromoethane

Concen: 55.327 ug/l

RT: 9.01 min Scan# 850

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:107 Resp: 98423

Ion Ratio Lower Upper




107 100

109 94.7 73.7 110.5



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 49.419 ug/l

RT: 11.43 min Scan# 1095

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion: 95 Resp: 167240

Ion Ratio Lower Upper

95 100

174 62.0 0.0 143.8



176 60.9 0.0 147.6




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.80 min Scan# 930
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
117	100		
82	59.1	45.0	67.4
119	32.9	24.8	37.2






Abundance Ion 117.00 (116.70 to 117.70): V
Ion 82.00 (81.70 to 82.70): VF0
Ion 119.00 (118.70 to 119.70): V




#64
Tetrachloroethene
Concen: 62.020 ug/l
RT: 8.18 min Scan# 766
Delta R.T. 0.03 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion	Ratio	Lower	Upper
164	100		
166	126.7	104.0	156.0
129	91.9	81.1	121.7
131	92.1	81.2	121.8

Abundance Ion 163.85 (163.55 to 164.55): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.90 (128.60 to 129.60): V
Ion 130.90 (130.60 to 131.60): V





#65

Chlorobenzene

Concen: 54.347 ug/l

RT: 9.82 min Scan# 932

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:112 Resp: 334489

Ion Ratio Lower Upper

112 100

114 34.0 27.8 41.6



Abundance Ion 112.00 (111.70 to 112.70): V

Ion 114.00 (113.70 to 114.70): V

150000



100000

50000

0

9.70 9.80 9.90 10.00

Time-->



#66

1,1,1,2-Tetrachloroethane

Concen: 53.252 ug/l

RT: 9.96 min Scan# 946

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion:131 Resp: 133937

Ion Ratio Lower Upper

131 100

133 92.4 49.5 148.7

119 67.1 34.5 103.5



Abundance Ion 131.00 (130.70 to 131.70): V

Ion 133.00 (132.70 to 133.70): V

Ion 119.00 (118.70 to 119.70): V

80000

60000


40000


20000

0

9.80 9.90 10.00 10.10

Time-->









#67
Ethyl Benzene
Concen: 52.421 ug/l
RT: 9.93 min Scan# 943
Delta R.T. 0.03 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 587775
Ion Ratio Lower Upper
91 100
106 29.0 22.5 33.7



#68
m/p-Xylenes
Concen: 101.057 ug/l
RT: 10.16 min Scan# 966
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 106 Resp: 398770
Ion Ratio Lower Upper
106 100
91 206.0 168.2 252.2









#69
o-Xylene
Concen: 53.352 ug/l
RT: 10.74 min Scan# 1025
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:106 Resp: 233489
Ion Ratio Lower Upper
106 100
91 220.9 115.9 347.7



#70
Styrene
Concen: 56.662 ug/l
RT: 10.82 min Scan# 1033
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion:104 Resp: 298696
Ion Ratio Lower Upper
104 100
78 47.4 38.4 57.6
103 46.0 37.0 55.6









#71
Bromoform
Concen: 54.161 ug/l
RT: 10.80 min Scan# 1031
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:173 Resp: 63023
Ion Ratio Lower Upper
173 100
175 49.7 24.8 74.4
254 8.2 0.0 0.0#




Abundance Ion 173.00 (172.70 to 173.70): V
Ion 175.00 (174.70 to 175.70): V
Ion 254.00 (253.70 to 254.70): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.57 min Scan# 1211
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion:152 Resp: 136222
Ion Ratio Lower Upper
152 100
115 57.3 33.3 99.9
150 159.8 0.0 328.2



Abundance Ion 152.00 (151.70 to 152.70): V
Ion 115.00 (114.70 to 115.70): V
Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 56.829 ug/l

RT: 11.14 min Scan# 1066

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion:105 Resp: 667117

Ion Ratio Lower Upper

105 100

120 24.9 13.0 39.0



#74

N-ethyl acetate

Concen: 59.640 ug/l

RT: 11.39 min Scan# 1091

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Tgt Ion: 43 Resp: 155437




Ion Ratio Lower Upper

43 100

70 34.5 29.2 43.8

55 18.9 16.5 24.7

61 26.0 20.7 31.1





#75

1,1,2,2-Tetrachloroethane

Concen: 60.611 ug/l

RT: 11.71 min Scan# 1124

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 83 Resp: 118087

Ion Ratio Lower Upper

83 100

131 8.5 4.8 14.3

85 68.4 33.9 101.6



#76

1,2,3-Trichloropropane

Concen: 55.489 ug/l

RT: 11.81 min Scan# 1134

Delta R.T. 0.01 min

Lab File: VF061140.D




Acq: 22 Dec 2018 12:49


Tgt Ion: 75 Resp: 80454

Ion Ratio Lower Upper

75 100

77 35.3 17.6 52.9





#77

Bromobenzene

Concen: 54.192 ug/l

RT: 11.52 min Scan# 1104

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:156 Resp: 136944

Ion Ratio Lower Upper

156 100

77 159.6 93.8 281.4




158 101.0 48.9 146.8



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 57.963 ug/l

RT: 11.61 min Scan# 1114

Delta R.T. 0.02 min

Lab File: VF061140.D


Acq: 22 Dec 2018 12:49

Tgt Ion: 91 Resp: 757525

Ion Ratio Lower Upper



91 100


120 21.1 10.4 31.2



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 55.914 ug/l

RT: 11.73 min Scan# 1126

Delta R.T. 0.01 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 91 Resp: 450269

Ion Ratio Lower Upper

91 100

126 29.9 16.4 49.1



#80

1,3,5-Trimethylbenzene

Concen: 53.348 ug/l

RT: 11.84 min Scan# 1137

Delta R.T. 0.01 min

Lab File: VF061140.D




Acq: 22 Dec 2018 12:49


Tgt Ion: 105 Resp: 515667

Ion Ratio Lower Upper

105 100

120 50.2 25.4 76.4





#81

trans-1,4-Dichloro-2-butene

Concen: 71.468 ug/l

RT: 11.89 min Scan# 1142

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050




Tot Ion: 75 Resp: 44527

Ion Ratio Lower Upper

75 100

53 83.5 74.1 111.1

89 55.0 42.6 64.0



#82

4-Chlorotoluene

Concen: 51.948 ug/l

RT: 11.92 min Scan# 1145

Delta R.T. 0.02 min

Lab File: VF061140.D





Acq: 22 Dec 2018 12:49


Tgt Ion: 91 Resp: 432757

Ion Ratio Lower Upper

91 100

126 33.0 16.0 47.9









#83
tert-Butylbenzene
Concen: 53.747 ug/l
RT: 12.15 min Scan# 1168
Delta R.T. 0.01 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:119 Resp: 534602
Ion Ratio Lower Upper
119 100
91 63.1 30.5 91.5
134 25.3 12.4 37.4




Abundance Ion 119.00 (118.70 to 119.70): V
Ion 91.00 (90.70 to 91.70): VF0
Ion 134.00 (133.70 to 134.70): V





#84
1,2,4-Trimethylbenzene
Concen: 53.532 ug/l
RT: 12.22 min Scan# 1175
Delta R.T. 0.01 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion:105 Resp: 516003
Ion Ratio Lower Upper
105 100
120 45.2 23.6 71.0



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 55.212 ug/l

RT: 12.33 min Scan# 1186

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:105 Resp: 716682

Ion Ratio Lower Upper




105 100

134 19.9 9.4 28.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 51.340 ug/l

RT: 12.47 min Scan# 1201

Delta R.T. 0.02 min

Lab File: VF061140.D

Acq: 22 Dec 2018 12:49


Tgt Ion:119 Resp: 553162

Ion Ratio Lower Upper

119 100

134 24.3 12.8 38.4



91 25.1 12.2 36.6




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 58.581 ug/l
 RT: 12.49 min Scan# 1203
 Delta R.T. 0.01 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion:146 Resp: 237778
 Ion Ratio Lower Upper
 146 100
 111 39.8 19.3 57.9
 148 64.4 32.4 97.2



#88
 1,4-Dichlorobenzene
 Concen: 53.926 ug/l
 RT: 12.58 min Scan# 1212
 Delta R.T. 0.01 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion:146 Resp: 250228
 Ion Ratio Lower Upper
 146 100
 111 40.7 19.4 58.4
 148 64.7 32.4 97.0









#89
n-Butylbenzene
Concen: 56.830 ug/l
RT: 12.86 min Scan# 1240
Delta R.T. 0.01 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 598490
Ion Ratio Lower Upper
91 100
92 64.2 32.8 98.4
134 22.8 13.2 39.5



#90
Hexachloroethane
Concen: 57.947 ug/l
RT: 12.94 min Scan# 1248
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion: 117 Resp: 154253
Ion Ratio Lower Upper
117 100
201 62.2 24.9 74.8









#91
 1,2-Dichlorobenzene
 Concen: 54.315 ug/l
 RT: 12.96 min Scan# 1250
 Delta R.T. 0.01 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
146	100		
111	46.0	23.2	69.5
148	62.4	32.4	97.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 63.709 ug/l
 RT: 13.66 min Scan# 1321
 Delta R.T. 0.01 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion	Ratio	Lower	Upper
75	100		
155	74.5	41.5	124.5
157	97.2	53.2	159.6









#93
 1,2,4-Trichlorobenzene
 Concen: 52.459 ug/l
 RT: 14.22 min Scan# 1378
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Tot Ion:180 Resp: 168070
 Ion Ratio Lower Upper
 180 100
 182 96.6 46.9 140.6
 145 35.2 20.5 61.6




Abundance Ion 180.00 (179.70 to 180.70): V
 Ion 182.00 (181.70 to 182.70): V
 Ion 145.00 (144.70 to 145.70): V






#94
 Hexachlorobutadiene
 Concen: 53.575 ug/l
 RT: 14.21 min Scan# 1377
 Delta R.T. 0.02 min
 Lab File: VF061140.D
 Acq: 22 Dec 2018 12:49

Tgt Ion:225 Resp: 109324
 Ion Ratio Lower Upper
 225 100
 223 66.5 32.8 98.3
 227 69.2 34.0 101.8



Abundance Ion 225.00 (224.70 to 225.70): V
 Ion 223.00 (222.70 to 223.70): V
 Ion 227.00 (226.70 to 227.70): V









#95
Naphthalene
Concen: 59.537 ug/l
RT: 14.48 min Scan# 1404
Delta R.T. 0.02 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:128 Resp: 348315
Ion Ratio Lower Upper
128 100
127 11.9 9.4 14.2
129 10.7 8.8 13.2




Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V



#96
1,2,3-Trichlorobenzene
Concen: 54.190 ug/l
RT: 14.61 min Scan# 1418
Delta R.T. 0.01 min
Lab File: VF061140.D
Acq: 22 Dec 2018 12:49

Tgt Ion:180 Resp: 160503
Ion Ratio Lower Upper
180 100
182 101.3 47.9 143.8
145 38.0 18.6 55.8



Abundance Ion 180.00 (179.70 to 180.70): V
Ion 182.00 (181.70 to 182.70): V
Ion 145.00 (144.70 to 145.70): V

