

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF122218\
 Data File : VF061141.D
 Acq On : 22 Dec 2018 13:29
 Operator : VA/AP
 Sample : VF1222SBL01
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1222SBL01

Quant Time: Dec 24 03:14:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.90	168	230250	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	5.62	114	399662	50.00	ug/l	0.03
63) Chlorobenzene-d5	9.80	117	372532	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	12.57	152	204208	50.00	ug/l	0.03

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.87	65	147444	49.79	ug/l	0.03
Spiked Amount	50.000		Recovery	=	99.58%	
35) Dibromofluoromethane	4.14	113	167826	52.42	ug/l	0.03
Spiked Amount	50.000		Recovery	=	104.84%	
50) Toluene-d8	7.58	98	438270	48.31	ug/l	0.03
Spiked Amount	50.000		Recovery	=	96.62%	
62) 4-Bromofluorobenzene	11.43	95	209370	50.06	ug/l	0.02
Spiked Amount	50.000		Recovery	=	100.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF122218\
Data File : VF061141.D
Acq On : 22 Dec 2018 13:29
Operator : VA/AP
Sample : VF1222SBL01
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF1222SBL01

Quant Time: Dec 24 03:14:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 00:56:59 2018
Response via : Initial Calibration

