

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF122218\
 Data File : VF061145.D
 Acq On : 22 Dec 2018 15:31
 Operator : VA/AP
 Sample : J6518-01RE
 Misc : 5.25g/5mL/MSVOA-F/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 337-SOILRE

Manual Integrations
 APPROVED

MMDadoda
 12/26/2018 11:05:17 AM

Quant Time: Dec 24 03:29:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	160350	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	5.62	114	309328	50.00	ug/l	0.03
63) Chlorobenzene-d5	9.79	117	221871	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.56	152	73892	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	130963	63.50	ug/l	0.02
Spiked Amount 50.000			Recovery	=	127.00%	
35) Dibromofluoromethane	4.13	113	144599	58.36	ug/l	0.02
Spiked Amount 50.000			Recovery	=	116.72%	
50) Toluene-d8	7.57	98	333547	47.50	ug/l	0.02
Spiked Amount 50.000			Recovery	=	95.00%	
62) 4-Bromofluorobenzene	11.43	95	113630	35.10	ug/l	0.02
Spiked Amount 50.000			Recovery	=	70.20%	
Target Compounds						
16) Acetone	2.25	43	10512m	25.497	ug/l	Qvalue
40) Benzene	4.65	78	9138	1.138	ug/l	94
52) Toluene	7.64	92	38616	7.265	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF122218\
Data File : VF061145.D
Acq On : 22 Dec 2018 15:31
Operator : VA/AP
Sample : J6518-01RE
Misc : 5.25g/5mL/MSVOA-F/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
337-SOILRE

Manual Integrations
APPROVED

MMDadoda
12/26/2018 11:05:17 AM

Quant Time: Dec 24 03:29:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 00:56:59 2018
Response via : Initial Calibration

