

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF122818\
 Data File : VF061228.D
 Acq On : 28 Dec 2018 11:29
 Operator : VA/AP
 Sample : J6498-01RE
 Misc : 6.22g/5mL/MSVOA-F/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SSSI-1218-S-DH-9-TRE

Quant Time: Jan 02 12:56:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 00:56:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	93162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	139142	50.00	ug/l	0.01
63) Chlorobenzene-d5	9.79	117	96076	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.56	152	27126	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	46751	39.02	ug/l	0.01
Spiked Amount 50.000			Recovery	=	78.04%	
35) Dibromofluoromethane	4.13	113	57027	51.16	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.32%	
50) Toluene-d8	7.56	98	132492	41.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.90%	
62) 4-Bromofluorobenzene	11.42	95	39962	27.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.88%	
Target Compounds						
16) Acetone	2.24	43	58308	243.419	ug/l #	89
25) 2-Butanone	4.35	43	16832	43.327	ug/l	98
52) Toluene	7.64	92	3906	1.634	ug/l	96
95) Naphthalene	14.47	128	14487	12.435	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF122818\
Data File : VF061228.D
Acq On : 28 Dec 2018 11:29
Operator : VA/AP
Sample : J6498-01RE
Misc : 6.22g/5mL/MSVOA-F/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SSSI-1218-S-DH-9-TRE

Quant Time: Jan 02 12:56:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F121918S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 00:56:59 2018
Response via : Initial Calibration

