

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010919\
 Data File : VF061305.D
 Acq On : 9 Jan 2019 17:08
 Operator : VA/AP
 Sample : K1082-28
 Misc : 4.36g/5mL/MSVOA-F/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SP-1-2

Quant Time: Jan 10 05:56:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 07:13:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-4.87
34) 1,4-Difluorobenzene	5.58	114	63	50.00	ug/l	-0.03
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-9.79
72) 1,4-Dichlorobenzene-d4	12.46	152	72	50.00	ug/l	-0.09
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
62) 4-Bromofluorobenzene	0.00	95	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
Target Compounds						
					Qvalue	
37) Ethyl Acetate	4.09	43	493	1840.173	ug/l #	1
39) Methylcyclohexane	5.44	83	79	85.820	ug/l #	15
40) Benzene	4.73	78	60	34.933	ug/l #	51
41) Methacrylonitrile	4.71	41	259	1731.164	ug/l #	24
43) Isopropyl Acetate	6.92	43	108	305.340	ug/l #	45
47) Bromodichloromethane	6.22	83	82	122.946	ug/l #	25
48) Methyl methacrylate	6.69	41	138	601.081	ug/l #	21
51) 4-Methyl-2-Pentanone	8.24	43	131	490.682	ug/l #	38
53) t-1,3-Dichloropropene	8.21	75	63	120.280	ug/l #	44
56) Ethyl methacrylate	8.65	69	65	188.260	ug/l #	20
59) 2-Hexanone	9.46	43	169	893.524	ug/l #	27
74) N-amyl acetate	11.40	43	311	215.942	ug/l #	48
76) 1,2,3-Trichloropropane	11.77	75	104	141.156	ug/l #	40
78) n-propylbenzene	11.67	91	86	11.385	ug/l #	55
79) 2-Chlorotoluene	11.67	91	86	20.434	ug/l #	41
81) trans-1,4-Dichloro-2-buten	11.77	75	104	323.014	ug/l #	17
82) 4-Chlorotoluene	11.86	91	60	14.030	ug/l #	42
86) p-Isopropyltoluene	12.41	119	67	11.518	ug/l #	50
89) n-Butylbenzene	12.79	91	61	10.206	ug/l #	27
92) 1,2-Dibromo-3-Chloropropan	13.79	75	86	506.683	ug/l #	1
96) 1,2,3-Trichlorobenzene	14.48	180	85	55.812	ug/l #	13

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF010919\
Data File : VF061305.D
Acq On : 9 Jan 2019 17:08
Operator : VA/AP
Sample : K1082-28
Misc : 4.36g/5mL/MSVOA-F/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SP-1-2

Quant Time: Jan 10 05:56:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 07:13:51 2019
Response via : Initial Calibration

