


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010919\
Data File : VF061292.D
Acq On : 9 Jan 2019 9:50
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Title : SW846 8260
Last Update : Wed Jan 09 07:13:51 2019



AutoFind: Scans 1093, 1094, 1095; Background Corrected with Scan 1087

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	28122	PASS
75	95	30	60	45.7	66266	PASS
95	95	100	100	100.0	144928	PASS
96	95	5	9	6.6	9618	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	69.4	100600	PASS
175	174	5	9	7.6	7619	PASS
176	174	95	101	96.9	97437	PASS
177	176	5	9	6.8	6656	PASS