


Data Path : Z:\voasrv\HPCHEM1\MSVOA_F\Data\VF031219\
 Data File : VF061903.D
 Acq On : 12 Mar 2019 12:33
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
 Title : SW846 8260
 Last Update : Tue Mar 05 04:50:44 2019



AutoFind: Scans 1091, 1092, 1093; Background Corrected with Scan 1086

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	19548	PASS
75	95	30	60	40.9	49085	PASS
95	95	100	100	100.0	119986	PASS
96	95	5	9	7.0	8401	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	83.6	100298	PASS
175	174	5	9	7.6	7576	PASS
176	174	95	101	96.3	96632	PASS
177	176	5	9	6.5	6253	PASS