

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF082318\
 Data File : VF060076.D
 Acq On : 24 Aug 2018 3:42
 Operator : VA/AP
 Sample : J4585-01RE
 Misc : 4.45g/5mL/MSVOA F/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 WS-GRAVEL-STAINRE

Quant Time: Aug 24 04:23:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	153428	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	241757	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.92	117	200246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	63346	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.02	65	147236	70.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	140.24%	
35) Dibromofluoromethane	4.30	113	136787	72.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	144.10%	
50) Toluene-d8	7.72	98	292253	54.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.42%	
62) 4-Bromofluorobenzene	11.51	95	95651	33.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	2.27	84	7740	2.22	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082318\
Data File : VF060076.D
Acq On : 24 Aug 2018 3:42
Operator : VA/AP
Sample : J4585-01RE
Misc : 4.45g/5mL/MSVOA F/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
WS-GRAVEL-STAINRE

Quant Time: Aug 24 04:23:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

