

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF082318\
 Data File : VF060087.D
 Acq On : 24 Aug 2018 8:43
 Operator : VA/AP
 Sample : J4581-17
 Misc : 5.7a/5mL/MSVOA F/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP1-Q3-D1

Quant Time: Aug 24 09:12:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	151300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	260979	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.92	117	269506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	139032	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.01	65	118108	57.04	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	114.08%	
35) Dibromofluoromethane	4.28	113	112367	54.83	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	109.66%	
50) Toluene-d8	7.71	98	260244	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
62) 4-Bromofluorobenzene	11.51	95	114773	37.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.56%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
64) Tetrachloroethene	8.30	164	33394	18.39	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082318\
Data File : VF060087.D
Acq On : 24 Aug 2018 8:43
Operator : VA/AP
Sample : J4581-17
Misc : 5.7a/5mL/MSVOA F/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP1-Q3-D1

Quant Time: Aug 24 09:12:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

