

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF082318\
 Data File : VF060091.D
 Acq On : 24 Aug 2018 10:32
 Operator : VA/AP
 Sample : J4594-01
 Misc : 5.03g/5mL/MSVOA F/SOIL
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 1-3-RO

Quant Time: Aug 24 16:38:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.05	168	123270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.77	114	202434	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.92	117	155458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	46706	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.02	65	110935	65.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.50%	
35) Dibromofluoromethane	4.29	113	73614	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
50) Toluene-d8	7.71	98	196699	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
62) 4-Bromofluorobenzene	11.51	95	63195	26.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.34	43	23380	52.73	ug/l #	91
17) Carbon Disulfide	1.84	76	6391	2.60	ug/l	95
20) Methylene Chloride	2.28	84	10298	7.43	ug/l	91
80) 1,3,5-Trimethylbenzene	11.91	105	11000	4.49	ug/l	95
84) 1,2,4-Trimethylbenzene	12.28	105	12258	4.71	ug/l	82
86) p-Isopropyltoluene	12.51	119	7369	2.67	ug/l	83
95) Naphthalene	14.52	128	5190	5.14	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082318\
Data File : VF060091.D
Acq On : 24 Aug 2018 10:32
Operator : VA/AP
Sample : J4594-01
Misc : 5.03a/5mL/MSVOA F/SOIL
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
1-3-RO

Quant Time: Aug 24 16:38:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

