

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082818\
 Data File : VF060145.D
 Acq On : 28 Aug 2018 12:56
 Operator : VA/AP
 Sample : VF0828SBL01
 Misc : 5.00g/5mL/MSVOA F/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0828SBL01

Quant Time: Aug 28 15:06:03 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	161302	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.75	114	226107	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.91	117	259005	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.63	152	167304	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	107509	48.70	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	97.40%	
35) Dibromofluoromethane	4.29	113	108841	61.30	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	122.60%	
50) Toluene-d8	7.71	98	271366	53.82	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	107.64%	
62) 4-Bromofluorobenzene	11.51	95	145665	54.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082818\
Data File : VF060145.D
Acq On : 28 Aug 2018 12:56
Operator : VA/AP
Sample : VF0828SBL01
Misc : 5.00µg/5mL/MSVOA F/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF0828SBL01

Quant Time: Aug 28 15:06:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

