

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF092618\
 Data File : VF060360.D
 Acq On : 26 Sep 2018 19:07
 Operator : VA/AP
 Sample : J5077-01
 Misc : 5.73/5mL/MSVOA F/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 PARCEL-186-092418-1

Quant Time: Sep 27 03:45:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F091918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 20 02:58:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	328630	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	537878	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	435544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	191702	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.01	65	239800	54.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.38%	
35) Dibromofluoromethane	4.29	113	267174	55.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.44%	
50) Toluene-d8	7.70	98	580557	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
62) 4-Bromofluorobenzene	11.50	95	219666	40.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.40%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF092618\
Data File : VF060360.D
Acq On : 26 Sep 2018 19:07
Operator : VA/AP
Sample : J5077-01
Misc : 5.73/5mL/MSVOA F/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
PARCEL-186-092418-1

Quant Time: Sep 27 03:45:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F091918S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 20 02:58:42 2018
Response via : Initial Calibration

