

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF021319\
 Data File : VF061626.D
 Acq On : 13 Feb 2019 17:05
 Operator : VA/AP
 Sample : VSTDIC020
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020

Quant Time: Feb 14 03:14:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:03:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	296182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	415670	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	370111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	210504	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	41228	11.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	22.38%	
35) Dibromofluoromethane	4.12	113	60301	18.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	36.88%	
50) Toluene-d8	7.56	98	171286	19.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	38.32%	
62) 4-Bromofluorobenzene	11.42	95	71810	18.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	36.34%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	0.97	85	75434	14.515	ug/l 99
3) Chloromethane	1.11	50	74494	18.303	ug/l 96
4) Vinyl Chloride	1.15	62	69526	19.653	ug/l # 88
5) Bromomethane	1.33	94	46026	17.504	ug/l 93
6) Chloroethane	1.39	64	34760	21.097	ug/l 93
7) Trichlorofluoromethane	1.49	101	61137	9.843	ug/l 99
8) Diethyl Ether	1.64	74	20458	21.037	ug/l 99
9) 1,1,2-Trichlorotrifluoroet	1.84	101	58120	17.332	ug/l 96
10) Methyl Iodide	1.91	142	87631	15.840	ug/l 99
11) Tert butyl alcohol	2.58	59	12685	81.232	ug/l 98
12) 1,1-Dichloroethene	1.80	96	51737	18.884	ug/l 95
13) Acrolein	2.00	56	16357	94.511	ug/l 99
14) Allyl chloride	2.10	41	65850	16.537	ug/l 95
15) Acrylonitrile	2.95	53	45008	104.776	ug/l 99
16) Acetone	2.25	43	46764	84.393	ug/l 95
17) Carbon Disulfide	1.82	76	161866	19.387	ug/l 99
18) Methyl Acetate	2.35	43	20944	17.688	ug/l 95
19) Methyl tert-butyl Ether	2.44	73	90522	16.228	ug/l 98
20) Methylene Chloride	2.19	84	21941	6.742	ug/l 96
21) trans-1,2-Dichloroethene	2.32	96	53805	19.012	ug/l 96
22) Diisopropyl ether	2.81	45	146818	16.104	ug/l 99
23) Vinyl Acetate	3.20	43	367574	82.263	ug/l 98
24) 1,1-Dichloroethane	2.88	63	93014	16.890	ug/l 97
25) 2-Butanone	4.35	43	101408	89.878	ug/l 99
26) 2,2-Dichloropropane	3.61	77	38841	14.475	ug/l 99
27) cis-1,2-Dichloroethene	3.48	96	69859	17.373	ug/l 95
28) Bromochloromethane	3.73	49	42158	15.399	ug/l 98
29) Tetrahydrofuran	4.08	42	46934	88.804	ug/l 96
30) Chloroform	3.88	83	102205	13.704	ug/l 91
31) Cyclohexane	3.70	56	67707	12.384	ug/l 97
32) 1,1,1-Trichloroethane	4.10	97	61801	12.281	ug/l 94
36) 1,1-Dichloropropene	4.29	75	85155	18.344	ug/l 99
37) Ethyl Acetate	4.12	43	40957	20.068	ug/l 95
38) Carbon Tetrachloride	4.00	117	58645	15.447	ug/l 92

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF021319\
 Data File : VF061626.D
 Acq On : 13 Feb 2019 17:05
 Operator : VA/AP
 Sample : VSTDIC020
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020

Quant Time: Feb 14 03:14:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:03:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.46	83	102255	21.690	ug/l	99
40) Benzene	4.65	78	234211	22.514	ug/l	98
41) Methacrylonitrile	4.75	41	22502	22.000	ug/l	91
42) 1,2-Dichloroethane	4.95	62	56199	15.614	ug/l	99
43) Isopropyl Acetate	6.97	43	48335	20.862	ug/l #	92
44) Trichloroethene	5.51	130	72082	22.686	ug/l	91
45) 1,2-Dichloropropane	6.24	63	54028	20.133	ug/l	97
46) Dibromomethane	6.10	93	33041	18.849	ug/l	93
47) Bromodichloromethane	6.40	83	70038	15.836	ug/l	96
48) Methyl methacrylate	6.73	41	27244	17.601	ug/l	94
49) 1,4-Dioxane	6.73	88	5032	463.116	ug/l #	81
51) 4-Methyl-2-Pentanone	8.26	43	179765	103.155	ug/l	98
52) Toluene	7.63	92	138360	21.301	ug/l	98
53) t-1,3-Dichloropropene	8.28	75	62185	18.169	ug/l	98
54) cis-1,3-Dichloropropene	7.31	75	80543	18.568	ug/l	99
55) 1,1,2-Trichloroethane	8.50	97	39869	19.841	ug/l	96
56) Ethyl methacrylate	8.63	69	44205	19.843	ug/l	96
57) 1,3-Dichloropropane	8.87	76	68121	19.474	ug/l	97
58) 2-Chloroethyl Vinyl ether	7.31	63	37912	100.886	ug/l	98
59) 2-Hexanone	9.50	43	127391	107.924	ug/l	98
60) Dibromochloromethane	8.73	129	55476	19.871	ug/l	99
61) 1,2-Dibromoethane	8.99	107	43347	20.683	ug/l	95
64) Tetrachloroethene	8.17	164	60212	21.501	ug/l	99
65) Chlorobenzene	9.81	112	155664	21.624	ug/l	98
66) 1,1,1,2-Tetrachloroethane	9.94	131	60779	20.471	ug/l	98
67) Ethyl Benzene	9.91	91	257083	19.616	ug/l	99
68) m/p-Xylenes	10.14	106	192413	42.317	ug/l	97
69) o-Xylene	10.72	106	113642	22.663	ug/l	96
70) Styrene	10.80	104	150506	21.675	ug/l	100
71) Bromoform	10.78	173	33342	23.727	ug/l	96
73) Isopropylbenzene	11.13	105	305972	18.805	ug/l	100
74) N-amyl acetate	11.38	43	76654	18.710	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.71	83	54125	17.821	ug/l	95
76) 1,2,3-Trichloropropane	11.80	75	37434	17.348	ug/l	94
77) Bromobenzene	11.51	156	76647	21.411	ug/l	91
78) n-propylbenzene	11.60	91	352333	18.027	ug/l	93
79) 2-Chlorotoluene	11.73	91	193058	17.347	ug/l	95
80) 1,3,5-Trimethylbenzene	11.83	105	249477	18.586	ug/l	96
81) trans-1,4-Dichloro-2-buten	11.87	75	22033	24.427	ug/l	92
82) 4-Chlorotoluene	11.90	91	204962	17.728	ug/l	94
83) tert-Butylbenzene	12.14	119	272012	19.710	ug/l	98
84) 1,2,4-Trimethylbenzene	12.21	105	235528	17.874	ug/l	97
85) sec-Butylbenzene	12.32	105	352852	19.131	ug/l	99
86) p-Isopropyltoluene	12.47	119	302155	20.029	ug/l	98
87) 1,3-Dichlorobenzene	12.49	146	149624	21.505	ug/l	98
88) 1,4-Dichlorobenzene	12.57	146	139098	21.015	ug/l	99
89) n-Butylbenzene	12.85	91	282604	18.497	ug/l	97
90) Hexachloroethane	12.92	117	67884	17.742	ug/l	83
91) 1,2-Dichlorobenzene	12.95	146	145459	22.446	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.65	75	8193	15.559	ug/l	74

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF021319\
Data File : VF061626.D
Acq On : 13 Feb 2019 17:05
Operator : VA/AP
Sample : VSTDICC020
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VSTDICC020

Quant Time: Feb 14 03:14:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 03:03:38 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.21	180	104155	22.889	ug/l	95
94) Hexachlorobutadiene	14.20	225	61679	21.971	ug/l	95
95) Naphthalene	14.47	128	172171	20.014	ug/l	99
96) 1,2,3-Trichlorobenzene	14.61	180	91765	22.451	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF021319\
 Data File : VF061626.D
 Acq On : 13 Feb 2019 17:05
 Operator : VA/AP
 Sample : VSTDIC020
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020

Quant Time: Feb 14 03:14:51 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:03:38 2019
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.88 min Scan# 431

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020

Tgt Ion: 168 Resp: 296182

Ion Ratio Lower Upper

168 100

99 47.3 34.9 52.3



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.88

100000

80000



60000

40000

20000

0

Time--> 4.70 4.80 4.90 5.00 5.10



#2

Dichlorodifluoromethane

Concen: 14.515 ug/l

RT: 0.97 min Scan# 34

Delta R.T. 0.00 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 85 Resp: 75434

Ion Ratio Lower Upper

85 100

87 31.7 16.1 48.2



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.97

25000

20000


15000


10000

5000

0

Time--> 0.90 1.00 1.10 1.20





#3

Chloromethane

Concen: 18.303 ug/l

RT: 1.11 min Scan# 48

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion: 50 Resp: 74494

Ion Ratio Lower Upper




50 100

52 35.0 26.2 39.4



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#4

Vinyl Chloride

Concen: 19.653 ug/l

RT: 1.15 min Scan# 52

Delta R.T. 0.00 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 62 Resp: 69526

Ion Ratio Lower Upper



62 100


64 27.5 27.5 41.3#



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0





#5

Bromomethane

Concen: 17.504 ug/l

RT: 1.33 min Scan# 71

Delta R.T. 0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion: 94 Resp: 46026

Ion Ratio Lower Upper

94 100

96 86.2 74.6 112.0



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0

12000

10000

8000

6000

4000



2000

0

1.33

1.20 1.30 1.40 1.50

Time-->



#6

Chloroethane

Concen: 21.097 ug/l

RT: 1.39 min Scan# 77

Delta R.T. -0.01 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 64 Resp: 34760

Ion Ratio Lower Upper

64 100

66 30.8 27.9 41.9



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0

12000

10000

8000

6000

4000


2000


0

1.39

1.30 1.40 1.50 1.60

Time-->





#7

Trichlorofluoromethane

Concen: 9.843 ug/l

RT: 1.49 min Scan# 87

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion:101 Resp: 61137

Ion Ratio Lower Upper




101 100

103 60.8 49.3 73.9



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 21.037 ug/l

RT: 1.64 min Scan# 102

Delta R.T. 0.00 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 74 Resp: 20458

Ion Ratio Lower Upper



74 100


45 122.8 60.8 182.3



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 17.332 ug/l

RT: 1.84 min Scan# 122

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion:101 Resp: 58120

Ion Ratio Lower Upper

101 100

85 42.4 29.8 44.6




151 71.6 56.3 84.5



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 15.840 ug/l

RT: 1.91 min Scan# 130

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion:142 Resp: 87631

Ion Ratio Lower Upper

142 100

127 45.2 36.6 55.0



141 14.2 11.0 16.6




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11
Tert butyl alcohol
Concen: 81.232 ug/l
RT: 2.58 min Scan# 198
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 59 Resp: 12685
Ion Ratio Lower Upper
59 100
57 16.3 12.5 18.7



#12
1,1-Dichloroethene
Concen: 18.884 ug/l
RT: 1.80 min Scan# 118
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 96 Resp: 51737
Ion Ratio Lower Upper
96 100
61 159.8 122.6 184.0
98 63.9 47.2 70.8





#13

Acrolein

Concen: 94.511 ug/l

RT: 2.00 min Scan# 139

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 56 Resp: 16357

Ion Ratio Lower Upper

56 100

55 67.7 55.0 82.6



#14

Allyl chloride

Concen: 16.537 ug/l

RT: 2.10 min Scan# 149

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05




Tgt Ion: 41 Resp: 65850


Ion Ratio Lower Upper

41 100

39 72.3 61.6 92.4

76 35.3 30.3 45.5





#15

Acrylonitrile

Concen: 104.776 ug/l

RT: 2.95 min Scan# 235

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020





Tgt Ion: 53 Resp: 45008

Ion Ratio Lower Upper

53 100

52 81.3 64.6 97.0

51 36.2 29.8 44.8



#16

Acetone

Concen: 84.393 ug/l

RT: 2.25 min Scan# 164

Delta R.T. 0.00 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 43 Resp: 46764

Ion Ratio Lower Upper

43 100

58 25.7 22.9 34.3









#17
Carbon Disulfide
Concen: 19.387 ug/l
RT: 1.82 min Scan# 120
Delta R.T. -0.02 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 161866
Ion Ratio Lower Upper
76 100
78 9.4 7.9 11.9



#18
Methyl Acetate
Concen: 17.688 ug/l
RT: 2.35 min Scan# 174
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 43 Resp: 20944
Ion Ratio Lower Upper
43 100
74 20.3 14.6 21.8









#19
Methyl tert-butyl Ether
Concen: 16.228 ug/l
RT: 2.44 min Scan# 183
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 90522
Ion Ratio Lower Upper
73 100
57 20.6 15.8 23.8



#20
Methylene Chloride
Concen: 6.742 ug/l
RT: 2.19 min Scan# 158
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 84 Resp: 21941
Ion Ratio Lower Upper
84 100
49 122.3 101.8 152.8
51 34.2 31.1 46.7
86 67.9 54.4 81.6





#21

trans-1,2-Dichloroethene

Concen: 19.012 ug/l

RT: 2.32 min Scan# 171

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion: 96 Resp: 53805

Ion Ratio Lower Upper

96 100

61 131.0 108.2 162.4




98 60.4 52.2 78.2



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 16.104 ug/l

RT: 2.81 min Scan# 221

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Tgt Ion: 45 Resp: 146818


Ion Ratio Lower Upper

45 100

43 38.9 31.4 47.2

87 19.8 16.3 24.5

59 9.6 8.7 13.1






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 82.263 ug/l

RT: 3.20 min Scan# 260

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion: 43 Resp: 367574

Ion Ratio Lower Upper

43 100

86 8.8 7.5 11.3



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0

3.20

150000



100000

50000

0

Time-->

3.00 3.10 3.20 3.30 3.40



#24

1,1-Dichloroethane

Concen: 16.890 ug/l

RT: 2.88 min Scan# 228

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion: 63 Resp: 93014

Ion Ratio Lower Upper

63 100

98 6.5 2.8 8.3

100 4.6 1.8 5.4



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0

2.88

40000

30000


20000


10000

0

Time-->

2.70 2.80 2.90 3.00 3.10





#25

2-Butanone

Concen: 89.878 ug/l

RT: 4.35 min Scan# 377

Delta R.T. 0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 101408

Ion Ratio Lower Upper

43 100

72 22.0 17.1 25.7



#26

2,2-Dichloropropane

Concen: 14.475 ug/l

RT: 3.61 min Scan# 302

Delta R.T. -0.01 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 77 Resp: 38841

Ion Ratio Lower Upper

77 100

97 22.7 11.7 35.0





#27

cis-1,2-Dichloroethene

Concen: 17.373 ug/l

RT: 3.48 min Scan# 289

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020





Tgt Ion: 96 Resp: 69859

Ion Ratio Lower Upper

96 100

61 128.0 0.0 241.6

98 64.1 0.0 130.2



#28

Bromochloromethane

Concen: 15.399 ug/l

RT: 3.73 min Scan# 314

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05




Tgt Ion: 49 Resp: 42158


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.4

130 94.9 77.7 116.5





#29

Tetrahydrofuran

Concen: 88.804 ug/l

RT: 4.08 min Scan# 350

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion: 42 Resp: 46934

Ion Ratio Lower Upper

42 100

72 41.4 35.2 52.8




71 43.4 32.9 49.3



Abundance Ion 42.00 (41.70 to 42.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 13.704 ug/l

RT: 3.88 min Scan# 329

Delta R.T. 0.00 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 83 Resp: 102205

Ion Ratio Lower Upper



83 100


85 70.8 51.1 76.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0









#31
Cyclohexane
Concen: 12.384 ug/l
RT: 3.70 min Scan# 311
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 56 Resp: 67707
Ion Ratio Lower Upper
56 100
69 31.2 23.8 35.8
84 83.4 68.6 103.0



#32
1,1,1-Trichloroethane
Concen: 12.281 ug/l
RT: 4.10 min Scan# 352
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 97 Resp: 61801
Ion Ratio Lower Upper
97 100
99 69.1 52.2 78.2
61 53.8 39.2 58.8









#33
 1,2-Dichloroethane-d4
 Concen: 11.189 ug/l
 RT: 4.85 min Scan# 428
 Delta R.T. -0.01 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020


Tgt Ion: 65 Resp: 41228
 Ion Ratio Lower Upper
 65 100
 67 53.3 0.0 104.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.61 min Scan# 505
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05

Tgt Ion: 114 Resp: 415670
 Ion Ratio Lower Upper
 114 100
 63 17.3 0.0 36.6
 88 16.7 0.0 34.4









#35
 Dibromofluoromethane
 Concen: 18.439 ug/l
 RT: 4.12 min Scan# 354
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020


Tgt Ion: 113 Resp: 60301
 Ion Ratio Lower Upper
 113 100
 111 94.9 78.8 118.2
 192 21.0 13.4 20.0#



#36
 1,1-Dichloropropene
 Concen: 18.344 ug/l
 RT: 4.29 min Scan# 371
 Delta R.T. -0.01 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05

Tgt Ion: 75 Resp: 85155
 Ion Ratio Lower Upper
 75 100
 110 41.3 20.5 61.6
 77 33.2 27.0 40.4









#37
Ethyl Acetate
Concen: 20.068 ug/l
RT: 4.12 min Scan# 354
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 40957
Ion Ratio Lower Upper
43 100
61 61.1 45.7 68.5
70 11.1 8.5 12.7



#38
Carbon Tetrachloride
Concen: 15.447 ug/l
RT: 4.00 min Scan# 342
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 117 Resp: 58645
Ion Ratio Lower Upper
117 100
119 103.9 75.6 113.4
121 32.5 25.2 37.8









#39
Methylcyclohexane
Concen: 21.690 ug/l
RT: 5.46 min Scan# 490
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 102255
Ion Ratio Lower Upper
83 100
55 81.6 66.6 99.8
98 50.1 40.4 60.6



#40
Benzene
Concen: 22.514 ug/l
RT: 4.65 min Scan# 407
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 78 Resp: 234211
Ion Ratio Lower Upper
78 100
77 25.0 19.0 28.6





#41

Methacrylonitrile

Concen: 22.000 ug/l

RT: 4.75 min Scan# 418

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020

Tgt Ion: 41 Resp: 22502


Ion Ratio Lower Upper

41 100

39 46.7 45.3 67.9

67 54.6 49.3 73.9

52 41.4 32.6 49.0






Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

Ion 67.00 (66.70 to 67.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#42

1,2-Dichloroethane

Concen: 15.614 ug/l

RT: 4.95 min Scan# 438

Delta R.T. -0.01 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 62 Resp: 56199

Ion Ratio Lower Upper



62 100


98 12.5 0.0 26.0



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 98.00 (97.70 to 98.70): VF0









#43

Isopropyl Acetate
 Concen: 20.862 ug/l
 RT: 6.97 min Scan# 643
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020




Tgt Ion: 43 Resp: 48335
 Ion Ratio Lower Upper
 43 100
 61 36.6 25.8 38.6
 87 0.0 0.4 0.6#




#44

Trichloroethene
 Concen: 22.686 ug/l
 RT: 5.51 min Scan# 495
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05

Tgt Ion: 130 Resp: 72082
 Ion Ratio Lower Upper
 130 100
 95 81.6 0.0 179.4





#45

1,2-Dichloropropane

Concen: 20.133 ug/l

RT: 6.24 min Scan# 569

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 63 Resp: 54028

Ion Ratio Lower Upper

63 100

65 33.0 24.9 37.3



#46

Dibromomethane

Concen: 18.849 ug/l

RT: 6.10 min Scan# 554

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05




Tgt Ion: 93 Resp: 33041


Ion Ratio Lower Upper

93 100

95 80.8 69.0 103.6

174 115.9 99.3 148.9





#47

Bromodichloromethane

Concen: 15.836 ug/l

RT: 6.40 min Scan# 585

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion: 83 Resp: 70038

Ion Ratio Lower Upper

83 100

85 63.2 53.4 80.2




127 9.3 7.5 11.3



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 17.601 ug/l

RT: 6.73 min Scan# 618

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion: 41 Resp: 27244

Ion Ratio Lower Upper

41 100

69 81.0 60.4 90.6



39 59.0 43.7 65.5




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#49

1,4-Dioxane

Concen: 463.116 ug/l

RT: 6.73 min Scan# 618

Delta R.T. 0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020





Tgt Ion: 88 Resp: 5032

Ion Ratio Lower Upper

88 100

43 35.2 32.3 48.5

58 53.5 59.4 89.2#



#50

Toluene-d8

Concen: 19.163 ug/l

RT: 7.56 min Scan# 703

Delta R.T. 0.00 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 98 Resp: 171286

Ion Ratio Lower Upper

98 100

100 68.0 55.4 83.2





#51

4-Methyl-2-Pentanone

Concen: 103.155 ug/l

RT: 8.26 min Scan# 774

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 179765

Ion Ratio Lower Upper

43 100

58 40.1 33.1 49.7



#52

Toluene

Concen: 21.301 ug/l

RT: 7.63 min Scan# 710

Delta R.T. 0.00 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 92 Resp: 138360

Ion Ratio Lower Upper

92 100

91 170.1 133.9 200.9





#53

t-1,3-Dichloropropene

Concen: 18.169 ug/l

RT: 8.28 min Scan# 776

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 75 Resp: 62185

Ion Ratio Lower Upper

75 100

77 31.4 26.2 39.2



#54

cis-1,3-Dichloropropene

Concen: 18.568 ug/l

RT: 7.31 min Scan# 677

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05




Tgt Ion: 75 Resp: 80543


Ion Ratio Lower Upper

75 100

77 31.7 26.4 39.6

39 44.6 36.0 54.0





#55

1,1,2-Trichloroethane

Concen: 19.841 ug/l

RT: 8.50 min Scan# 798

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 39869


Ion Ratio Lower Upper

97 100

83 85.7 69.2 103.8

85 58.0 43.0 64.4

99 63.7 47.2 70.8






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 19.843 ug/l

RT: 8.63 min Scan# 811

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion: 69 Resp: 44205

Ion Ratio Lower Upper

69 100

41 62.1 52.9 79.3



39 31.2 24.5 36.7




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 19.474 ug/l

RT: 8.87 min Scan# 835

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion: 76 Resp: 68121

Ion Ratio Lower Upper




76 100

78 33.1 27.8 41.6



Abundance Ion 76.00 (75.70 to 76.70): VF0

Ion 78.00 (77.70 to 78.70): VF0



#58

2-Chloroethyl Vinyl ether

Concen: 100.886 ug/l

RT: 7.31 min Scan# 677

Delta R.T. 0.00 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 63 Resp: 37912

Ion Ratio Lower Upper



63 100


106 29.4 24.6 36.8



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 106.00 (105.70 to 106.70): V





#59

2-Hexanone

Concen: 107.924 ug/l

RT: 9.50 min Scan# 899

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 127391

Ion Ratio Lower Upper

43 100

58 58.1 28.5 85.5



#60

Dibromochloromethane

Concen: 19.871 ug/l

RT: 8.73 min Scan# 821

Delta R.T. 0.00 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 129 Resp: 55476

Ion Ratio Lower Upper

129 100

127 74.6 37.8 113.4





#61

1,2-Dibromoethane

Concen: 20.683 ug/l

RT: 8.99 min Scan# 848

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion: 107 Resp: 43347

Ion Ratio Lower Upper




107 100

109 101.6 77.4 116.0



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 18.170 ug/l

RT: 11.42 min Scan# 1094

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion: 95 Resp: 71810

Ion Ratio Lower Upper

95 100

174 90.1 0.0 165.8



176 87.5 0.0 167.0




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.78 min Scan# 928
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion:117 Resp: 370111
Ion Ratio Lower Upper
117 100
82 50.7 39.3 58.9
119 33.1 26.6 40.0



#64
Tetrachloroethene
Concen: 21.501 ug/l
RT: 8.17 min Scan# 764
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion:164 Resp: 60212
Ion Ratio Lower Upper
164 100
166 130.9 103.9 155.9
129 87.3 71.3 106.9
131 85.1 67.1 100.7









#65
Chlorobenzene
Concen: 21.624 ug/l
RT: 9.81 min Scan# 931
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion:112 Resp: 155664
Ion Ratio Lower Upper
112 100
114 31.4 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.471 ug/l
RT: 9.94 min Scan# 944
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion:131 Resp: 60779
Ion Ratio Lower Upper
131 100
133 94.5 48.9 146.6
119 68.4 34.8 104.3





#67

Ethyl Benzene

Concen: 19.616 ug/l

RT: 9.91 min Scan# 941

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 91 Resp: 257083

Ion Ratio Lower Upper

91 100

106 29.5 24.1 36.1



#68

m/p-Xylenes

Concen: 42.317 ug/l

RT: 10.14 min Scan# 964

Delta R.T. -0.01 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 106 Resp: 192413

Ion Ratio Lower Upper

106 100

91 195.2 153.0 229.6









#69
o-Xylene
Concen: 22.663 ug/l
RT: 10.72 min Scan# 1023
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion:106 Resp: 113642
Ion Ratio Lower Upper
106 100
91 205.5 100.1 300.1



#70
Styrene
Concen: 21.675 ug/l
RT: 10.80 min Scan# 1031
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion:104 Resp: 150506
Ion Ratio Lower Upper
104 100
78 42.1 33.2 49.8
103 46.5 37.2 55.8





#71

Bromoform

Concen: 23.727 ug/l

RT: 10.78 min Scan# 1029

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion:173 Resp: 33342

Ion Ratio Lower Upper

173 100

175 45.5 24.0 72.0




254 8.8 8.2 12.2



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.55 min Scan# 1209

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion:152 Resp: 210504

Ion Ratio Lower Upper

152 100

115 54.0 24.3 72.8



150 156.9 0.0 312.0




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 18.805 ug/l

RT: 11.13 min Scan# 1065

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion:105 Resp: 305972

Ion Ratio Lower Upper




105 100

120 27.0 13.4 40.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 18.710 ug/l

RT: 11.38 min Scan# 1090

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Tgt Ion: 43 Resp: 76654


Ion Ratio Lower Upper

43 100

70 38.0 31.5 47.3

55 19.8 16.7 25.1

61 26.6 23.0 34.6






Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0





#75

1,1,2,2-Tetrachloroethane

Concen: 17.821 ug/l

RT: 11.71 min Scan# 1123

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020





Tgt Ion: 83 Resp: 54125

Ion Ratio Lower Upper

83 100

131 10.9 4.4 13.2

85 66.5 31.5 94.5



#76

1,2,3-Trichloropropane

Concen: 17.348 ug/l

RT: 11.80 min Scan# 1133

Delta R.T. 0.00 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 75 Resp: 37434

Ion Ratio Lower Upper

75 100

77 36.7 16.6 49.7





#77

Bromobenzene

Concen: 21.411 ug/l

RT: 11.51 min Scan# 1103

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion:156 Resp: 76647

Ion Ratio Lower Upper

156 100

77 114.6 64.3 192.7

158 95.6 50.9 152.7




Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V


11.51

Time-->



11.51

Time-->



#78

n-propylbenzene

Concen: 18.027 ug/l

RT: 11.60 min Scan# 1112

Delta R.T. -0.01 min

Lab File: VF061626.D


Acq: 13 Feb 2019 17:05

Tgt Ion: 91 Resp: 352333

Ion Ratio Lower Upper

91 100

120 20.7 12.1 36.3




Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V


11.60

Time-->



11.60

Time-->



#79

2-Chlorotoluene

Concen: 17.347 ug/l

RT: 11.73 min Scan# 1125

Delta R.T. -0.01 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC020

Tgt Ion: 91 Resp: 193058

Ion Ratio Lower Upper

91 100

126 36.6 19.8 59.4



#80

1,3,5-Trimethylbenzene

Concen: 18.586 ug/l

RT: 11.83 min Scan# 1136

Delta R.T. 0.00 min

Lab File: VF061626.D




Acq: 13 Feb 2019 17:05


Tgt Ion: 105 Resp: 249477

Ion Ratio Lower Upper

105 100

120 52.8 27.9 83.6









#81
trans-1,4-Dichloro-2-butene
Concen: 24.427 ug/l
RT: 11.87 min Scan# 1140
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 75 Resp: 22033
Ion Ratio Lower Upper
75 100
53 72.8 53.7 80.5
89 45.4 41.6 62.4



#82
4-Chlorotoluene
Concen: 17.728 ug/l
RT: 11.90 min Scan# 1143
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 91 Resp: 204962
Ion Ratio Lower Upper
91 100
126 34.3 19.1 57.1









#83
tert-Butylbenzene
Concen: 19.710 ug/l
RT: 12.14 min Scan# 1167
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020

Tgt Ion:119 Resp: 272012
Ion Ratio Lower Upper
119 100
91 51.5 26.3 78.8
134 24.7 12.9 38.7




Abundance Ion 119.00 (118.70 to 119.70): V
Ion 91.00 (90.70 to 91.70): VF0
Ion 134.00 (133.70 to 134.70): V






#84
1,2,4-Trimethylbenzene
Concen: 17.874 ug/l
RT: 12.21 min Scan# 1174
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion:105 Resp: 235528
Ion Ratio Lower Upper
105 100
120 49.9 26.1 78.3



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 19.131 ug/l

RT: 12.32 min Scan# 1185

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC020

Tgt Ion:105 Resp: 352852

Ion Ratio Lower Upper

105 100

134 21.3 10.8 32.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V

12.32

250000

200000



150000

100000

50000

0

Time--> 12.20 12.30 12.40



#86

p-Isopropyltoluene

Concen: 20.029 ug/l

RT: 12.47 min Scan# 1200

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion:119 Resp: 302155

Ion Ratio Lower Upper

119 100

134 26.5 13.7 41.1

91 20.4 10.8 32.4



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VF0

300000

250000

200000


150000


100000

50000

0

Time--> 12.40 12.50 12.60





#87

1,3-Dichlorobenzene

Concen: 21.505 ug/l

RT: 12.49 min Scan# 1202

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC020


Tgt Ion:146 Resp: 149624

Ion Ratio Lower Upper

146 100

111 33.4 17.5 52.6




148 64.1 31.4 94.3



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#88

1,4-Dichlorobenzene

Concen: 21.015 ug/l

RT: 12.57 min Scan# 1211

Delta R.T. 0.00 min

Lab File: VF061626.D

Acq: 13 Feb 2019 17:05


Tgt Ion:146 Resp: 139098

Ion Ratio Lower Upper

146 100

111 34.9 18.4 55.4



148 64.5 32.2 96.6




Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V









#89
n-Butylbenzene
Concen: 18.497 ug/l
RT: 12.85 min Scan# 1239
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 282604
Ion Ratio Lower Upper
91 100
92 60.6 31.4 94.3
134 26.4 12.3 36.9



#90
Hexachloroethane
Concen: 17.742 ug/l
RT: 12.92 min Scan# 1246
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion: 117 Resp: 67884
Ion Ratio Lower Upper
117 100
201 58.6 36.2 108.6









#91
 1,2-Dichlorobenzene
 Concen: 22.446 ug/l
 RT: 12.95 min Scan# 1249
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020


Tgt Ion: 146 Resp: 145459
 Ion Ratio Lower Upper
 146 100
 111 37.3 20.2 60.5
 148 65.3 32.3 96.9



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 15.559 ug/l
 RT: 13.65 min Scan# 1320
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05

Tgt Ion: 75 Resp: 8193
 Ion Ratio Lower Upper
 75 100
 155 120.1 48.9 146.8
 157 148.6 58.3 174.8









#93
 1,2,4-Trichlorobenzene
 Concen: 22.889 ug/l
 RT: 14.21 min Scan# 1377
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC020


Tgt Ion:180 Resp: 104155
 Ion Ratio Lower Upper
 180 100
 182 92.6 48.6 145.9
 145 27.4 15.3 45.9



#94
 Hexachlorobutadiene
 Concen: 21.971 ug/l
 RT: 14.20 min Scan# 1376
 Delta R.T. 0.00 min
 Lab File: VF061626.D
 Acq: 13 Feb 2019 17:05

Tgt Ion:225 Resp: 61679
 Ion Ratio Lower Upper
 225 100
 223 67.1 30.9 92.7
 227 67.4 32.1 96.3









#95
Naphthalene
Concen: 20.014 ug/l
RT: 14.47 min Scan# 1403
Delta R.T. 0.00 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC020

Tgt Ion:128 Resp: 172171
Ion Ratio Lower Upper
128 100
127 11.8 9.8 14.6
129 11.1 9.4 14.2




Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V



#96
1,2,3-Trichlorobenzene
Concen: 22.451 ug/l
RT: 14.61 min Scan# 1417
Delta R.T. -0.01 min
Lab File: VF061626.D
Acq: 13 Feb 2019 17:05

Tgt Ion:180 Resp: 91765
Ion Ratio Lower Upper
180 100
182 98.1 49.0 147.2
145 33.6 15.4 46.1



Abundance Ion 180.00 (179.70 to 180.70): V
Ion 182.00 (181.70 to 182.70): V
Ion 145.00 (144.70 to 145.70): V

