

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF030119\
 Data File : VF061816.D
 Acq On : 1 Mar 2019 21:42
 Operator : VA/AP
 Sample : K1654-01RE
 Misc : 5.01g/5mL/MSVOA-F/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SOIL-PILERE

Manual Integrations
 APPROVED

MMDadoda
 3/4/2019 4:02:10 PM

Quant Time: Mar 02 03:41:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:27:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	251674	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.59	114	421637	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.76	117	339144	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	172299	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	113293	61.39	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	122.78%	
35) Dibromofluoromethane	4.10	113	163879	52.34	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	104.68%	
50) Toluene-d8	7.54	98	397236	44.13	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	88.26%	
62) 4-Bromofluorobenzene	11.40	95	173459	47.63	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	95.26%	
Target Compounds						
16) Acetone	2.24	43	3416	8.627	ug/l	94
20) Methylene Chloride	2.24	84	24465m	9.488	ug/l	
52) Toluene	7.61	92	7001	1.057	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF030119\
Data File : VF061816.D
Acq On : 1 Mar 2019 21:42
Operator : VA/AP
Sample : K1654-01RE
Misc : 5.01µ/5mL/MSVOA-F/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SOIL-PILERE

Manual Integrations
APPROVED

MMDadoda
3/4/2019 4:02:10 PM

Quant Time: Mar 02 03:41:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 03:27:56 2019
Response via : Initial Calibration

