


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF030119\
 Data File : VF061800.D
 Acq On : 1 Mar 2019 12:24
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Title : SW846 8260
 Last Update : Thu Feb 14 03:27:56 2019



AutoFind: Scans 1091, 1092, 1093; Background Corrected with Scan 1085

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.2	17433	PASS
75	95	30	60	38.9	44485	PASS
95	95	100	100	100.0	114461	PASS
96	95	5	9	6.3	7229	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.5	102418	PASS
175	174	5	9	7.0	7172	PASS
176	174	95	101	97.9	100248	PASS
177	176	5	9	6.9	6905	PASS