


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF061918\
 Data File : VF059205.D
 Acq On : 19 Jun 2018 23:48
 Operator : VA/AP
 Sample : BFB
 Misc : 5.40g/5mL/MSVOA-F/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Title : SW846 8260
 Last Update : Thu Jun 07 10:52:00 2018



AutoFind: Scans 1112, 1113, 1114; Background Corrected with Scan 1107

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	24165	PASS
75	95	30	60	47.0	56863	PASS
95	95	100	100	100.0	120901	PASS
96	95	5	9	7.3	8780	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.2	90861	PASS
175	174	5	9	7.8	7056	PASS
176	174	95	101	98.7	89672	PASS
177	176	5	9	6.5	5870	PASS