

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF062018\
 Data File : VF059253.D
 Acq On : 21 Jun 2018 5:47
 Operator : VA/AP
 Sample : J3570-03
 Misc : 5.17g/5mL/MSVOA-F/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 20180373-AMENDED-COMPOSITE-

Manual Integrations
 APPROVED

apatel
 6/21/2018 9:46:46 AM

Quant Time: Jun 21 08:09:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 10:52:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	172216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	268783	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	235563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	121314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	130883	62.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.32%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	
50) Toluene-d8	7.71	98	259021	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
62) 4-Bromofluorobenzene	11.51	95	144701	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
Target Compounds						
16) Acetone	2.33	43	142046	284.84	ug/l #	91
17) Carbon Disulfide	1.87	76	168573	52.91	ug/l	97
20) Methylene Chloride	2.27	84	8746m	7.32	ug/l	
25) 2-Butanone	4.50	43	27082	35.03	ug/l #	90
95) Naphthalene	14.53	128	12019	2.14	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF062018\
Data File : VF059253.D
Acq On : 21 Jun 2018 5:47
Operator : VA/AP
Sample : J3570-03
Misc : 5.17g/5mL/MSVOA-F/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
20180373-AMENDED-COMPOSITE-

Manual Integrations
APPROVED

apatel
6/21/2018 9:46:46 AM

Quant Time: Jun 21 08:09:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 10:52:00 2018
Response via : Initial Calibration

