

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF062018\
 Data File : VF059229.D
 Acq On : 20 Jun 2018 15:39
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 21 07:10:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 10:52:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.566	0.555	1.9	74	0.00
3 P	Chloromethane	0.438	0.362	17.4	60	0.00
4 C	Vinyl Chloride	0.395	0.353	10.6#	65	0.00
5 T	Bromomethane	0.230	0.225	2.2	74	0.00
6 T	Chloroethane	0.164	0.171	-4.3	73	0.00
7 T	Trichlorofluoromethane	0.615	0.725	-17.9	81	0.00
8 T	Diethyl Ether	0.155	0.182	-17.4	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.392	0.366	6.6	68	0.00
10 T	Methyl Iodide	0.686	0.717	-4.5	75	0.00
11 T	Tert butyl alcohol	0.039	0.047	-20.5	93	0.00
12 CM	1,1-Dichloroethene	0.329	0.345	-4.9#	74	0.00
13 T	Acrolein	0.034	0.033	2.9	84	0.00
14 T	Allyl chloride	0.647	0.660	-2.0	77	0.00
15 T	Acrylonitrile	0.077	0.080	-3.9	70	0.00
16 T	Acetone	0.145	0.203	-40.0#	104	0.00
17 T	Carbon Disulfide	0.925	0.833	9.9	64	0.02
18 T	Methyl Acetate	0.249	0.278	-11.6	88	0.00
19 T	Methyl tert-butyl Ether	1.023	1.253	-22.5	90	0.00
20 T	Methylene Chloride	0.347	0.366	-5.5	81	0.00
21 T	trans-1,2-Dichloroethene	0.361	0.364	-0.8	75	0.00
22 T	Diisopropyl ether	1.271	1.398	-10.0	84	0.01
23 T	Vinyl Acetate	0.838	1.029	-22.8	89	0.01
24 P	1,1-Dichloroethane	0.771	0.857	-11.2	85	0.00
25 T	2-Butanone	0.224	0.255	-13.8	90	0.00
26 T	2,2-Dichloropropane	0.559	0.590	-5.5	78	0.00
27 T	cis-1,2-Dichloroethene	0.556	0.556	0.0	73	0.00
28 T	Bromochloromethane	0.347	0.358	-3.2	82	0.00
29	Tetrahydrofuran	0.099	0.107	-8.1	82	0.00
30 C	Chloroform	1.102	1.213	-10.1#	80	0.00
31 T	Cyclohexane	0.673	0.585	13.1	64	0.00
32 T	1,1,1-Trichloroethane	0.823	0.859	-4.4	76	0.00
33 S	1,2-Dichloroethane-d4	0.606	0.707	-16.7	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.351	0.392	-11.7	75	0.00
36 T	1,1-Dichloropropene	0.485	0.510	-5.2	73	0.00
37 T	Ethyl Acetate	0.299	0.355	-18.7	89	0.00
38 T	Carbon Tetrachloride	0.525	0.613	-16.8	78	0.00
39 T	Methylcyclohexane	0.494	0.477	3.4	65	0.00
40 TM	Benzene	1.125	1.101	2.1	70	0.00
41 T	Methacrylonitrile	0.147	0.171	-16.3	84	0.00
42 TM	1,2-Dichloroethane	0.505	0.632	-25.1#	86	0.00
43 T	Isopropyl Acetate	0.401	0.438	-9.2	83	0.00
44 TM	Trichloroethene	0.351	0.377	-7.4	77	0.00
45 C	1,2-Dichloropropane	0.324	0.313	3.4#	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF062018\
 Data File : VF059229.D
 Acq On : 20 Jun 2018 15:39
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 21 07:10:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 10:52:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.244	0.283	-16.0	79	0.00
47 T	Bromodichloromethane	0.555	0.652	-17.5	81	0.00
48 T	Methyl methacrylate	0.247	0.294	-19.0	84	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	62	0.00
50 S	Toluene-d8	1.031	0.988	4.2	64	0.00
51 T	4-Methyl-2-Pentanone	0.275	0.328	-19.3	90	0.00
52 CM	Toluene	0.810	0.796	1.7#	70	0.00
53 T	t-1,3-Dichloropropene	0.538	0.623	-15.8	88	0.00
54 T	cis-1,3-Dichloropropene	0.618	0.675	-9.2	78	0.00
55 T	1,1,2-Trichloroethane	0.294	0.337	-14.6	80	0.00
56 T	Ethyl methacrylate	0.370	0.421	-13.8	83	0.00
57 T	1,3-Dichloropropane	0.502	0.585	-16.5	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.026	0.027	-3.8	77	0.00
59 T	2-Hexanone	0.217	0.269	-24.0	95	0.00
60 T	Dibromochloromethane	0.422	0.500	-18.5	82	0.00
61 T	1,2-Dibromoethane	0.341	0.406	-19.1	84	0.00
62 S	4-Bromofluorobenzene	0.529	0.562	-6.2	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	73	0.00
64 T	Tetrachloroethene	0.367	0.374	-1.9	74	0.00
65 PM	Chlorobenzene	0.990	1.046	-5.7	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.423	-7.4	80	0.00
67 C	Ethyl Benzene	1.720	1.686	2.0#	71	0.00
68 T	m/p-Xylenes	0.644	0.615	4.5	76	0.00
69 T	o-Xylene	0.696	0.695	0.1	74	0.00
70 T	Styrene	1.053	1.050	0.3	74	0.00
71 P	Bromoform	0.267	0.320	-19.9	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.462	3.390	2.1	75	0.00
74 T	N-amyl acetate	1.237	1.254	-1.4	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.821	0.851	-3.7	84	0.00
76 T	1,2,3-Trichloropropane	0.606	0.639	-5.4	86	0.00
77 T	Bromobenzene	0.871	0.904	-3.8	81	0.00
78 T	n-propylbenzene	4.097	3.935	4.0	74	0.00
79 T	2-Chlorotoluene	2.421	2.407	0.6	78	0.00
80 T	1,3,5-Trimethylbenzene	2.865	2.776	3.1	75	0.00
81 T	trans-1,4-Dichloro-2-butene	0.393	0.402	-2.3	84	0.00
82 T	4-Chlorotoluene	2.577	2.576	0.0	79	0.00
83 T	tert-Butylbenzene	2.845	2.935	-3.2	78	0.00
84 T	1,2,4-Trimethylbenzene	2.969	2.926	1.4	73	0.00
85 T	sec-Butylbenzene	3.866	3.895	-0.8	78	0.00
86 T	p-Isopropyltoluene	3.175	3.220	-1.4	76	0.00
87 T	1,3-Dichlorobenzene	1.608	1.559	3.0	77	0.00
88 T	1,4-Dichlorobenzene	1.550	1.630	-5.2	82	0.00
89 T	n-Butylbenzene	3.325	3.295	0.9	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF062018\
Data File : VF059229.D
Acq On : 20 Jun 2018 15:39
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_F
LabSampleId :
VSTDCCC050

Quant Time: Jun 21 07:10:33 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 10:52:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.782	0.815	-4.2	79	0.00
91 T	1,2-Dichlorobenzene	1.550	1.637	-5.6	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.161	0.195	-21.1	92	0.00
93 T	1,2,4-Trichlorobenzene	1.084	1.082	0.2	77	0.00
94 T	Hexachlorobutadiene	0.627	0.692	-10.4	82	0.00
95 T	Naphthalene	2.315	2.781	-20.1	93	0.00
96 T	1,2,3-Trichlorobenzene	1.002	1.185	-18.3	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6