

Data Path : W:\HPCHEM1\MSVOA H\DATA\VH092115\
 Data File : VH057088.D
 Acq On : 21 Sep 2015 11:47
 Operator : SY/FY
 Sample : VH0921MBL01
 Misc : 5.00g/10mL/100uL/5mL/MSVOA_H/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_H
 ClientSampleId :
 VH0921MBL01

Quant Time: Sep 22 04:54:46 2015
 Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 18 15:34:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.82	168	900911	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.55	114	1610419	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.70	117	1191326	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.49	152	329572	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.79	65	792130	49.24	ug/l	-0.02
Spiked Amount	50.000	Range	66 - 150	Recovery	=	98.48%
35) Dibromofluoromethane	4.06	113	676375	52.75	ug/l	-0.02
Spiked Amount	50.000	Range	76 - 130	Recovery	=	105.50%
48) Toluene-d8	7.50	98	1823659	49.46	ug/l	-0.02
Spiked Amount	50.000	Range	78 - 121	Recovery	=	98.92%
62) 4-Bromofluorobenzene	11.35	95	611627	49.19	ug/l	0.00
Spiked Amount	50.000	Range	70 - 131	Recovery	=	98.38%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA H\DATA\VH092115\
Data File : VH057088.D
Acq On : 21 Sep 2015 11:47
Operator : SY/FY
Sample : VH0921MBL01
Misc : 5.00g/10mL/100uL/5mL/MSVOA_H/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_H
ClientSampleId :
VH0921MBL01

Quant Time: Sep 22 04:54:46 2015
Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 18 15:34:01 2015
Response via : Initial Calibration

