

Data Path : W:\HPCHEM1\MSVOA_H\DATA\VH092215\
 Data File : VH057109.D
 Acq On : 22 Sep 2015 17:54
 Operator : SY/FY
 Sample : G3767-10ME
 Misc : 4.99g/10mL//100uL/5mL/MSVOA_H/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_H
 ClientSampleId :
 ERD45-CS10-3ME

Quant Time: Sep 23 11:38:56 2015
 Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 18 15:34:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.83	168	993006	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.55	114	1715947	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.71	117	1379298	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.49	152	424023	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.80	65	893774	50.40	ug/l	-0.01
Spiked Amount 50.000	Range 66 - 150		Recovery	=	100.80%	
35) Dibromofluoromethane	4.07	113	805418	58.95	ug/l	-0.01
Spiked Amount 50.000	Range 76 - 130		Recovery	=	117.90%	
48) Toluene-d8	7.50	98	2311035	58.83	ug/l	-0.02
Spiked Amount 50.000	Range 78 - 121		Recovery	=	117.66%	
62) 4-Bromofluorobenzene	11.35	95	806072	60.84	ug/l	0.00
Spiked Amount 50.000	Range 70 - 131		Recovery	=	121.68%	

Target Compounds						Qvalue
43) Trichloroethene	5.47	130	49152	4.31	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_H\DATA\VH092215\
Data File : VH057109.D
Acq On : 22 Sep 2015 17:54
Operator : SY/FY
Sample : G3767-10ME
Misc : 4.99g/10mL//100uL/5mL/MSVOA_H/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_H
ClientSampleId :
ERD45-CS10-3ME

Quant Time: Sep 23 11:38:56 2015
Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 18 15:34:01 2015
Response via : Initial Calibration

