

Data Path : W:\HPCHEM1\MSVOA_H\DATA\VH092215\
 Data File : VH057110.D
 Acq On : 22 Sep 2015 18:22
 Operator : SY/FY
 Sample : G3767-13
 Misc : 4.98g/10mL//100uL/5mL/MSVOA_H/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_H
 ClientSampleId :
 ERD45-CS11-3

Quant Time: Sep 23 11:44:36 2015
 Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 18 15:34:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.83	168	879134	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.55	114	1491995	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.71	117	1203388	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.48	152	343213	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.80	65	358332	22.82	ug/l	-0.01
Spiked Amount 50.000	Range 66 - 150		Recovery	=	45.64%#	
35) Dibromofluoromethane	4.08	113	329839	27.77	ug/l	0.00
Spiked Amount 50.000	Range 76 - 130		Recovery	=	55.54%#	
48) Toluene-d8	7.50	98	894667	26.19	ug/l	-0.02
Spiked Amount 50.000	Range 78 - 121		Recovery	=	52.38%#	
62) 4-Bromofluorobenzene	11.35	95	311201	27.02	ug/l	0.00
Spiked Amount 50.000	Range 70 - 131		Recovery	=	54.04%#	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
27) cis-1,2-Dichloroethene	3.43	96	89207	7.16	ug/l	97
43) Trichloroethene	5.47	130	1055140	106.36	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_H\DATA\VH092215\
Data File : VH057110.D
Acq On : 22 Sep 2015 18:22
Operator : SY/FY
Sample : G3767-13
Misc : 4.98g/10mL//100uL/5mL/MSVOA_H/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_H
ClientSampleId :
ERD45-CS11-3

Quant Time: Sep 23 11:44:36 2015
Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 18 15:34:01 2015
Response via : Initial Calibration

