

Data Path : W:\HPCHEM1\MSVOA H\DATA\VH092215\
 Data File : VH057114.D
 Acq On : 22 Sep 2015 20:11
 Operator : SY/FY
 Sample : G3712-06ME
 Misc : 5.01g/10mL//100uL/5mL/MSVOA_H/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_H
 ClientSampleId :
 ERD45-CS4-3ME

Quant Time: Sep 24 12:00:26 2015
 Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 18 15:34:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.84	168	764855	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.56	114	1342913	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.71	117	1046727	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.48	152	306119	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.80	65	259528	19.00	ug/l	0.00
Spiked Amount 50.000	Range 66 - 150		Recovery =	38.00%#		
35) Dibromofluoromethane	4.08	113	218942	20.48	ug/l	0.00
Spiked Amount 50.000	Range 76 - 130		Recovery =	40.96%#		
48) Toluene-d8	7.50	98	656553	21.35	ug/l	-0.02
Spiked Amount 50.000	Range 78 - 121		Recovery =	42.70%#		
62) 4-Bromofluorobenzene	11.35	95	213533	20.59	ug/l	0.00
Spiked Amount 50.000	Range 70 - 131		Recovery =	41.18%#		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA H\DATA\VH092215\
Data File : VH057114.D
Acq On : 22 Sep 2015 20:11
Operator : SY/FY
Sample : G3712-06ME
Misc : 5.01g/10mL//100uL/5mL/MSVOA_H/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_H
ClientSampleId :
ERD45-CS4-3ME

Quant Time: Sep 24 12:00:26 2015
Quant Method : W:\HPCHEM1\MSVOA_H\METHODS\82H091715W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 18 15:34:01 2015
Response via : Initial Calibration

