

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL010220\
 Data File : VL034544.D
 Acq On : 2 Jan 2020 12:44
 Operator : SY/AP
 Sample : VL0102ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0102ABL01

Manual Integrations
 APPROVED

apatel
 1/3/2020 11:37:45 AM

Quant Time: Jan 03 05:31:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 3
 QLast Update : Mon Dec 30 14:58:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	907602	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1814872	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	1784195	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.46	95	1482032	9.95	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.50%
Target Compounds						
15) Acetone	3.85	43	30016	0.227	ppbv	Ovalue # 68
20) Methylene Chloride	4.34	84	10028m	0.192	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL010220\
Data File : VL034544.D
Acq On : 2 Jan 2020 12:44
Operator : SY/AP
Sample : VL0102ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0102ABL01

Manual Integrations
APPROVED

apatel
1/3/2020 11:37:45 AM

Quant Time: Jan 03 05:31:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 30 14:58:50 2019
Response via : Initial Calibration

