

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL010816\
 Data File : VL026651.D
 Acq On : 8 Jan 2016 18:00
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

apatel
 1/11/2016 5:13:32 PM

Quant Time: Jan 08 23:06:31 2016
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jan 08 2016
 QLast Update : Fri Jan 08 22:06:23 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.63	49	809162	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.30	114	2453651	10.00	ppbv	-0.01
55) Chlorobenzene-d5	13.71	117	3184011	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.25	95	2612795	10.18	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.87	62	6217	0.10	ppbv	# 87
32) 1,1,1-Trichloroethane	7.57	97	21570	0.11	ppbv	95
35) Carbon Tetrachloride	8.13	117	22290	0.11	ppbv	88
38) Trichloroethene	9.04	130	10687	0.11	ppbv	90
52) Tetrachloroethene	12.76	164	12648m	0.12	ppbv	
63) 1,1,2,2-Tetrachloroethane	15.45	83	39100	0.13	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL010816\
Data File : VL026651.D
Acq On : 8 Jan 2016 18:00
Operator : SY/AP
Sample : VSTDICC0.1
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

apatel
1/11/2016 5:13:32 PM

Quant Time: Jan 08 23:06:31 2016
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jan
QLast Update : Fri Jan 08 22:06:23 2016
Response via : Initial Calibration

