

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL010816\
 Data File : VL026654.D
 Acq On : 8 Jan 2016 23:36
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Jan 12 03:41:26 2016
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Jan 11 17:14:02 2016
 QLast Update : Mon Jan 11 17:14:02 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.62	49	743582	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	8.29	114	2344981	10.00	ppbv	-0.02
55) Chlorobenzene-d5	13.71	117	3117158	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.25	95	2539702	10.22	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	4.59	43	42623	0.35	ppbv	# 17

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL010816\
Data File : VL026654.D
Acq On : 8 Jan 2016 23:36
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jan 12 03:41:26 2016
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 11 17:14:02 2016
QLast Update : Mon Jan 11 17:14:02 2016
Response via : Initial Calibration

