

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL011216\  
 Data File : VL026669.D  
 Acq On : 12 Jan 2016 10:18  
 Operator : SY/AP  
 Sample : VL0112ABL01  
 Misc : 400mL/MSVOA L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 ClientSampleId :  
 VL0112ABL01

Quant Time: Jan 13 04:23:52 2016  
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA\_L Mon Jan 11 17:14:02 2016  
 QLast Update : Mon Jan 11 17:14:02 2016  
 Response via : Initial Calibration

| Internal Standards      | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------|-------|------|----------|-------|-------|----------|
| 1) Bromochloromethane   | 6.65  | 49   | 817348   | 10.00 | ppbv  | 0.02     |
| 33) 1,4-Difluorobenzene | 8.32  | 114  | 2449132  | 10.00 | ppbv  | 0.01     |
| 55) Chlorobenzene-d5    | 13.74 | 117  | 3219505  | 10.00 | ppbv  | 0.02     |

System Monitoring Compounds

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| 68) 1-Bromo-4-Fluorobenzene | 16.29  | 95    | 2682747  | 10.45    | ppbv | 0.02    |
| Spiked Amount               | 10.000 | Range | 65 - 135 | Recovery | =    | 104.50% |

| Target Compounds | R.T. | QIon | Response | Conc | Units | Ovalue |
|------------------|------|------|----------|------|-------|--------|
| 15) Acetone      | 4.63 | 43   | 48544    | 0.36 | ppbv  | # 1    |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL011216\  
Data File : VL026669.D  
Acq On : 12 Jan 2016 10:18  
Operator : SY/AP  
Sample : VL0112ABL01  
Misc : 400mL/MSVOA L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
VL0112ABL01

Quant Time: Jan 13 04:23:52 2016  
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M  
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Mon Jan 11 17:14:02 2016  
QLast Update : Mon Jan 11 17:14:02 2016  
Response via : Initial Calibration

