

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL011216\
 Data File : VL026678.D
 Acq On : 12 Jan 2016 17:13
 Operator : SY/AP
 Sample : H1075-02DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-6DL

Quant Time: Jan 13 07:42:19 2016
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 11 17:14:02 2016
 QLast Update : Mon Jan 11 17:14:02 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.68	49	668872	10.00	ppbv	0.05
33) 1,4-Difluorobenzene	8.36	114	2122187	10.00	ppbv	0.05
55) Chlorobenzene-d5	13.78	117	2908083	10.00	ppbv	0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.32	95	2412138	10.41	ppbv	0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.10%

Target Compounds						Ovalue
3) Chlorodifluoromethane	3.60	51	37869	0.43	ppbv	95
10) Heptane	9.38	43	13707	0.11	ppbv	93
36) Benzene	8.06	78	32189	0.15	ppbv	97
44) 2,2,4-Trimethylpentane	9.12	57	65621	0.19	ppbv #	76
53) Toluene	11.27	91	269440	0.88	ppbv	99
58) Ethyl Benzene	14.45	91	47530	0.09	ppbv #	87
59) m/p-Xylene	14.73	91	156199	0.36	ppbv	95
60) o-Xylene	15.53	91	65919	0.14	ppbv	100
72) 4-Ethyltoluene	17.86	105	21818	0.04	ppbv #	91
74) 1,2,4-Trimethylbenzene	18.86	105	65256	0.13	ppbv #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL011216\
Data File : VL026678.D
Acq On : 12 Jan 2016 17:13
Operator : SY/AP
Sample : H1075-02DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-6DL

Quant Time: Jan 13 07:42:19 2016
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL010816AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 11 17:14:02 2016
QLast Update : Mon Jan 11 17:14:02 2016
Response via : Initial Calibration

