

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
 Data File : VL041892.D
 Acq On : 14 Jan 2025 14:30
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 14 21:21:01 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Jan 14 16:31:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	146186	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.968	114	430874	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.891	117	379402	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.383	95	297462	9.954	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	10599	0.532	ppbv	99
3) Chlorodifluoromethane	1.476	51	10582	0.507	ppbv	98
4) Chloromethane	1.534	50	3990	0.533	ppbv	98
5) Vinyl Chloride	1.580	62	4097	0.540	ppbv	94
6) Bromomethane	1.670	94	1852	0.512	ppbv #	78
7) Chloroethane	1.706	64	1532	0.519	ppbv #	77
8) Dichlorotetrafluoroethane	1.557	85	9372	0.539	ppbv	95
9) Propene	1.486	41	4690	0.546	ppbv	85
10) Heptane	4.988	43	11230	0.489	ppbv	95
11) Trichlorofluoromethane	1.881	101	10873	0.525	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.143	101	8027	0.494	ppbv	90
13) Ethanol	1.725	45	478m	0.759	ppbv	
14) Bromoethene	1.784	108	3071	0.517	ppbv #	94
15) Acetone	1.839	43	10273	0.585	ppbv	99
16) 1,3-Butadiene	1.612	39	4038	0.512	ppbv	99
17) tert-Butyl alcohol	2.046	59	11158	0.516	ppbv #	94
18) 1,1-Dichloroethene	2.036	96	3803	0.520	ppbv	95
19) Isopropyl Alcohol	1.891	45	6317	0.607	ppbv #	32
20) Methylene Chloride	2.065	84	4286	0.618	ppbv	93
21) Allyl Chloride	2.101	41	6553	0.539	ppbv	96
22) trans-1,2-Dichloroethene	2.347	96	4015	0.486	ppbv	89
23) Vinyl Acetate	2.467	43	6381m	0.835	ppbv	
24) 1,1-Dichloroethane	2.415	63	8512	0.533	ppbv	97
25) Ethyl Acetate	2.845	43	23529	0.497	ppbv	97
26) Hexane	2.829	57	9246m	0.494	ppbv	
27) Carbon Disulfide	2.153	76	8847	0.477	ppbv #	1
28) Methyl tert-Butyl Ether	2.447	73	7621	0.532	ppbv	97
29) Chloroform	2.855	83	15360	0.519	ppbv	92
30) Cyclohexane	3.871	84	7579	0.467	ppbv	92
31) cis-1,2-Dichloroethene	2.726	61	10275	0.521	ppbv #	89
32) 1,1,1-Trichloroethane	3.379	97	14198	0.501	ppbv	98
34) 2-Butanone	2.567	43	13981	0.510	ppbv	94
35) Carbon Tetrachloride	3.774	117	12590	0.464	ppbv	89
36) Benzene	3.667	78	19954	0.505	ppbv	95
37) 1,2-Dichloroethane	3.227	62	11613	0.523	ppbv	97
38) Trichloroethene	4.561	130	8199	0.523	ppbv #	89
39) 1,2-Dichloropropane	4.324	63	7526m	0.543	ppbv	
40) 1,4-Dioxane	4.616	88	4031m	0.576	ppbv	
41) Tetrahydrofuran	3.072	42	7183	0.481	ppbv	94
42) Bromodichloromethane	4.499	83	15081m	0.507	ppbv	
43) Methyl Methacrylate	4.891	69	7332m	0.499	ppbv	
44) 2,2,4-Trimethylpentane	4.645	57	31243	0.489	ppbv	95
45) t-1,3-Dichloropropene	6.428	75	6465m	0.470	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
 Data File : VL041892.D
 Acq On : 14 Jan 2025 14:30
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 14 21:21:01 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Jan 14 16:31:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	7986m	0.443	ppbv	
47) 1,1,2-Trichloroethane	6.597	97	7324	0.498	ppbv #	84
48) Dibromochloromethane	7.393	129	10548m	0.465	ppbv	
49) Bromoform	9.493	173	8162	0.438	ppbv	97
50) 4-Methyl-2-Pentanone	5.794	43	19221m	0.515	ppbv	
51) 2-Hexanone	7.454	43	13390	0.466	ppbv #	97
52) Tetrachloroethene	8.238	164	6821m	0.498	ppbv	
53) Toluene	6.946	91	20118	0.467	ppbv	96
54) 1,2-Dibromoethane	7.665	107	10511	0.520	ppbv	89
56) 1,1,1,2-Tetrachloroethane	8.930	131	9524	0.503	ppbv #	1
57) Chlorobenzene	8.930	112	18701	0.534	ppbv #	90
58) Ethyl Benzene	9.364	91	28584	0.482	ppbv #	88
59) m/p-Xylene	9.558	91	45003m	0.946	ppbv	
60) o-Xylene	9.972	91	22370	0.474	ppbv	97
61) Styrene	9.879	104	9788	0.463	ppbv	99
62) Isopropylbenzene	10.545	105	36561	0.501	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.969	83	15485	0.521	ppbv	98
64) n-propylbenzene	10.995	120	9128	0.495	ppbv	93
65) tert-Butylbenzene	11.536	119	33122	0.498	ppbv	98
66) Benzyl Chloride	11.620	91	1935	0.470	ppbv	99
67) sec-Butylbenzene	11.775	105	45326	0.495	ppbv	98
69) p-Isopropyltoluene	11.931	119	35136	0.466	ppbv	98
70) n-Butylbenzene	12.287	91	35409	0.473	ppbv	96
71) 2-Chlorotoluene	10.908	91	26436	0.490	ppbv	99
72) 4-Ethyltoluene	11.131	105	27363	0.485	ppbv	97
73) 1,3,5-Trimethylbenzene	11.209	105	23183	0.465	ppbv	93
74) 1,2,4-Trimethylbenzene	11.545	105	25777	0.478	ppbv	89
75) 1,3-Dichlorobenzene	11.610	146	16708	0.522	ppbv	100
76) 1,4-Dichlorobenzene	11.678	146	15738	0.503	ppbv	96
77) 1,2-Dichlorobenzene	11.953	146	16364	0.516	ppbv	93
78) Hexachloro-1,3-Butadiene	13.892	225	14162	0.544	ppbv	96
79) Naphthalene	13.520	128	28052	0.516	ppbv	98
80) Naphthalene,2-methyl-	14.468	142	10859	0.422	ppbv	97
81) 1,2,4-Trichlorobenzene	13.445	180	14136	0.533	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
Data File : VL041892.D
Acq On : 14 Jan 2025 14:30
Operator : SY/MD
Sample : VSTDICC0.5
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 14 21:21:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Jan 14 16:31:18 2025
Response via : Initial Calibration

