

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
 Data File : VL041913.D
 Acq On : 15 Jan 2025 02:53
 Operator : SY/MD
 Sample : MDL 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL 0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 04:26:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Jan 14 23:15:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	126935	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	367914	10.000	ppbv	-0.01
55) Chlorobenzene-d5	8.885	117	320329	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	249196	9.877	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	213m	0.032	ppbv	
32) 1,1,1-Trichloroethane	3.373	97	806m	0.033	ppbv	
52) Tetrachloroethene	8.228	164	431m	0.037	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.969	83	836	0.033	ppbv #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
Data File : VL041913.D
Acq On : 15 Jan 2025 02:53
Operator : SY/MD
Sample : MDL 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Quant Time: Jan 15 04:26:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Jan 14 23:15:42 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025

