

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
 Data File : VL041897.D
 Acq On : 14 Jan 2025 18:34
 Operator : SY/MD
 Sample : VL0114ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0114ABL01

Quant Time: Jan 15 03:24:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 14 23:15:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.793	49	140579	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	414150	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	360177	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	281871	9.936	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.400%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011425\
Data File : VL041897.D
Acq On : 14 Jan 2025 18:34
Operator : SY/MD
Sample : VL0114ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0114ABL01

Quant Time: Jan 15 03:24:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Jan 14 23:15:42 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.793 min Scan# 83
Delta R.T. -0.004 min
Lab File: VL041897.D
Acq: 14 Jan 2025 18:34




Instrument :
MSVOA_L
ClientSampleId :
VL0114ABL01


Tgt Ion: 49 Resp: 140579
Ion Ratio Lower Upper
49 100
128 49.5 24.3 73.0
130 63.4 31.1 93.2



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.965 min Scan# 1198
Delta R.T. -0.010 min
Lab File: VL041897.D
Acq: 14 Jan 2025 18:34

Tgt Ion: 114 Resp: 414150
Ion Ratio Lower Upper
114 100
63 24.8 19.8 29.6
88 18.4 14.4 21.6





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.888 min Scan# 27

Delta R.T. -0.006 min

Lab File: VL041897.D

Acq: 14 Jan 2025 18:34

Instrument :

MSVOA_L

ClientSampleId :

VL0114ABL01





Tgt Ion:117 Resp: 360177

Ion Ratio Lower Upper

117 100

82 65.0 54.2 81.4

119 31.7 25.0 37.6



#68

1-Bromo-4-Fluorobenzene

Concen: 9.936 ppbv

RT: 10.383 min Scan# 3181

Delta R.T. -0.007 min

Lab File: VL041897.D

Acq: 14 Jan 2025 18:34

Tgt Ion: 95 Resp: 281871

Ion Ratio Lower Upper

95 100

174 59.1 49.0 73.4

175 4.8 3.8 5.8

