

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011623\
 Data File : VL040095.D
 Acq On : 16 Jan 2023 12:49
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 01/17/2023
 Supervised By : Mahesh Dadoda 01/17/2023

Quant Time: Jan 16 13:25:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2023
 QLast Update : Mon Jan 16 11:28:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	1083007	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	2913980	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	2566968	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.789	95	2139227	10.910	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 109.100%	
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.912	62	6520m	0.080	ppbv	
32) 1,1,1-Trichloroethane	5.999	97	13925m	0.070	ppbv	
35) Carbon Tetrachloride	6.494	117	13945m	0.077	ppbv	
38) Trichloroethene	7.298	130	10427m	0.078	ppbv	
52) Tetrachloroethene	10.622	164	6083m	0.069	ppbv	
54) 1,2-Dibromoethane	10.015	107	6963m	0.061	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.066	83	7622m	0.068	ppbv	
79) Naphthalene	21.426	128	8159m	0.041	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011623\
Data File : VL040095.D
Acq On : 16 Jan 2023 12:49
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/17/2023
Supervised By :Mahesh Dadoda 01/17/2023

Quant Time: Jan 16 13:25:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Jan 16 11:28:27 2023
Response via : Initial Calibration

