

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011623\
 Data File : VL040096.D
 Acq On : 16 Jan 2023 13:26
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/17/2023
 Supervised By :Mahesh Dadoda 01/17/2023

Quant Time: Jan 16 13:59:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Jan 16 11:28:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	1105720	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	2804348	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	2497342	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.793	95	2085299	10.931	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	109.300%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.912	62	2182m	0.026	ppbv	
32) 1,1,1-Trichloroethane	6.005	97	4409m	0.022	ppbv	
35) Carbon Tetrachloride	6.494	117	4077m	0.023	ppbv	
38) Trichloroethene	7.288	130	3647m	0.028	ppbv	
52) Tetrachloroethene	10.642	164	2061m	0.024	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.076	83	2898m	0.027	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011623\
Data File : VL040096.D
Acq On : 16 Jan 2023 13:26
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Jan 16 13:59:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Jan 16 11:28:27 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/17/2023
Supervised By : Mahesh Dadoda 01/17/2023

