

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL012218\
 Data File : VL031420.D
 Acq On : 22 Jan 2018 16:37
 Operator : SY/AP
 Sample : VSTDIC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03

Manual Integrations
 APPROVED

MMDadoda
 1/24/2018 5:10:14 PM

Quant Time: Jan 22 17:10:40 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL012218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 22 2018
 QLast Update : Mon Jan 22 16:42:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.90	49	1242722	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.44	114	2849275	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.41	117	2842612	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.75	95	2297133	10.23	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.40	62	1728m	0.02	ppbv	
32) 1,1,1-Trichloroethane	6.77	97	5580m	0.02	ppbv	
35) Carbon Tetrachloride	7.28	117	6158m	0.02	ppbv	
38) Trichloroethene	8.12	130	2228	0.02	ppbv #	76
52) Tetrachloroethene	11.55	164	2241m	0.02	ppbv	
63) 1,1,2,2-Tetrachloroethane	14.01	83	7569m	0.03	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL012218\
Data File : VL031420.D
Acq On : 22 Jan 2018 16:37
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
1/24/2018 5:10:14 PM

Quant Time: Jan 22 17:10:40 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL012218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan
QLast Update : Mon Jan 22 16:42:46 2018
Response via : Initial Calibration

