

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012324\
 Data File : VL041009.D
 Acq On : 23 Jan 2024 23:24
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Jan 24 01:02:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Jan 24 00:50:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.156	49	1008118	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.613	114	2791588	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.423	117	2369263	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.735	95	1720311	9.613	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.100%
Target Compounds						
					Qvalue	
15) Acetone	3.465	43	25851	0.201	ppbv #	88
20) Methylene Chloride	3.899	84	8610	0.174	ppbv #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012324\
Data File : VL041009.D
Acq On : 23 Jan 2024 23:24
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jan 24 01:02:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Jan 24 00:50:15 2024
Response via : Initial Calibration

