

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012422\
Data File : VL038281.D
Acq On : 25 Jan 2022 4:45
Operator : SY/AP
Sample : OC012022 CAN 10595
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC012022 CAN 10595

Quant Time: Jan 25 05:19:08 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	776631	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2011977	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	1677829	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1273427	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012422\
Data File : VL038281.D
Acq On : 25 Jan 2022 4:45
Operator : SY/AP
Sample : OC012022 CAN 10595
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC012022 CAN 10595

Quant Time: Jan 25 05:19:08 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

