


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL012422\
 Data File : VL038252.D
 Acq On : 24 Jan 2022 8:49
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
 Last Update : Mon Jan 24 14:12:48 2022



AutoFind: Scans 3615, 3616, 3617; Background Corrected with Scan 3591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.8	129296	PASS
75	95	30	66	55.9	304362	PASS
95	95	100	100	100.0	544277	PASS
96	95	5	9	6.6	35962	PASS
173	174	0.00	2	1.2	3653	PASS
174	95	50	120	57.5	313109	PASS
175	174	4	9	7.3	22813	PASS
176	174	93	101	97.4	304960	PASS
177	176	5	9	6.5	19869	PASS