

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012424\
 Data File : VL041017.D
 Acq On : 24 Jan 2024 15:45
 Operator : SY/MD
 Sample : P1187-13 LOD 0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT1-2024

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
 Supervised By :Semsettin Yesilyurt 01/25/2024

Quant Time: Jan 25 03:20:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Thu Jan 25 03:13:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.160	49	1002956	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.623	114	2529049	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.436	117	2145908	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.741	95	1620158m	9.996	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.000%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.896	62	5651m	0.091	ppbv	
32) 1,1,1-Trichloroethane	5.973	97	13583	0.081	ppbv	88
35) Carbon Tetrachloride	6.459	117	14214m	0.084	ppbv	
38) Trichloroethene	7.259	130	6972m	0.083	ppbv	
52) Tetrachloroethene	10.577	164	5811m	0.084	ppbv	
54) 1,2-Dibromoethane	9.979	107	7637m	0.064	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.015	83	13581m	0.072	ppbv	
79) Naphthalene	21.361	128	5238m	0.034	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012424\
Data File : VL041017.D
Acq On : 24 Jan 2024 15:45
Operator : SY/MD
Sample : P1187-13 LOD 0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT1-2024

Quant Time: Jan 25 03:20:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Jan 25 03:13:06 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024

