

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012622\
 Data File : VL038289.D
 Acq On : 26 Jan 2022 15:26
 Operator : SY/AP
 Sample : N1268-03DL2 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-1DL2

Quant Time: Jan 27 03:54:07 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL012422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
 QLast Update : Mon Jan 24 14:12:48 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	820823	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	2301517	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.05	117	1837287	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	1381571	9.91	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
10) Heptane	8.09	43	22873	0.167	ppbv	97
13) Ethanol	3.61	45	1411	0.247	ppbv #	11
26) Hexane	5.67	57	58993	0.626	ppbv #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012622\
Data File : VL038289.D
Acq On : 26 Jan 2022 15:26
Operator : SY/AP
Sample : N1268-03DL2 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-1DL2

Quant Time: Jan 27 03:54:07 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

